

WAVETEK

Wavetek Ltd

*Hurricane Way, Norwich
Norfolk NR6 6JB, United Kingdom
Tel: +(44) (0) 1603 404824
Fax: +(44) (0) 1603 483670*

Fax Form

Date: 25 September 1996

To Fax No:

Attention:

Company:

Copy To:

From: Andy Eddy

Subject: 1271 Drift

No. Pages (inc. Cover): 1

Message:

I do not like to say this, but we unfortunately know of a problem with the 100 & 1KV attenuator on the DC pcb. Some of the attenuator resistors seem to have a long term drift. If you open the 1271 you will invalidate the NAMAS certification, so it is best we replace the attenuator resistors when it is most convenient to you.

Regards

