

Supplement

Fluke 50 Series II CD-ROM

PN: 1276106

Supplement Issue: 1

Print Date: 10/99

Date: 10/99

This supplement contains information necessary to ensure the accuracy of the following manual(s) on this CD:

- 51 & 52 Series II Users Manual
- 53 & 54 Series II Users Manual

51 & 52 Users Manual:

On page 13, under Environmental, replace the Humidity section with:

Humidity	Non condensing (<10°C) (<50°F) 95% RH: 10°C to 30°C (50°F to 86°F) 75% RH: 30°C to 40°C (86°F to 104°F) 45% RH: 40°C to 50°C (104°F to 122°F) (Without Condensation)
----------	---

On page 15, under Electrical, following the Temperature Coefficient section, add:

Electromagnetic Compatibility	Susceptibility: $\pm 2^{\circ}\text{C}$ ($\pm 3.6^{\circ}\text{F}$) for 80 MHz to 200 MHz in 1.5 V/m field, for 200 MHz to 1000 MHz in 3 V/m field. Emissions: Commercial Limits per EN50081-1
-------------------------------	---

53 & 54 Users Manual:

On page 19, under Environmental, replace the Humidity section with:

Humidity	Non condensing (<10°C) (<50°F) 95% RH: 10°C to 30°C (50°F to 86°F) 75% RH: 30°C to 40°C (86°F to 104°F) 45% RH: 40°C to 50°C (104°F to 122°F) (Without Condensation)
----------	---

On page 21, under Electrical, following the Temperature Coefficient section, add:

Electromagnetic Compatibility	Susceptibility: $\pm 2^{\circ}\text{C}$ ($\pm 3.6^{\circ}\text{F}$) for 80 MHz to 200 MHz in 1.5 V/m field, for 200 MHz to 1000 MHz in 3 V/m field. Emissions: Commercial Limits per EN50081-1
-------------------------------	---