Notice

Hewlett-Packard to Agilent Technologies Transition

This documentation supports a product that previously shipped under the Hewlett-Packard company brand name. The brand name has now been changed to Agilent Technologies. The two products are functionally identical, only our name has changed. The document still includes references to Hewlett-Packard products, some of which have been transitioned to Agilent Technologies.



Printed in USA March 2000

By internet, phone, or fax, get assistance with all your test and measurement needs.

Table 1-1 Contacting Agilent

- -

Online assistance: www.agilent.com/find/assist

| United States (tel) 1 800 452 4844 | Latin America (tel) (305) 269 7500 (fax) (305) 269 7599 | Canada (tel) 1 877 894 4414 (fax) (905) 282-6495 | Europe (tel) (+31) 20 547 2323 (fax) (+31) 20 547 2390 |
|--|---|---|---|
| New Zealand | Japan | Australia | |
| (tel) 0 800 738 378 | (tel) (+81) 426 56 7832 | (tel) 1 800 629 485 | |
| (fax) (+64) 4 495 8950 | (fax) (+81) 426 56 7840 | (fax) (+61) 3 9210 5947 | |

Asia Call Center Numbers

| Country | Phone Number | Fax Number |
|-------------------------------|--|---|
| Singapore | 1-800-375-8100 | (65) 836-0252 |
| Malaysia | 1-800-828-848 | 1-800-801664 |
| Philippines | (632) 8426802 1-800-16510170 (PLDT Subscriber Only) | (632) 8426809 1-800-16510288 (PLDT Subscriber Only) |
| Thailand | (088) 226-008 (outside Bangkok) (662) 661-3999 (within Bangkok) | (66) 1-661-3714 |
| Hong Kong | 800-930-871 | (852) 2506 9233 |
| Taiwan | 0800-047-866 | (886) 2 25456723 |
| People's Republic of China | 800-810-0189 (preferred) 10800-650-0021 | 10800-650-0121 |
| India | 1-600-11-2929 | 000-800-650-1101 |

Getting Started and Quick Reference Guide

Agilent Technologies 8591C Cable TV Analyzer



Part Number: 08591-90130 Printed in USA June 1995

© Copyright 1994, 1995 Agilent Technologies, Inc

•

Notice.

The information contained in this document is subject to change without notice.

Agilent Technologies makes no warranty of any kind with regard to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Warranty

This Hewlett-Packard instrument product is warranted against defects in material and workmanship for a period of one year from date of shipment. During the warranty period, Hewlett-Packard Company will, at its option, either repair or replace products that prove to be defective.

For warranty service or repair, this product must be returned to a service facility designated by HP. Buyer shall prepay shipping charges to HP and HP shall pay shipping charges to return the product to Buyer. However, Buyer shall pay all shipping charges, duties, and taxes for products returned to HP from another country.

HP warrants that its software and firmware designated by HP for use with an instrument will execute its programming instructions when properly installed on that instrument. HP does not warrant that the operation of the instrument, or software, or firmware will be uninterrupted or error-free. ITSheadLimitation of Warranty

The foregoing warranty shall not apply to defects resulting from improper or inadequate maintenance by Buyer, Buyer-supplied software or interfacing, unauthorized modification or misuse, operation outside of the environmental specifications for the product, or improper site preparation or maintenance.

NO OTHER WARRANTY IS EXPRESSED OR IMPLIED. HP SPECIFICALLY DISCLAIMS THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE.

ITSheadExclusive Remedies

THE REMEDIES PROVIDED HEREIN ARE BUYER'S SOLE AND EXCLUSIVE REMEDIES. HP SHALL NOT BE LIABLE FOR ANY DIRECT, INDIRECT, SPECIAL, INCIDENTAL, OR CONSEQUENTIAL DAMAGES, WHETHER BASED ON CONTRACT, TORT, OR ANY OTHER LEGAL THEORY.

Contents

| 1. | About This Guide | |
|----|--|---|
| | Other HP 8591C Cable TV Analyzer Documentation | 1-3 |
| | Documentation Options | 1-4 |
| | How to Order Guides | 1-4 |
| | Before You Begin | 1-5 |
| | Conventions | 1-5 |
| | Safety Notes | 1-6 |
| | General Safety Considerations | 1-7 |
| | Symbols Used on the Cable TV Analyzer | 1-8 |
| 2. | Preparing for Use | |
| | What You'll Learn in This Chapter | 2-2 |
| | Preparing Your Cable TV Analyzer for Use | 2-3 |
| | Step 1. Perform an initial inspection | 2-4 |
| | Step 2. Set the line voltage selector switch | 2-6 |
| | Step 3. Check the fuse | 2-8 |
| | Step 4. Connect the power cable | 2-10 |
| | Step 5. Turn on the cable TV analyzer for the first time | 2-12 |
| | Stan 6 Denform the frequency and emplitude | |
| | Step 6. Ferform the frequency and amplitude | |
| | self-calibration routines | 2-14 |
| | self-calibration routines | 2-14 |
| | Step 6. Perform the frequency and ampirtude self-calibration routines | 2-14 2-16 |
| 3. | step 6. Perform the frequency and ampirtude self-calibration routines | 2-14 2-16 |
| 3. | Step 6. Perform the frequency and ampirtude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter | 2-14 2-16 3-2 |
| 3. | Step 6. Perform the frequency and ampirtude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV | 2-14 2-16 3-2 |
| 3. | Step 6. Perform the frequency and ampirtude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer | 2-14 2-16 3-2 3-3 |
| 3. | Step 6. Perform the frequency and ampirtude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features | 2-14 2-16 3-2 3-3 3-4 |
| 3. | Step 6. Perform the frequency and amplitude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features Rear-Panel Features | 2-14 2-16 3-2 3-3 3-4 3-6 |
| 3. | Step 6. Perform the frequency and amplitude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features Rear-Panel Features Data Controls | 2-14 2-16 3-2 3-3 3-4 3-6 3-8 |
| 3. | Step 6. Perform the frequency and amplitude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features Rear-Panel Features Data Controls Hold Key | 2-14 2-16 3-2 3-3 3-4 3-6 3-8 3-8 3-8 |
| 3. | Step 6. Perform the frequency and amplitude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features Data Controls Hold Key Knob | 2-14 2-16 3-2 3-3 3-4 3-6 3-8 3-8 3-8 3-8 |
| 3. | Step 6. Perform the frequency and ampirtude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features Data Controls Hold Key Mumber/Units Keypad | 2-14 2-16 3-2 3-3 3-4 3-6 3-8 3-8 3-8 3-8 3-8 3-9 |
| 3. | Step 6. Perform the frequency and amplitude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features Rear-Panel Features Data Controls Hold Key Number/Units Keypad Step Keys | 2-14 2-16 3-2 3-3 3-4 3-6 3-8 3-8 3-8 3-8 3-8 3-8 3-8 3-9 3-9 3-9 3-9 |
| 3. | Step 6. Perform the frequency and amplitude self-calibration routines Step 7. Perform the tracking-generator self-calibration routine (Option 011 only) Quick Tour What You'll Learn in This Chapter Getting Acquainted with the HP 8591C Cable TV Analyzer Front-Panel Features Rear-Panel Features Data Controls Knob Number/Units Keypad Step Keys Fine-Focus Control | 2-14 2-16 3-2 3-3 3-4 3-6 3-8 3-8 3-8 3-8 3-8 3-9 3-9 3-9 3-10 |

Quick Reference

Contents-1

Cable TV Analyzer Functions

| Menu and Softkey Overview | 3-15 |
|--|--------|
| Cable TV Measurements and System Monitor Personality | 3 - 16 |
| Making a Measurement | 3 - 17 |
| Measurement Summary | 3-20 |
| Improving Accuracy with Self-Calibration Routines | 3 - 21 |
| Warmup Time | 3 - 21 |
| Calibrating the Analyzer | 3-22 |
| Calibrating the Tracking Generator (Option 011 only) . | 3-24 |
| When Is Self-Calibration Needed? | 3 - 26 |
| Printing or Plotting | 3-27 |
| Memory Card Insertion and Battery Replacement | 3-28 |
| Changing the Memory Card Battery | 3-29 |
| Procedure to Change the Memory Card Battery | 3-30 |
| Analyzer Battery Information | 3-32 |
| Softkey Descriptions | |
| What You'll Find in This Chapter | 4-2 |
| Service Functions | 4-3 |
| Service Calibration Functions | 4-3 |
| Service Diagnostic Functions | 4-4 |
| Service Diagnostic Functions (continued) | 4-5 |
| Cable TV Analyzer Functions | 4-6 |
| | |

5. Key Menus

4.

Index

Figures

| 2-1. | Setting the Line Voltage Selector Switch | 2-7 |
|-------|--|-----|
| 2-2. | Checking the Line Fuse | 2-8 |
| 2-3. | Reference Connector | -12 |
| 3-1. | HP 8591C Cable TV Analyzer | 3-3 |
| 3-2. | Front-Panel Feature Overview | 3-4 |
| 3-3. | Rear-Panel Feature Overview | 3-6 |
| 3-4. | Adjusting the Fine Focus | -10 |
| 3-5. | Screen Annotation | -11 |
| 3-6. | Relationship between Frequency and Amplitude | -18 |
| 3-7. | Reading the Amplitude and Frequency | -20 |
| 3-8. | Inserting the Memory Card | -28 |
| 3-9. | Memory Card Battery Date Code Location | -30 |
| 3-10. | Memory Card Battery Replacement | -31 |
| 3-11. | Rear-Panel Battery Information Label | -32 |

Quick Reference

Tables

| 2-1. | Accessories Supplied with the Cable TV Analyzer | 2-5 |
|--------|--|---------|
| 2-2. | Power Requirements | 2-6 |
| 2-3. | AC Power Cables Available | 2-11 |
| 3-1. | Screen Annotation | 3 - 12 |
| 3-2. | Screen Annotation for Trace, Trigger, and Sweep Modes | 3-13 |
| 4-1. | Commands Not Available with Analog+ Operation | 4-12 |
| 4 - 2. | Default Configuration Values | 4-37 |
| 4-3. | Compatibility of FFT with Other Functions | 4-54 |
| 4-4. | Commands Altered/Not Available within the Gate Utility . | 4-59 |
| 4-5. | Model Specific Preset Conditions | 4-90 |
| 4-6. | Common Preset Conditions | 4-90 |
| 4-7. | Preset Conditions for All Models | 4-92 |
| 4-8. | Functions That Exit the Windows Display Format | 4 - 136 |

Contents-4

Quick Reference

1

About This Guide

About This Guide

Welcome to the *HP 8591C Cable TV Analyzer Getting Started and Quick Reference Guide*. This guide is divided into two basic sections:

| Getting Started | Chapter 1, Chapter 2, and Chapter 3 instruct you how to prepare the cable TV analyzer for use, and introduce you to the cable TV analyzer and its features. |
|-----------------|---|
| Quick Reference | Chapter 4 and Chapter 5 provide you with a quick reference for commonly used functions of the HP 8591C Cable TV Analyzer. Refer to the HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide and HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide for more detailed information. |

Other HP 8591C Cable TV Analyzer Documentation

In addition to this guide, other printed guides will help you learn about the HP 8591C Cable TV Analyzer and locate information quickly.

HP 8590 E-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Calibration Guide

• Tells you how to test your spectrum analyzer to determine if the spectrum analyzer meets its specifications.

HP 8591C Cable TV Analyzer User's Guides There are two user's guides included with the cable TV analyzer:

Cable TV Measurements

- Tells you how to make cable TV measurements with your cable TV analyzer.
- Describes the cable TV analyzer mode features.

Spectrum Analyzer Reference

- Tells you how to make measurements using the spectrum analyzer mode.
- Describes the spectrum analyzer mode features.
- Tells you what to do in case of a failure.

HP 8590 E-Series and L-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Programmer's Guide

Describes analyzer operation via a remote controller (computer).

Documentation Options

Option 910: Additional User's Documentation

Provides an additional copy of the user's guides, the calibration guide, and the quick reference guide.

Option 915: Assembly-Level and Component-Level Repair Service Guides

Describes troubleshooting and repair of the spectrum analyzer.

Option 915 consists of two manuals:

HP 8590 E-Series and L-Series Spectrum Analyzers, and HP 8591C Cable TV Analyzer Assembly-Level Repair Service Guide

• Describes adjustment and assembly-level repair of the analyzer.

HP 8590 E-Series and L-Series Spectrum Analyzers, and HP 8591C Cable TV Analyzer Component-Level Repair Service Guide

• Provides information for component-level repair of the analyzer.

How to Order Guides

Each of the guides listed above can be ordered individually. To order, contact your local HP Sales and Service Office.

Before You Begin

Become familiar with the conventions and safety information provided in this chapter before you begin using your cable TV analyzer.

Conventions

This guide uses the following conventions:

| (Front-Panel Key) | A boxed, uppercase name in this typeface represents a key physically located on the instrument. |
|-------------------------|--|
| Softkey | A boxed word written in this typeface indicates a "softkey," a key whose label is determined by the instrument's firmware. |
| Softkey ON OFF (ON) | A boxed word written in this typeface with the words ON and OFF can turn a function on or off. The underlined function is shown parenthetically. |
| Softkey AUTO MAN (AUTO) | A boxed word written in this typeface with the words AUTO and MAN can either be auto-coupled or have its value manually changed. The underlined function is shown parenthetically. |
| Screen Text | Text printed in this typeface indicates text displayed on the spectrum analyzer screen. |

Safety Notes The following safety notes are used throughout this manual. Familiarize yourself with each of the notes and its meaning before operating this instrument. CAUTION Caution denotes a hazard. It calls attention to a procedure that, if not correctly performed or adhered to, could result in damage to or destruction of the instrument. Do not proceed beyond a caution sign until the indicated conditions are fully understood and met. WARNING Warning denotes a hazard. It calls attention to a procedure that, if not correctly performed or adhered to, could result in injury or loss of life. Do not proceed beyond a warning sign until the indicated conditions are fully understood and met.

General Safety Considerations

| W A R N I N G | • Before this instrument is switched on, make sure it has been properly grounded through the protective conductor of the ac power cable to a socket outlet provided with protective earth contact. |
|---------------|--|
| | Any interruption of the protective (grounding) conductor, inside or outside the instrument, or disconnection of the protective earth terminal can result in personal injury. |
| | • There are many points in the instrument which can, if contacted, cause personal injury. Be extremely careful. Any adjustments or service procedures that require operation of the instrument with protective covers removed should be performed only by trained service personnel. |
| | • No operator serviceable parts inside. Refer servicing to qualified personnel. To prevent electrical shock, do not remove covers. |
| | • If this instrument is used in a manner not specified by Hewlett-Packard Co., the protection provided by the instrument may be impaired. |
| CAUTION | Before this instrument is switched on, make sure its primary power circuitry has been adapted to the voltage of the ac power source. |
| | Failure to set the ac power input to the correct voltage could cause damage to the instrument when the ac power cable is plugged in. |

Symbols Used on the Cable TV Analyzer



The instruction documentation symbol. The product is marked with this symbol when it is necessary for the user to refer to the instructions in the documentation.

"CE" The CE mark is a registered trademark of the European Community. (If this mark is accompanied by a year, it indicates when the design was proven.) 2

Preparing for Use

Preparing for Use

What You'll
Learn in ThisThis chapter describes the process of getting the cable TV analyzer ready to
use when you have just received it. In this chapter you will:Chapter• Perform an initial inspection

- Set the cable TV analyzer for the selected AC power source
- Turn the cable TV analyzer on for the first time
- Perform automatic self-calibration routines

Preparing Your Cable TV Analyzer for Use

Prepare the cable TV analyzer for use by performing the following procedure:

- 1. Perform an initial inspection.
- 2. Set the line voltage selector switch.
- 3. Check the fuse.
- 4. Connect the power cable.
- 5. Turn on the cable TV analyzer for the first time.
- 6. Perform the frequency and amplitude self-calibration routines.
- 7. Perform the tracking-generator self-calibration routine. (For Option 011 only.)

Detailed information for all of the steps in this process are included in this chapter in the following pages.

Reducing Damage Caused by ESD

The following suggestions may help reduce ESD damage that occurs during testing and servicing operations:

- Before connecting any coaxial cable to an cable TV analyzer connector for the first time each day, momentarily ground the center and outer conductors of the cable.
- Personnel should be grounded with a resistor-isolated wrist strap before touching the center pin of any connector and before removing any assembly from the unit.
- Be sure that all instruments are properly earth-grounded to prevent a buildup of static charge.

Refer to the *HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide* for more information about static-safe accessories that can be obtained from Hewlett-Packard.

Step 1. Perform an initial inspection

Inspect the shipping container for damage. If the shipping container or cushioning material is damaged, keep it until you have verified that the contents are complete and you have tested the cable TV analyzer mechanically and electrically.

Table 2-1 contains the accessories shipped with the cable TV analyzer. If the contents are incomplete or if the cable TV analyzer does not pass the verification tests in the calibration guide, notify the nearest Hewlett-Packard office. If the shipping container is damaged or the cushioning material shows signs of stress, also notify the carrier. Keep the shipping materials for the carrier's inspection. The HP office will arrange for repair or replacement without waiting for a claim settlement.

If the shipping materials are in good condition, retain them for possible future use. You may wish to ship the cable TV analyzer to another location or to return it to Hewlett-Packard for service. See "Returning the Cable TV Analyzer for Service" in Chapter 9 of the *HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide* for more information about shipping materials.

NOTE

Complete instructions for installing your cable TV analyzer in an equipment rack are provided in a service note that is included with Options 908 and 909 Rack Mounting Kits.

| Description | HP Part Number | Comments |
|---|----------------------|--|
| Cable TV RF/video measurements personality | HP 85721A | Shipped with analyzer. |
| Cable, 75 $oldsymbol{\Omega}$, BNC | 5062-6452 | Shipped with analyzer. |
| Memory card holder | 9222-1545 | Shipped with analyzer. |
| Memory card, 128 kilobyte | HP 82215A | Shipped with analyzer. |
| Adapter, BNC m to Type F f | 1250-2477 | Shipped with analyzer. (Option 107 only) |
| Coupler | 0955-0704 | Shipped with analyzer. (Option 107 only) |
| Po wer cable | See Table 2-3. | Shipped with analyzer. |
| Reference connector | 1250-1499 | Shipped connected between the 10 MHz REF OUT and the EXT REF IN on the rear panel of the analyzer. |
| Soft carrying pack | 9211-7102 | Shipped with analyzer. |
| Documentation: | | |
| • HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide | * | Shipped with analyzer. |
| • HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide | * | Shipped with analyzer. |
| • HP 8590 E-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Calibration Guide | * | Shipped with analyzer. |
| • HP 8591C Cable TV Analyzer Getting Started and Quick Reference Guide this book | * | Shipped with analyzer. |
| HP 8590 E-Series and L-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Programmer's Guide | * | Shipped with analyzer. |
| HP 8590 E-Series and L-Series Spectrum Analyzers, and HP 8591C Cable TV Analyzer Assembly-Level Repair Service Guide | * | Shipped with analyzer. <i>(Option 915 only)</i> |
| HP 8590 E-Series and L-Series Spectrum Analyzers, and HP 8591C Cable TV Analyzer Component-Level Repair Service Guide | * | Shipped with analyzer. <i>(Option 915 only)</i> |
| st Contact your nearest sales and service center for | current part number. | |

Table 2.1. Accessories Supplied with the Cable TV Analyzer

Getting Started

Step 2. Set the line voltage selector switch The cable TV analyzer is a portable instrument and requires no physical installation other than connection to a power source. Do not connect AC power until you have verified that the line voltage is correct, the proper fuse is installed, and the line voltage selector switch is properly positioned, as described in the following paragraphs. Damage to the equipment could result.

| Characteristic | 115 V Requirement | 230 V Requirement |
|----------------|-------------------|-------------------|
| Input Voltage | 90 to 132 V rms | 198 to 264 V rms |
| Frequency | 47 to 440 Hz | 47 to 66 Hz |
| Po wer | <500 VA, <180 W | <500 VA, <180 W |

Table 2.2. Power Requirements



pu138e

Figure 2.1. Setting the Line Voltage Selector Switch

CAUTIONBefore connecting the cable TV analyzer to the power source, you must set
the rear-panel voltage selector switch correctly to adapt the cable TV analyzer
to the power source. An improper selector switch setting can damage the
cable TV analyzer when it is turned on.

Set the instrument's rear-panel voltage selector switch to the line voltage range (115 V or 230 V) corresponding to the available AC voltage. See Figure 2-1. Insert a small screwdriver or similar tool in the slot and slide the switch up or down so that the proper voltage label is visible.

Getting Started

Step 3. Check the fuse



Figure 2.2. Checking the Line Fuse

The recommended fuse is size 5 by 20 mm, rated F5A, 250 V (IEC approved). This fuse may be used with input line voltages of 115 V or 230 V. Its HP part number is 2110-0709.

With an input line voltage of 115 V an alternate fuse can be used. In areas where the recommended fuse is not available, a size 5 by 20 mm, rated fast blow, 5 A, 125 V (UL/CSA approved) fuse may be substituted. Its HP part number is 2110-0756.

The line fuse is housed in a small container beside the rear-panel power connector. See Figure 2-2. The container provides space for storing a spare fuse, as shown in the figure.

To check the fuse, insert the tip of a screwdriver in the slot at the middle of the container and pry gently to extend the container.

NOTE

The fuse container is attached to the line module; it cannot be removed.

The fuse closest to the cable TV analyzer is the fuse in use. If the fuse is defective or missing, install a new fuse in the proper position and reinsert the fuse container.

Getting Started

| | Step 4. Connect the power cable |
|---------------|--|
| | The cable TV analyzer is equipped with a three-wire power cable, in accordance with international safety standards. When connected to an appropriate power line outlet, this cable grounds the instrument cabinet. |
| W A R N I N G | Failure to ground the cable TV analyzer properly can result in personal injury. Before turning on the cable TV analyzer, you must connect its protective earth terminals to the protective conductor of the main power cable. Insert the main power cable plug only into a socket outlet that has a protective earth contact. DO NOT defeat the earth-grounding protection by using an extension cable, power cable, or autotransformer without a protective ground conductor. |
| | If you are using an autotransformer, make sure its common terminal is connected to the protective earth contact of the power source outlet socket. |
| | Various power cables are available to connect the cable TV analyzer to the types of AC power outlets unique to specific geographic areas. The cable appropriate for the area to which the cable TV analyzer is originally shipped |

appropriate for the area to which the cable TV analyzer is originally shipped is included with the unit. You can order additional AC power cables for use in different areas. Table 2-3 lists the available AC power cables, illustrates the plug configurations, and identifies the geographic area in which each cable is appropriate.

Table 2.3. AC Power Cables Available

| PLUG TYPE * * | CABLE HP PART NUMBER | PLUG DESCRIPTION | CABLE LENGTH CM (INCHES) | CABLE COLOR | FOR USE IN COUNTRY |
|---|--|---|--|--|---|
| 250V | 8120-1351 8120-1703 | Straight [*] BS1363A 90 [°] | 229 (90) 229 (90) | Mint Gray Mint Gray | Great Britain, Cyprus, Nigeria, Singapore, Zimbabwe |
| 250V E L N | 8120-1369 8120-0696 | Straight [*] NZSS198/ASC112 90 [°] | 201 (79) 221 (87) | Gray Gray | Argentina, Australia, New Zealand, Mainland China |
| | 8120-1689 8120-1692 | Straight [*] CEE7-Y11 90° | 201 (79) 201 (79) | Mint Gray Mint Gray | East and West Europe, Central African Republic, United Arab Republic (unpolarized in many nations) |
| | 8120-1348 8120-1538 | Straight [*] NEMA5—15P 90° | 203 (80) 203 (80) | Black Black | United States Canada, Japan (100 V or 200 V), Brazil, Colombia, Mexico, Philippines, Saudia Arabia, Taiwan |
| | 8120-1378 8120-4753 8120-1521 8120-4754 | Straight [*] NEMA5-15P Straight 90° 90° | 203 (80) 230 (90) 203 (80) 230 (90) | Jade Gray Jade Gray Jade Gray Jade Gray | |
| 250V | 8120-5182 8120-5181 | Straight [*] NEMA5-15P 90 [°] | 200 (78) 200 (78) | Jade Gray Jade Gray | Israel |
| | | | | | |
| | | | | | |
| * Part number for plug is industry identifier for plug only. Number shown for cable is HP Part Number for complete cable, including plug. ** E = Farth Ground: L = Line: N = Neutral | | | | | |

FORMAT80

Getting Started

Step 5. Turn on the cable TV analyzer for the first time

When you turn the cable TV analyzer on for the first time, you should perform frequency and amplitude self-calibration routines to generate correction factors and indicate that the unit is functioning correctly. The cable TV analyzer should be allowed to warm up for 30 minutes before performing the self-calibration routines. See "When Is Self-Calibration Needed?" in Chapter 3 for helpful guidelines on how often the self-calibration routines should be performed.

Perform the following steps:

1. Ensure the reference connector is connected between the 10 MHz OUTPUT and EXT REF IN rear-panel connectors. See Figure 2-3.



pu126el

Figure 2.3. Reference Connector

If you wish to use an external 10 MHz source as the reference frequency, disconnect the reference connector from the rear-panel and connect an external reference source to the EXT REF IN connector on the rear panel.

2. Press (LINE).

After a few seconds, the screen displays the firmware revision date in the YYMMDD format. For example, 930522 indicates May 22, 1993. The baud rate (**RS232:** XXXX) is also displayed.

If your cable TV analyzer is equipped with Option 021 (HP-IB interface), the appropriate interface address (HP-IB ADRS: XX) in place of the baud rate, also appears on the screen.

NOTE

Record the firmware date and keep it for reference. If you should ever need to call Hewlett-Packard for service or with any questions regarding your cable TV analyzer, it will be helpful to have the firmware date readily available.

Step 6. Perform the frequency and amplitude self-calibration routines

- 1. To meet cable TV analyzer specifications, allow a 30 minute warmup before attempting to make any calibrated measurements. Be sure to calibrate the cable TV analyzer only *after* the cable TV analyzer has met the operating temperature conditions.
- 2. Connect the 75 Ω coaxial cable between the front-panel CAL OUT and the INPUT 75 Ω connector.

NOTE

Remove all connections to the GATE TRIGGER INPUT rear-panel connector before performing the self-calibration routines.

3. Perform the frequency and amplitude self-calibration routine by pressing (CAL) and CAL FREQ & AMPTD.

During the frequency routine, CAL: SWEEP, CAL: FREQ, and CAL: SPAN and CAL: FM GAIN + OFFSET are displayed as the sequence progresses.

During the amplitude routine, CAL; AMPTD, CAL: 3 dB BW, CAL: ATTEN, and CAL: LOGAMP are displayed as the sequence progresses. CAL: DONE appears when the routine is completed. Any failures or discrepancies produce a message on the screen. See Error Messages.

4. When the frequency and amplitude self-calibration routines have been completed successfully, store the correction factors by pressing CAL_STORE.

The self-calibration routines calibrate the cable TV analyzer by generating correction factors. The softkey CAL STORE stores the correction factors in the area of cable TV analyzer memory that is stored when the cable TV analyzer is turned off; the cable TV analyzer will automatically apply these factors in future measurements. If CAL STORE is not pressed, the correction factors remain in effect until the cable TV analyzer is turned off.

Step 7. Perform the tracking-generator self-calibration routine (Option 011 only)

This procedure is for cable TV analyzers equipped with Option 011 only, the tracking-generator self-calibration routine should be performed prior to using the tracking generator.

NOTE

Since the tracking generator calibration routine depends on the accuracy of the absolute amplitude level of the cable TV analyzer, the cable TV analyzer amplitude calibration should be done prior to using CAL TRK GEN.

1. To calibrate the tracking generator, connect the tracking generator output (RF OUT 75 Ω) to the cable TV analyzer INPUT 75 Ω connector, using an appropriate cable.

NOTE

A low-loss cable should be used for accurate calibration. Use the 75 Ω cable shipped with the cable TV analyzer.

2. Press the following cable TV analyzer keys: (CAL), More 1 of 4,

More 2 of 4, then CAL TRK GEN. TG SIGNAL NOT FOUND will be displayed if the tracking generator output is not connected to the cable TV analyzer input.

3. To store this data in the area of cable TV analyzer memory that is stored when the cable TV analyzer is turned off, press CAL STORE.

When the self-calibration routines have been completed successfully, the cable TV analyzer is ready for normal operation.

Preparing for Use
Preparing Your Cable TV Analyzer for Use
3

Quick Tour

Quick Tour

What You'll Learn in This Chapter

This chapter introduces the basic functions and features of the HP 8591C Cable TV Analyzer. In this chapter, you will:

- Get acquainted with the front-panel and rear-panel features.
- Get acquainted with the menus and softkeys
- Learn about screen annotation
- Learn about the cable TV measurements and system monitor personality
- Make a basic measurement using the calibration signal
- Learn how to improve measurement accuracy by using self-calibration routines.
- Learn how to insert the memory card and about the memory-card battery.
- Learn about the cable TV analyzer battery.



pu115c

Figure 3.1. HP 8591C Cable TV Analyzer

The HP 8591C Cable TV Analyzer is a small, lightweight test instrument that covers the 1 MHz to 1.8 GHz frequency range. The HP 8591C Cable TV Analyzer is a portable instrument ideal for field use. It comes in a durable carrying case that makes it easy to transport and protects it from moisture and dirt. The cable TV analyzer is fully operational within the case so you never need to remove it.

The HP 8591C Cable TV Analyzer comes equipped with the HP 85721A Cable TV Measurements and System Monitor Personality installed. The Cable TV Measurements and System Monitor Personality is a downloadable program (DLP) consisting of measurement routines useful for cable television applications.

Front-Panel Features

The following section provides a brief overview of front-panel features illustrated in Figure 3-2. Refer to HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide for more information about these features.



Figure 3.2. Front Panel Feature Overview

pu114c

| 1 | Active function block | is the space on the screen that indicates the active function. |
|----|---|---|
| 2 | Message block | is the space on the screen where $\ensuremath{\texttt{MEAS}}$ $\ensuremath{\texttt{UNCAL}}$ and the asterisk $ \ensuremath{^*} $ appear. |
| 3 | Softkey labels | are the annotation on the screen next to the unlabeled keys. |
| 4 | Softkeys | are the unlabeled keys next to the screen. |
| 5 | (FREQUENCY), (SPAN), and (AMPLITUDE) | are the three large dark-gray keys that activate the primary cable TV analyzer functions and access menus of related functions. |
| 6 | INSTRUMENT STATE | functions affect the state of the entire cable TV analyzer. |
| 7 | (COPY) | prints or plots screen data. |
| 8 | CONTROL | functions access menus that allow you to adjust the resolution bandwidth, adjust the sweep time, store and manipulate trace data, and control the instrument display. |
| 9 | MARKER | functions control the markers, read out frequencies and amplitudes along the cable TV-analyzer trace, automatically locate the signals of highest amplitude, and keep a signal at the marker position in the center of the screen. |
| 10 | WINDOWS | keys, turn on the windows display mode. |
| 11 | DATA keys, STEP keys, and knob | allow you to change the numeric value of an active function. |
| 12 | INPUT 75 Ω | is the signal input for the cable TV analyzer. |
| 13 | PROBE PWR | provides power for high-impedance ac probes or other accessories. |
| 14 | CAL OUT | provides a calibration signal of 300 MHz at 29 dBmV. |
| 15 | VOL·INTEN | The VOL-INTEN knob changes the brightness of the display and the volume of the internal speaker. |
| 16 | TVIN | Provides the input for the built-in Television-tuner, Option 107. |
| 17 | Memory card reader | reads from or writes to a memory card. |
| 18 | RF OUT 75 Ω | supplies a source output for the built-in tracking generator. For Option 011 the output frequency range of the cable TV analyzer is 1 MHz to 1.8 Ghz. |
| 19 | (LINE) | turns the instrument on and off. |

Rear-Panel Features

The following section provides a brief overview of rear-panel features illustrated in Figure 3-3. Refer to HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide for more information about these features.



Figure 3-3. Rear Panel Feature Overview

| 1 | EXT ALC INPUT | allows the use of an external detector or power meter for automatic leveling control of the tracking generator. <i>For Option 011 only.</i> |
|----|--------------------------|---|
| 2 | TV TRIG OUTPUT (TTL) | provides TV trigger output using TTL and negative-edge triggering. <i>Standard except for Option 701.</i> |
| 3 | GATE TRIGGER INPUT (TTL) | accepts a TTL signal which acts as the gate trigger. For Option 107 only. |
| 4 | TV MON OUTPUT | connector provides a connection for an external television monitor. |
| 5 | EARPHONE | connector provides a connection for an earphone jack instead of using the internal speaker. |
| 6 | GATE OUTPUT (TTL) | provides a TTL signal which indicates gate status when the when the gate is in edge trigger mode. <i>For Option 107 only.</i> |
| 7 | EXT REF IN | accepts an external frequency source to provide the 10 MHz, -2 to +10 dBm frequency reference used by the cable TV analyzer. |
| 8 | 10 MHz REF OUTPUT | provides a 10 MHz, 0 dBm minimum, time-based reference signal. |
| 9 | VOLTAGE SELECTOR | adapts the unit to the power source: 115 V or 230 V. |
| 10 | MONITOR OUTPUT | drives an external monitor with a signal spectrum analyzer display that has a 15.7 kHz horizontal synchronizing rate. |
| 11 | Power input | is the input for the line power source. |
| 12 | AUX INTERFACE | provides a nine-pin "D" subminiature connector for control of external devices. |
| 13 | Interface connectors | interfaces for RS-232 and parallel $ \mbox{standard} $ and HP-IB and parallel $ \mbox{Option O41} $ interface buses. |
| 14 | AUX IF OUTPUT | is a 50 $\Omega,$ 21.4 MHz IF output that is the down-converted signal of the RF input of the cable TV analyzer. |
| 15 | AUX VIDEO OUTPUT | provides detected video output before the analog-to-digital conversion proportional to vertical deflection of the trace. |
| 16 | EXT TRIG INPUT (TTL) | accepts the positive edge of an external voltage input that triggers the cable TV analyzer internal sweep source. |
| 17 | HIGH SWEEP IN/OUT (TTL) | indicates when the cable TV analyzer is sweeping or can be grounded to stop sweeping. |
| 18 | EXT KEYBOARD | connector is provided with the interface connector. |
| 19 | SWEEP OUTPUT | provides a voltage ramp proportional to the sweep and the cable TV analyzer span $ 0\$ V to $10\$ V . |

Data Controls

Data controls are used to change values for functions such as center frequency, start frequency, resolution bandwidth, and marker position.

The data controls will change the active function in a manner prescribed by that function. For example, you can change center frequency in fine steps with the knob, in discrete steps with the step keys, or to an exact value with the number/units keypad. For example, resolution bandwidth, which can be set to discrete values only, is changed to predetermined values with any of the data controls.

Hold Key Deactivate functions with HOLD which is found under the (DISPLAY) key. The active function readout is blanked, indicating that no entry will be made inadvertently by using the knob, step keys, or keypad. (Pressing a function key re-enables the data controls.)

KnobThe knob allows continuous change of functions such as center frequency,
reference level, and marker position. It also changes the values of many
functions that change in increments only. Clockwise rotation of the knob
increases values. For continuous changes, the extent of alteration is
determined by the size of the measurement range; the speed at which the
knob is turned does not affect the rate at which the values are changed.

The knob enables you to change the center frequency, start or stop frequency, or reference level in smooth scrolling action. The smooth scrolling feature is designed to move the trace display to the latest function value as the knob is turned. When either center frequency or reference level is adjusted, the signal will shift right or left or up or down with the rotation of the knob before a new sweep is actually taken. An asterisk is placed in the message block (the upper right-hand corner of the cable TV analyzer display) to indicate that the data on-screen does not reflect data at the current setting.

NOTE

When using the knob to change frequency or amplitude settings, the trace data is shifted. Therefore, when using MAX HOLD A, MAX HOLD B, or MIN HOLD C, moving the center frequency with the knob will not simulate a drifting signal.

Number/Units Keypad

The number/units keypad allows entry of exact values for many of the cable TV analyzer functions. You may include a decimal point in the number portion. If not, the decimal point is placed at the end of the number.

Numeric entries must be terminated with a units key. The units keys change the active function in a manner prescribed by that function. For example, the units keys for frequency span are (HZ), (MHZ), (kHZ), and (HZ), whereas the units for reference level are (+dBm), (-dBm), (mV), and (μV) .

NOTE

If an entry from the number/units keypad does not coincide with an allowed function value (for example, that of a 12 MHz bandwidth), the cable TV analyzer defaults to the nearest allowable value.

Step Keys

The step keys allow discrete increases or decreases of the active function value. The step size depends upon the cable TV analyzer measurement range or on a preset amount. Each press results in a single step change. For those parameters with fixed values, the next value in a sequence is selected each time a step key is pressed. Changes are predictable and can be set for some functions. Out-of-range values or out-of-sequence values will not occur using these keys.

Fine-Focus Control

The fine-focus control is located on the side of the cable TV analyzer. Use the following procedure to adjust the fine-focus control:

- 1. Adjust the front-panel intensity control for a comfortable viewing intensity.
- 2. Use an adjustment tool or small screwdriver to access the fine-focus adjustment. See Figure 3-4. Adjust for a focused display.



Figure 3.4. Adjusting the Fine Focus

Screen Annotation

Figure 3-5 shows an example of the annotation that may appear on a cable TV analyzer screen. The screen annotation is referenced by numbers and is listed in Table 3-1. The function key column indicates which front-panel key or softkey activates the function related to the screen annotation. Refer to the *HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide* for more information on a specific function key.

Refer to Table 3-2 for the screen annotation codes for trace, trigger, and sweep modes.



Figure 3.5. Screen Annotation

Quick Tour Screen Annotation

| ltem | Description | Function Key |
|------|--|--|
| 1 | detector mode | DETECTOR SMP PK |
| 2 | reference level | REF LVL |
| 3 | time and date display | Time Date |
| 4 | screen title | Change Title |
| 5 | RF attenuation | ATTEN AUTO MAN |
| 6 | preamplifier gain | EXTERNAL PREAMPG |
| 7 | external keyboard entry | Refer to "External Keyboard" in the <i>HP 8591C Cable TV</i> Analyzer Spectrum Analyzer Reference User's Guide. |
| 8 | marker readout | (MKR) , $(MKR \rightarrow)$, $(MKR FCTN)$, or $(PEAK SEARCH)$ |
| 9 | measurement uncalibrated or function-in-progress messages | (AUTO COUPLE) |
| 10 | service request | See Appendix A. |
| 11 | remote operation | See " LOCAL " in the <i>HP 8591C Cable TV Analyzer</i> Spectrum Analyzer Reference User's Guide. |
| 12 | frequency span or stop frequency | (SPAN), STOP FREQ |
| 13 | sweep time | SWP TIME AUTO MAN |
| 14 | frequency offset | FREQ OFFSET |
| 15 | video bandwidth | VID BW AUTO MAN |
| 16 | resolution bandwidth | RES BW AUTO MAN |
| 17 | center frequency or start frequency | CENTER FREQ, START FREQ |
| 18 | threshold | THRESHLD ON OFF |
| 19 | correction factors on | CORRECT ON OFF |

Table 3-1. Screen Annotation

| ltem | Description | Function Key |
|------|---------------------------------|--|
| 20 | amplitude correction factors on | See "Using Amplitude-Correction Factors" in the <i>HP 8591C</i> |
| | | Cable TV Analyzer Spectrum Analyzer Reference User's |
| | | Guide. |
| 21 | trigger | TRIG |
| 22 | trace mode | (TRACE) |
| 23 | video average | VID AVG ON OFF |
| 24 | display line | DSP LINE ON OFF |
| 25 | amplitude offset | REF LVL OFFSET |
| 26 | amplitude scale | SCALE LOG LIN |
| 27 | active function block | Refer to the description of the softkey function that was activated. |

Table 3.1. Screen Annotation (continued)

Table 3.2. Screen Annotation for Trace, Trigger, and Sweep Modes

| Trace Mode | Trigger Mode | Sweep Mode |
|-------------------------------|--------------|------------------|
| W = clear write traces A/B/C | F = free run | C = continuous |
| M = maximum hold traces A/B | L = line | S = single sweep |
| V = view traces A/B/C | V = video | |
| S = store blank traces A/B/C | E = external | |
| M = minimum hold trace C | T = TV | |

More about annotations

- Item 21 refers to the trigger and sweep modes of the cable TV analyzer. The first letter ("F") indicates the cable TV analyzer is in free-run trigger mode. The second letter ("S") indicates the cable TV analyzer is in single-sweep mode.
- Item 22 refers to the trace modes of the cable TV analyzer. The first letter ("W") indicates that the cable TV analyzer is in clear-write mode. The second letter is "A," representing trace A. The trace B trace mode is "SB", indicating trace B ("B") is in the store-blank mode ("S"). The trace mode annotation for trace C is displayed under the trace mode annotation of trace A. In Figure 3-5, the trace C trace mode is "SC", indicating trace C ("C") is in the store blank mode ("S").
- The WINDOWS display mode splits the screen into two separate displays. Only one of these
 displays is active at a time. The currently active window will have a solid line around the graticule
 rather than a broken line. The complete annotation is not available for each window because of
 space limitations.
- The display will be compressed slightly when using the PAL or NTSC format for the MONITOR OUTPUT, instead of the normal format. The PAL and NTSC formats have less vertical resolution than the cable TV analyzer display. The top and bottom of the cable TV analyzer display are compressed slightly so that all of the information can be fit into the size required by the MONITOR OUTPUT.

Menu and Softkey Overview

The keys labeled AMPLITUDE, FREQUENCY, CAL, and MKR are all examples of front-panel keys. Pressing most front-panel keys accesses menus of functions that are displayed along the right side of the display. These menus are called softkey menus.

Softkey menus list functions other than those accessed directly by the front-panel keys. To activate a function on the softkey menu, press the unlabeled key immediately to the right of the annotation on the screen. The unlabeled keys next to the annotation on the display screen are called softkeys.

Front-panel keys are designated with a box around the key label, for example, (AMPLITUDE); softkeys are designated by shading on the key label, for example, **REF LVL**. The softkeys that are displayed depend on which front-panel key is pressed and which menu level is enabled.

If a softkey function's value can be changed, it is called an active function. The function label of the active function appears in inverse video. For example, press (AMPLITUDE). This calls up the softkey menu of related amplitude functions. Note the function labeled **REF_LVL** appears in inverse video. **REF_LVL** also appears in the active function block, indicating that it is the active amplitude function and can now be changed using any of the data entry controls.

A softkey with ON and OFF in its label can be used to turn the softkey's function on or off. To turn the function on, press the softkey so that ON is underlined. To turn the function off, press the softkey so that OFF is underlined.

A function with AUTO and MAN in the label can either be auto-coupled or have its value manually changed. The function's value can be changed manually by pressing the softkey until MAN is underlined, and then changing its value with the numeric keypad, knob, or step keys. To auto-couple a function, press the softkey so that AUTO is underlined.

A summary of all front-panel keys and their related softkeys can be found in the HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide.

Cable TV Measurements and System Monitor Personality

The HP 85721A Cable TV Measurements and System Monitor Personality is a downloadable program (DLP) consisting of measurement routines useful for cable television applications. The cable TV RF/video measurements personality is installed in the HP 8591C Cable TV Analyzer.

The tests included in the cable TV measurements and system monitor personality are:

- Visual carrier level and visual-to-aural level difference test.
- Visual carrier frequency and visual-to-aural frequency difference test.
- Carrier-to-noise test.
- Composite second order (CSO) test.
- Composite triple beat (CTB) test.
- Hum test.
- Depth of modulation test.
- Color tests (chroma-luma delay inequality, differential gain, differential phase).
- In-channel frequency response test.
- FM deviation test.
- Cross modulation test.
- System frequency response test.
- View ingress test.
- View co-channel test.

Each measurement typically performs the following tasks:

- Sets the analyzer to the state required for making the measurement.
- Prompts for any adjustments or values.
- Performs the measurement.
- Displays the measurement results.
- Provides the ability to print the measurement results for FCC proof of performance.

Refer to the *HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide* for detailed information about cable TV measurements.

Making a Measurement

| CAUTION | Do not exceed the maximum input power. |
|---------|--|
| | The maximum input power for the HP 8591C is +72 dBmV (0.2 watt) continuous, 100 VDC (with 10 dB or more attenuation). |
| | Let's begin using the cable TV analyzer by measuring an input signal. Since the 300 MHz calibration signal (CAL OUT) is readily available, we will use it as our input signal. |
| | You cannot hurt the cable TV analyzer by using the calibration signal and pressing any of the keys described in this section. Don't be afraid to play with the knob, step keys, or number/units keypad. (If you have experimented with other keys and wish to return to a known state, press the green (<u>PRESET</u> key.) |
| | 1. First, turn the instrument on by pressing (LINE). Wait for the power-up process to complete. |
| | 2. Press the green PRESET key. |
| | 3. Connect the cable TV analyzer CAL OUT to the INPUT 75 Ω with an appropriate cable. |
| | 4. Set the frequency. |
| | Press the (FREQUENCY) key. CENTER appears on the left side of the screen, indicating that the center-frequency function is active. The CENTER FREQ softkey label appears in inverse video to indicate that center frequency is the active function. The active function block is the space on the screen within the graticule where the center frequency messages appear. Functions appearing in this block are active: their values can be changed with the knob, step keys, or number/units keypad. Set the center frequency to 300 MHz with the DATA keys by pressing 300 (MHz). The knob and step keys can also be used to set the center frequency. |
| | 5. Set the span. |
| | Press (SPAN). SPAN is now displayed in the active function block, and the SPAN softkey label appears in inverse video to indicate it is the active function. Reduce the span to 20 MHz by using the knob, pressing the down key ((\square), or pressing 20 (MHz). |

6. Set the amplitude.

When the peak of a signal does not appear on the screen, it may be necessary to adjust the amplitude level on the screen. Press <u>AMPLITUDE</u>. REF LEVEL .0 dBm appears in the active function block. The REF LVL softkey label appears in inverse video to indicate that reference level is the active function. The reference level is the top graticule line on the display and is set to 0.0 dBm. Changing the value of the reference level changes the amplitude level of the top graticule line.

If desired, use the reference level function to place the signal peak on the screen using the knob, step keys, or number/units keypad. (Marker functions determine the frequency and amplitude of a signal.)

Figure 3-6 demonstrates the relationship between center frequency and reference level. The box in the figure represents the cable TV analyzer screen. Changing the center frequency changes the horizontal placement of the signal on the screen. Changing the reference level changes the vertical placement of the signal on the screen. Increasing the span increases the frequency range that appears horizontally on the screen.



pu119c

Figure 3.6. Relationship between Frequency and Amplitude

7. Set the marker.

You can place a diamond-shaped marker on the signal peak to find the signal's frequency and amplitude.

To activate a marker, press the (MKR) key (located in the MARKER section of the front panel). The MARKER NORMAL label appears in inverse video to show that the marker is the active function. Turn the knob to place the marker at the signal peak.

You can also use the (PEAK SEARCH) key, which automatically places a marker at the highest point on the trace.

Readouts of marker amplitude and frequency appear in the active function block and in the upper-right corner of the display. Look at the marker readout to determine the amplitude of the signal.

If another function is activated, the frequency and amplitude can still be identified by looking at the marker readout in the upper-right corner of the screen.

Measurement Summary

- 1. Connect the cable TV analyzer CAL OUT to the INPUT 75 Ω and press the (PRESET) key.
- 2. Set the center frequency by pressing the following keys: (FREQUENCY), 300 (MHz).
- 3. Set the span by pressing the following keys: (SPAN), 20 (MHz).
- 4. The calibration signal is 20 dB (two graticule divisions) below the top of the screen using these cable TV analyzer settings. If desired, adjust the reference level: press (AMPLITUDE) to activate the reference level, and use the knob or step keys to change the reference level.
- 5. Determine the amplitude and frequency of the signal. You can either press (PEAK SEARCH) or press (MKR) and move the marker to the signal peak. Read the amplitude and frequency. The display screen should look like the one in Figure 3-7. Frequency is displayed horizontally, and amplitude (power) is displayed vertically.



Figure 3.7. Reading the Amplitude and Frequency

Improving Accuracy with Self-Calibration Routines

Data from the self-calibration routine is necessary for cable TV analyzer operation. Executing the self-calibration routine regularly ensures that the cable TV analyzer is using current calibration data that improves the cable TV analyzer frequency and amplitude accuracy. Press the CAL key to view the self-calibration routine menus. The last softkey on this menu, labeled **More 1 of 4**, provides access to additional self-calibration functions. For more detailed information on the self-calibration softkeys, refer to Chapter 7 of the *HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide*.

The self-calibration routines add correction factors to internal circuitry. The addition of the correction factors is required to meet frequency and amplitude specifications.

When the correction factors are added to internal circuitry, CORR (corrected) appears on the left side of the screen.

Warmup Time

In order for the cable TV analyzer to meet its specifications, allow the cable TV analyzer to warm up for 30 minutes after being turned on before attempting to make any calibrated measurements. Be sure to calibrate the cable TV analyzer only *after* it has met operating temperature conditions.

Calibrating the Analyzer

The cable TV analyzer frequency and amplitude self-calibration routines are initiated by the CAL FREQ & AMPTD softkey in the menu located under the (CAL) key.

1. To calibrate the instrument, connect the cable TV analyzer CAL OUT to the INPUT 75 Ω connector with an appropriate cable.

NOTE

A low-loss cable should be used for accurate calibration. Use the 75 Ω cable shipped with the cable TV analyzer.

- 2. On the cable TV analyzer, press (CAL) and CAL FREQ & AMPTD. Cal signal not found will be displayed if CAL OUT is not connected to the cable TV analyzer input. The frequency and amplitude self-calibration functions take approximately 5 minutes to finish (9 minutes with Option 130), at which time the internal adjustment data is in working RAM.
- 3. To save this data in the area of cable TV analyzer memory that is saved when the cable TV analyzer is turned off, press CAL STORE.

NOTE

| To interrupt the calibration routines started by ${\tt CAL}$ ${\tt FREQ}$, ${\tt CAL}$ | AMPTD, or |
|---|-----------------------------------|
| CAL FREQ & AMPTD, press (PRESET), (CAL), More 1 | of 4, and CAL FETCH. |
| CAL FETCH retrieves the previous correction factors. Improperly i | interrupting the self-calibration |
| routines may result in corrupt correction factors. (If this occurs, press rerun the frequency and amplitude self-calibration routines.) | CAL FREQ & AMPTD to |

The frequency and amplitude self-calibration functions can be done separately by using the CAL FREQ or CAL AMPTD softkeys instead of CAL FREQ & AMPTD.

NOTE

If the frequency calibration **CAL FREQ** and the amplitude calibration **CAL AMPTD** self-calibration routines are used, the frequency calibration should be performed before the amplitude calibration, unless the frequency data is known to be accurate.

If the self-calibration routines cannot be performed, see "Check the basics" in Chapter 9 of the HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide.

Calibrating the Tracking Generator (Option 011 only)

In order for the tracking generator to meet its specifications, allow the cable TV analyzer to warm up for 30 minutes after being turned on before attempting to make any calibrated measurements. Be sure to calibrate the cable TV analyzer and the tracking generator only *after* the cable TV analyzer has met operating temperature conditions.

NOTE

Since the tracking-generator self-calibration routine uses the absolute amplitude level of the cable TV analyzer, the cable TV analyzer amplitude should be calibrated prior to using **CAL TRK GEN**.

1. To calibrate the tracking generator, connect the tracking generator output (RF OUT 75 Ω) to the cable TV analyzer INPUT 75 Ω connector, using an appropriate cable and adapters.

NOTE

A low-loss cable should be used for accurate calibration. Use the 75 Ω cable shipped with the cable TV analyzer.

- Press the following cable TV analyzer keys: CAL, More 1 of 4, More 2 of 4, and CAL TRK GEN. TG SIGNAL NOT FOUND will be displayed if the tracking generator output is not connected to the cable TV analyzer input.
- 3. To save this calibration data in the area of cable TV analyzer memory that is saved when the cable TV analyzer is turned off, press (CAL) then CAL_STORE.

When Is Self-Calibration Needed?

While it is difficult to provide general advice for your specific measurement needs, the following suggestions may help you decide when to use the self-calibration features:

- 1. Perform the frequency and amplitude self-calibration routines whenever the instrument experiences significant environmental changes such as temperature (± 5 °C), humidity, shock, or vibration (such as may occur during shipping or transport). This is especially important if the frequency and amplitude self-calibration routines were performed last in a different environment.
- 2. If the environment is relatively stable (for example, a lab environment), use **CAL FREQ & AMPTD** monthly. After being turned off overnight, the cable TV analyzer will need to warm up, but should not require self-calibration.
- 3. To achieve optimal amplitude accuracy for relative measurements:
 - a. Keep the cable TV analyzer in a stable environment.
 - b. Use CAL FREQ & AMPTD before beginning the first measurement.
 - c. Keep the cable TV analyzer turned on between measurements.
 - d. Do not use CAL FREQ & AMPTD before subsequent measurements (the amplitude drift is normally smaller than the self-calibration uncertainty).
- 4. If you change the input signal for EXT REF IN, run the frequency and amplitude self-calibration routines using CAL OUT. Amplitude calibration is required to improve IF centering.
- 5. If accurate self-calibration is needed temporarily in a different environment, use CAL FREQ & AMPTD, but do not press CAL STORE. The temporary correction factors will be used until the cable TV analyzer is turned off or until CAL FETCH is pressed.

Printing or Plotting

You may wish to obtain a permanent record of data displayed on the cable TV analyzer screen. This can be done using the \bigcirc where \bigcirc key of the analyzer, and a printer or plotter.

Refer to the *HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide* for more information about printing or plotting.

Memory Card Insertion and Battery Replacement

Use the following information to ensure that the memory card is inserted correctly. Improper insertion causes error messages to occur, but generally does not damage the card or instrument. Care must be taken, however, not to force the card into place. The cards are easy to insert when installed properly.

- 1. Locate the arrow printed on the card's label.
- 2. Insert the card with its arrow matching the raised arrow on the bezel around the card-insertion slot. See Figure 3-8.



pu122c

Figure 3.8. Inserting the Memory Card

3. Press the card into the slot. When correctly inserted, about 19 mm (0.75 in) of the card is exposed from the slot.

Changing the Memory Card Battery

It is recommended that the memory card battery be changed every 2 years. The battery is a lithium commercial CMOS type battery, part number CR 2016 or HP part number 1420-0383.

NOTE

The minimum lifetime of the battery (under ordinary conditions) is more than 2 years.

The date that the memory card battery was installed is either engraved on the side of the memory card or written on a label on the memory card.

If the memory card does not have a label with the date that the battery was installed, use the date code engraved on side of the memory card. The date code engraved on the memory card consists of numbers and letters engraved in the black plastic on the side of the memory card. See Figure 3-9. The first number indicates the year, the following two characters indicate the month, and the following number indicates the week in the month that the memory card battery was installed. For example, 80C3 indicates the battery was installed in the third week in October in 1988.

Quick Tour Memory Card Insertion and Battery Replacement



pu121c

Figure 3.9. Memory Card Battery Date Code Location

| Procedure to Change the Memory Card Battery | The battery is located beside the card's write-protect switch on the end opposite the connector. |
|--|---|
| C A U T I O N | The battery power enables the memory card's memory to retain data. You can lose the data when the battery is removed. Replace the battery while the card is installed in a powered-up instrument. |
| | 1. Locate the groove along the edge of the battery clip. See Figure 3-10. |
| | 2. Gently pry the battery clip out of the card. The battery fits within this clip. |
| | 3. Replace the battery, making sure the plus (+) sign on the battery is on the same side as the plus (+) sign on the clip. |
| | 4. Insert the battery clip into the memory card, holding the clip as oriented in Figure 3-10. (Face the "open" edge of the clip toward the write-protect switch on the memory card.) |
| | 5. Write the date that the battery was replaced on the memory card label. This will help you to remember when the battery should be replaced. |





pu125c

Figure 3-10. Memory Card Battery Replacement

Analyzer Battery Information

The HP 8591C cable TV analyzers use a 3.6 V lithium battery to enable the cable TV analyzer memory to retain data. The date when the battery was installed is on a label on the rear panel of the cable TV analyzer. See Figure 3-11.

The minimum life expectancy of the battery is 8 years at 25 °C, or 1 year at 55 °C. If you experience problems with the battery or the recommended time period for battery replacement has elapsed, see "Returning the Cable TV Analyzer for Service" in Chapter 9 of the HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide.

If you wish to replace the battery yourself, you can purchase the service documentation that provides all necessary test and maintenance information. The battery is soldered onto the cable TV analyzer processor board.

You can order the service documentation for an HP 8591C cable TV analyzer through your HP Sales and Service office. The documentation is described under "Options" in the front of the HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide.

After replacing the cable TV analyzer battery, write the date of battery replacement on the rear-panel label.

| E | BATTERY LIFE |
|-------------|-----------------------------------|
| \triangle | 8 YEARS AT 25°C 1 YEAR AT 55°C |
| | INSTALLED |
| | |

pu132e

Figure 3.11. Rear Panel Battery Information Label

4

Softkey Descriptions

Softkey Descriptions

What You'll Find in This Chapter

This chapter describes functions and controls of the cable TV analyzer. The front-panel keys and softkey functions are listed alphabetically followed by a brief description. Refer to the HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide and HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide for more detailed information.

Brief descriptions for service functions have also been included in this chapter. However, for more detailed descriptions and information about the use of each function, refer to the HP 8590 E-Series and L-Series Spectrum Analyzers, and HP 8591C Cable TV Analyzer Assembly-Level Repair Service Guide. A listing of all service calibration and service diagnostic functions has been provided in the following "Service Functions" section.

All softkeys and their relationship to the front-panel keys are illustrated in Chapter 5.

Service Functions

Two types of functions are available for service use only:

- Service calibration functions
- Service diagnostic functions

These service functions are designed for service use only. However, brief descriptions for each function are provided in this chapter. For more detailed descriptions and information about the use of each function, refer to the service guide.

Service documentation can be obtained by ordering Option 915 through your HP Sales and Service office. Option 915 is described in more detail in the HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide.

Service Calibration Functions

Service Cal accesses the following service calibration softkeys:

CAL TIMEBASE EDIT FLATNESS EXECUTE TITLE EXIT Flatness Data IDNUM INIT FLT SET ATTN ERROR STOR PWR ON UNITS

Quick Reference

Service Diagnostic Functions

Service Diag accesses the following diagnostic softkeys:

 ϕ LOCK ON OFF +10V REF DETECTOR -10V REF DETECTOR 2v REF DETECTOR (Option 011 only) ALC TEST ANALYZER GAINS AUXB BINARY SPAN COARSE TUNE DAC DACS DISPLAY CAL DATA DROOP FINE TUNE DAC FM COIL DRIVE FM GAIN
Softkey Descriptions
Service Functions

Service Diagnostic Functions (continued) FM OFFST FM SPAN FREQ DIAG FRQ DISC NORM OFF GND REF DETECTOR MAIN COIL DR MAIN SPAN STP GAIN ZERO SWEEP RAMP SWEEP TIME DAC X FINE TUNE DAC

Cable TV Analyzer Functions

| % AM ON OFF | determines the percent of amplitude modulation of the largest displayed signal and its two sidebands. The sidebands are assumed to be entirely from amplitude modulation. If there are differences in the sideband amplitude, the larger value is used. The measurement runs continuously, re-executing at the end of each sweep. |
|----------------------|--|
| | Key Access: (MEAS/USER) |
| φ LOCK ON OFF | turns off the analyzer phase locking. The counter is turned off so frequency correction is no longer done at the start of each sweep. Turning the phase locking off increases measurement speed, but it decreases frequency accuracy. This is a service diagnostic function and is for service use. |
| | Key Access: (CAL) |
| +10V REF DETECTOR | displays the output of the +10 V reference from the A7 analog interface assembly as a horizontal line at the top graticule. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| -10V REF DETECTOR | displays the output of the -10 V reference from the A7 analog interface assembly as a horizontal line at the bottom graticule. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| 2v REF DETECTOR | displays the output of the 2 V reference produced on the A16 processor/video assembly as a horizontal line at the top graticule. This is a service diagnostic function and is for service use only. Key Access: (CAL) |

Quick Reference

| 9 kHz EMI BW | allows a 6 dB resolution bandwidth of 9 kHz. This bandwidth is useful when performing electromagnetic interference (EMI) measurements. |
|-------------------------------|---|
| | Key Access: (BW) |
| 120 kHz EMI BW | allows a 6 dB resolution bandwidth of 120 kHz. This bandwidth is useful when performing electromagnetic interference (EMI) measurements. |
| | Key Access: BW |
| 200 Hz EMI BW | For Option 130 only. allows a 6 dB resolution bandwidth of 200 Hz. This bandwidth is useful when performing electromagnetic interference (EMI) measurements. |
| | Key Access: BW |
| A<>B | exchanges the contents of the trace A register with the trace B register and puts trace A in view mode. |
| | Key Access: (TRACE) |
| $A-B \rightarrow A$ ON OFF | when ON is underlined, subtracts the data in trace B from the measured data in trace A. A minus sign $(-)$ appears between the trace A status and the trace B status in the screen annotation while the function is active. |
| | Key Access: (TRACE) |
| $A \rightarrow C$ | copies trace A into trace C. Key Access: (TRACE) |

| ABCDEF | accesses the softkey menu for selecting screen title characters A through F. |
|--------------------|--|
| | Key Access: (CAL), (CONFIG), (DISPLAY), (RECALL), or (SAVE) |
| ABORT | allows you to stop the measurement. |
| | Key Access: SYSTEM MEAS |
| ACPGRAPH ON OFF | turns the adjacent channel power graph ON or OFF. With ACPGRAPH ON, the ACP ratio graph is calculated and displayed and the numeric results are not displayed. The value of the ACP ratio is displayed at the selected marker frequency offset from the center frequency. This graph function is used after doing an adjacent channel power measurement with the ADJ CHAN POWER or ADJ CHAN PWR extd softkeys. |
| ADJ CHAN POWER | Key Access: (MEAS/USER) measures the power leakage into the adjacent channels and calculates the adjacent channel power ratio of both the upper and lower channels, with respect to the total power. The total power of the carrier is displayed. The adjacent channel power ratios of both channels are also displayed and the larger number is marked with a \land . (Vertical lines on the display indicate the bandwidth edges of the three different channels being measured.) Key Access: (MEAS/USER) |

ADJ CHAN PWR extd

measures the power leakage into the adjacent channels and calculates the adjacent channel power ratio of both the upper and lower channels, with respect to the reference channel. ADJ CHAN POWER extd has an extended dynamic

range compared with ADJ CHAN POWER. The extended range is measured by taking two different sweeps with different reference levels and combining the trace data. The displayed dynamic range is 104 dB and the log scale is set to 13 dB/ division.

Key Access: MEAS/USER

AIR

allows you to select the AIR tune configuration. The AIR is also called off-the-air or over-the air tune configuration. It refers to signals that are broadcast over the air and received with an antenna. The AIR frequency assignments are defined in the Federal Communications Commission's (FCC) channel identification plan, part 76.612.

Key Access: TV RECEIVER

ALC MTR Option 011 only. INT XTAL activates the auto

activates the automatic leveling control (ALC) function for internal (INT) leveling or external (XTAL or MTR) leveling. The external leveling input (located on the rear panel of the analyzer) can be used with a power meter or crystal that has a positive or negative voltage output. See specifications and characteristics in your calibration guide for the leveling input characteristics. External leveling increases the amplitude accuracy by improving the effective source match. The meter (MTR) position narrows ALC loop bandwidth so an HP power meter can be used.

Key Access: (AUX CTRL)

ALC TEST Option 011 only. breaks the leveling loop of the automatic leveling control in the tracking generator. This is a service diagnostic function and is for service use only. Refer to the service guide for more information. Key Access: (CAL) ALL DLP saves all the downloadable programs and key definitions \rightarrow CARD that are in analyzer memory onto the memory card. If the downloadable program was stored using a prefix, the file name for the downloadable program consists of d(prefix)_(register number). If no prefix was specified, the data is stored with the file name d_(register number). Key Access: (SAVE) ALL allows you to select all tests for the currently selected TESTS channel. Key Access: SYSTEM MEAS Amp Cor accessed by (CAL). Amp Cor accesses the menus for entering and editing the current amplitude-correction factors. Key Access: (CAL) AMP COR when accessed by (SAVE), AMP COR stores the current

when accessed by <u>(SAVE)</u>, <u>AMP</u> COR stores the current amplitude-correction factors table in analyzer memory or on the memory card. When accessed by <u>(RECALL)</u>, <u>AMP</u> COR recalls the amplitude-correction factors table from either analyzer memory or the memory card. Amplitude-correction factors are saved with an "a" before the memory-card file name. Screen titles are not recalled with the data. Refer to "To Save a Limit-Line Table or Amplitude Correction Factors" in Chapter 3 for more information.

Key Access: (RECALL) or (SAVE)

| AMP COR | |
|-----------|--|
| ON OFF | turns the current amplitude-correction factors on or off. |
| | Key Access: CAL |
| AMPLITUDE | |
| | activates the reference level function and accesses the amplitude menu. The softkeys accessed when you press (<u>AMPLITUDE</u>) change reference level, input attenuation, vertical scale, mixer level, amplitude units, input impedance, and amplitude offset. |
| Amptd | |
| Units | accesses the softkeys that change the amplitude units. The amplitude units can be changed by pressing dBm, dBmV, |
| | dBuV, Volts, or Watts. |
| | Key Access: (AMPLITUDE) |
| ANALOG+ | |
| ON OFF | turns on the analog+ display mode. This is a digital implementation of an analog display, combining the advantages of both types of displays. The display is made up of 401 horizontal points or trace elements. In the analog+ display mode each trace element can display from 1 to 40 dots, or measurements. |
| | The following table shows commands that are not available with Analog+ operation. |

| Command | Description |
|----------------------------|--|
| A <> B | exchanges trace A and trace B |
| $A-B \rightarrow A$ on off | puts the difference between trace A and trace B into trace A |
| $B-DL \rightarrow B$ | puts the difference between trace B and the display line in trace B |
| CLEAR WRITE B | initially erases trace B and then displays it continuously |
| CLEAR WRITE C | initially erases trace C and then displays it continuously |
| DEMOD ON OFF | turns AM or FM demodulation on and off |
| FFT | changes zero span data to the frequency domain using an FFT |
| MKR CNT ON OFF | turns the marker counter on and off |
| MK PAUSE ON OFF | stops the analyzer sweep at the marker for the defined of the dwell time |
| MAX HOLD A | updates trace A with the maximum level at each point, each sweep |
| MAX HOLD B | updates trace B with the maximum level at each point, each sweep |
| MIN HOLD C | updates trace C with the minimum level at each point, each sweep |
| NORMLIZE ON OFF | the difference between A and B is added to the display line and put into trace A |
| MK TRACK ON OFF | moves the signal closest to the marker to center screen and keeps it centered |
| SPAN ZOOM | finds the highest signal peak, turns on marker track, and activates span |
| THRESHLD ON OFF | sets a lower boundary to the active trace |
| TRACKING PEAK | starts a routine that adjusts tracking to peak the tracking generator signal |
| VID AVG ON OFF | starts a routine that digitally averages display signals and noise |
| VIEW A | holds and displays the trace A data when used after blanking trace A |
| VIEW B | holds and displays the trace B data |
| VIEW C | holds and displays the trace C data |

Table 4·1. Commands Not Available with Analog + Operation

Key Access: (DISPLAY)

| ANALYZER ADDRESS | <i>Option 041 only.</i> allows you to set the HP-IB address of the analyzer. The analyzer address is set to 18 by pressing DEFAULT CONFIG . |
|---------------------|--|
| ΔΝΔΙΥΖΕΒ | Key Access: (CONFIG) |

ANALYZER GAINS

displays the current value of the gain for various functional blocks within the analyzer. The gain values will vary depending on the current analyzer settings. This is a service diagnostic function and is for service use only. Refer to the service guide for more information.

Key Access: (CAL)

Analyzer Input

allows you to select an external pad, an external preamplifier, and total power at the input to the analyzer. An external 50 Ω to 75 Ω pad is used for analyzers with 50 Ω input impedances. The **EXT PAD YES NO** function compensates for amplitude loss caused by an external pad.

An external preamplifier is used for the Carrier-to-Noise measurement.

The external pad and preamplifier settings are stored in non-volatile analyzer memory. This means that the analyzer retains the selection, even when power is turned off, until a new selection is made.

The TTL PWR **C** INPUT measures the total input power and displays the result at the bottom of the screen.

Press CABLE TV ANALYZER, Setup to access Analyzer Input.

Key Access: CABLE TV ANALYZER

| ANNOTATN ON OFF | turns the screen annotation on or off. However, softkey annotation will remain on the screen. The screen annotation may not be required for prints or plots, or during remote operation. Refer to the HOLD softkey description regarding softkey annotation. |
|--------------------|---|
| | Key Access: (DISPLAY) |
| APND | |
| CAT ITEM | starts the DLP editor function and allows the highlighted item from the catalog of analyzer memory to be appended to the end of the item that is currently in the analyzer DLP editor memory. If the item to be appended will not fit in the available memory space, it will not be appended. |
| | Key Access: (RECALL) or (SAVE) |
| ATTEN AUTO MAN | sets the input attenuation in 10 dB increments. The analyzer input attenuator, which is normally coupled (linked) to the reference level control, reduces the power level of the analyzer input signal at the input mixer. The attenuator is recoupled when AUTO is underlined. |
| To prevent dama | age to the input mixer, the power level at the input mixer |

CAUTION

To prevent damage to the input mixer, the power level at the input mixer must not exceed +72 dBmV. To prevent signal compression, power at the input to the input mixer must be kept below 39 dBmV.

NOTE

To protect the mixer from possible damage, O dB RF attenuation (no input power reduction to the mixer) can be selected only from the number/units keypad.

Key Access: (AMPLITUDE) or (AUTO COUPLE)

AUTO

ALL

couples the following functions: resolution bandwidth, video bandwidth, attenuation, sweep time, center-frequency step, video bandwidth, and video-bandwidth to resolution-bandwidth ratio.

Key Access: (AUTO COUPLE)

(AUTO COUPLE)

accesses the softkey menu of functions that can be coupled. (Coupled functions are functions that are linked: if one function is changed, the other function is changed.) The functions that can be auto-coupled are listed below:

- Resolution bandwidth couples to span.
- Video bandwidth couples to resolution bandwidth when the analyzer has a video-bandwidth to resolution-bandwidth ratio of 0.3.
- Sweep time couples to span, resolution bandwidth, and video bandwidth.
- RF attenuation couples to reference level.
- Center frequency step size couples to 10% of span.

Aux Conn Control

accesses the softkey menu used to control the auxiliary outputs and input. The auxiliary outputs are controlled by pressing CNTL A 0 1, CNTL B 0 1, CNTL C 0 1,

and CNTL D O 1. The status of the auxiliary input (control line I), can be displayed on the analyzer screen with DISPLAY CNTL I.

Key Access: (AUX CTRL)

| | accesses the softkey menu used for control of the auxiliary interface connector. (AUX CTRL) accesses demodulation functions. |
|----------------------|---|
| | For Option 011. It accesses tracking generator functions. |
| AUXB | displays the voltage level present at an unused input to the Test Point MUX circuitry located on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| $B \rightarrow C$ | |
| | copies trace B into trace C. |
| | Key Access: (TRACE) |
| B <> C | exchanges the contents of trace B with trace C and puts trace B in view mode. |
| | Key Access: (TRACE) |
| $B-DL \rightarrow B$ | subtracts the display line from trace B and places the result in trace B. The B-DL \rightarrow B function is a math operation. See the A-B \rightarrow A ON OFF softkey description for information about math operations. |
| | Key Access: (TRACE) |
| BAUD | |
| RATE | sets the data transmission speed. (Also see the description for the COPY key.) The baud rate is set to 1200 by pressing DEFAULT CONFIG. |
| | Key Access: (CONFIG) |

| BINARY SPAN | displays the output of the span DAC that is located on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
|----------------|---|
| | Key Access: CAL |
| BLANK A | stores the amplitude data for trace A and removes it from the screen. The trace A register will not be updated as the analyzer sweeps. |
| | Key Access: (TRACE) |
| BLANK B | stores the amplitude data for trace B and removes it from the screen. The trace B register will not be updated as the analyzer sweeps. |
| | Key Access: (TRACE) |
| BLANK C | stores the amplitude data for trace C and removes it from the screen. The trace C register will not be updated as the analyzer sweeps. |
| | Key Access: (TRACE) |
| BLANK CARD | deletes all the files from the memory card. Pressing BLANK CARD displays the message: If you are sure, |
| | press key again to purge data. Press BLANK CARD again if you want to delete all files from the memory card. |
| | Key Access: (Config) |

| (BW) | |
|-------|---|
| | activates the resolution bandwidth function and accesses the softkeys that control the bandwidth |
| | functions: RES BW AUTU MAN, VID BW AUTU MAN, |
| | VBW/RBW RATIO, VID AVG ON OFF, and the |
| | EMI BW Menu . (Also see the RES BW AUTO MAN softkey description.) |
| (CAL) | |
| | accesses the softkey menus used for the self-calibration, service-diagnostics, and service-calibration functions. For more information about self-calibrating the analyzer, see "Improving Accuracy with Self-Calibration Routines" in Chapter 2. |
| CAL | |
| AMPTD | initiates an amplitude self-calibration routine. Connect CAL OUT to the analyzer input before pressing CAL AMPTD. Remove the cable from the rear panel GATE INPUT before starting the self-calibration routine. |
| | |

NOTE

If the frequency calibration and amplitude calibration self-calibration routines are both used, the **CAL FREQ** softkey function should always be initiated before the **CAL AMPTD** softkey function.

Key Access: (CAL)

| CAL FETCH | retrieves stored self-calibration correction factors from the last calibration saved using CAL STORE . |
|---------------------|---|
| | Key Access: CAL |
| CAL FREQ | initiates a frequency self-calibration routine. Connect CAL OUT to the analyzer input before pressing CAL FREQ. Remove the cable from the rear panel GATE INPUT before starting the self-calibration routine. |
| | Key Access: (CAL) |
| CAL FREQ & AMPTD | initiates both the frequency and amplitude self-calibration routines. Connect CAL OUT to the analyzer input before pressing CAL FREQ & AMPTD. Remove the cable from the rear panel GATE INPUT before starting the self-calibration routine. |
| | Key Access: CAL |
| CAL STORE | stores the correction factors from the last calibration. The stored correction factors are automatically retrieved when the analyzer is turned on. If correction factors are not stored, they will be retained only until the analyzer is turned off. See the description for the CAL FETCH softkey. |
| | Key Access: CAL |
| CAL TIMEBASE | changes the setting of the 10 MHz reference (standard timebase) DAC that is located on the A25 counter lock assembly. This is a service calibration function and is for service use only. |
| | Key Access: CAL |
| | |

| CAL TRK GEN | <i>Option 011 only.</i> performs the tracking peak self-calibration. The analyzer should be amplitude calibrated by pressing CAL AMPTD prior to using the CAL TRK GEN function. Connect the tracking generator output to the analyzer input before pressing CAL TRK GEN . |
|------------------------|---|
| | Key Access: (CAL) |
| Card Config | accesses the softkey menu that allows you to catalog, format, and delete data from a memory card. |
| | Key Access: (CONFIG) |
| CARD → DISPLAY | recalls into analyzer memory a display image saved on the memory card. It does not recall the associated instrument state, but the restored display can be viewed and copied. Before recalling a display that was saved under a prefix other than the current prefix, change the current prefix to the prefix used when the display was saved. Key Access: (RECALL) |
| $CARD \rightarrow DLP$ | recalls into the analyzer memory a downloadable program (DLP) saved on the memory card. Before recalling data that was saved under a prefix other than the current prefix, change the current prefix to the prefix used when the data was saved. |

Key Access: (RECALL)

| CARD → STATE | recalls into analyzer memory a state saved on the memory card. CARD \rightarrow STATE also displays the time and date when the state data was stored. Before recalling a state that was saved under a prefix other than the current prefix, change the current prefix to the prefix used when the state was saved. |
|--------------------------|--|
| | If the windows display mode is being used, the instrument state can only be recalled into the active window. |
| | Key Access: (RECALL) |
| Card \rightarrow Trace | recalls into analyzer memory a trace saved on the memory card. Limit lines and amplitude correction factors are recalled by pressing Card \rightarrow Trace, LIMIT LINES or AMP COR. |
| | Key Access: (RECALL) |
| CARRIER LVL&FRQ | initiates the Carrier Level and Frequency measurement. |
| | Key Access: CHANNEL MEAS |
| CARRIER/ NOISE | initiates the Carrier to Noise measurement Key Access: CHANNEL MEAS |
| CATALOG ALL | catalogs all the programs and variables stored in analyzer memory in bytes along with the remaining memory available in bytes. |
| | Key Access: (RECALL) or (SAVE) |
| CATALOG AMP COR | catalogs the amplitude correction factor files that are on the memory card. |
| | Key Access: (RECALL) or (SAVE) |

| Catalog Card | accesses a menu with the cataloging functions for the memory card. |
|---------------------|---|
| | Key Access: (RECALL) or (SAVE) |
| CATALOG CARD | allows you to view a listing of the files stored on a RAM card or a ROM card. |
| CATALOG | |
| DISPLAY | catalogs all of the display images that are on the memory card. A display image can be recalled to the analyzer by using the CARD \rightarrow DISPLAY softkey. |
| | Key Access: (RECALL) or (SAVE) |
| CATALOG | |
| DLP | catalogs all of the downloadable programs (DLPs) that are in analyzer memory or on the memory card. Downloadable programs can be saved in analyzer memory by either loading in a downloadable program from the memory card or defining a function using remote programming commands (FUNCDEF or ACTDEF). |
| | Key Access: (RECALL) or (SAVE) |
| Catalog | |
| Internal | accesses a menu that has the cataloging functions for analyzer memory. |
| | Key Access: (RECALL) or (SAVE) |
| CATALOG LMT LINE | catalogs the limit-lines on the memory card. Press CATALOG REGISTER to catalog limit-line tables stored in analyzer memory (limit-line tables saved in analyzer memory are stored in trace registers). Key Access: (RECALL) or (SAVE) |
| | |

| CATALOG ON EVENT | displays the "on event" programming commands and their status. The on event programming commands are as follows: | | |
|---------------------|--|--|--|
| | ONEOS | Performs command list at end of every sweep. | |
| | ONSWP | Performs command list at beginning of every sweep. | |
| | TRMATH | Performs trace math. | |
| | ONCYCLE | Performs command list periodically. | |
| | ONDELAY | Performs command list once after a time period. | |
| | ONMKR | Performs command list at the marker. | |
| | ONSRQ | Performs command list on every service request. | |
| | ONTIME | Performs command list at a specific time. | |
| | ONPWRUP | Performs command list once at power up. | |

NOTE

If you get into an infinite loop with ONPWRUP, press (PRESET) to abort the command list, then press ERASE DLP MEM to clear all DLPs.

Key Access: (RECALL) or (SAVE)

CATALOG PREFIX

catalogs all of the saved data, that has the specified prefix, that is either on the memory card or in analyzer memory.

Key Access: (RECALL) or (SAVE)

| CATALOG REGISTER | displays the status of state and trace registers in analyzer memory. |
|---------------------|---|
| | Key Access: (RECALL) or (SAVE) |
| CATALOG STATES | catalogs all of the states stored on the memory card. Key Access: (RECALL) or (SAVE) |
| CATALOG TRACES | catalogs all of the traces stored on the memory card. |
| CATALOG VARIABLS | catalogs all of the variables saved in analyzer memory. Key Access: (RECALL) or (SAVE) |
| CENTER FREQ | activates the center-frequency function to allow the selection of frequency that will be at the center of the screen. |
| | Key Access: (AUX CTRL), (MEAS/USER), or (FREQUENCY) |
| CF STEP AUTO MAN | changes the step size for the center frequency function. Key Access: (AUTO COUPLE) or (FREQUENCY) |
| Change Prefix | allows you to enter a prefix that can be used for saving and recalling data to and from the memory card, and for cataloging by the prefix. The prefix can be from one to seven characters long. The longer the prefix, the shorter the register number must be. The total length of the prefix and register number cannot exceed eight characters. The prefix can be any character; however, the underscore should not be the first character of the prefix. Key Access: (CONFIG), (DISPLAY), (RECALL), or (SAVE) |
| | |

| Change Title | allows you to write a 53-character screen title across the top of the screen. |
|---------------------|---|
| | Key Access: (CAL) or (DISPLAY) |
| CHANNEL BANDWDTH | allows the user to enter the channel bandwidth to set up the analyzer when using the measurement functions under the Power Menu softkey. |
| | Key Access: (MEAS/USER) |
| CHANNEL POWER | measures the power and power spectral density in the channel bandwidth specified by the user. |
| | Key Access: (MEAS/USER) |
| CHANNEL SELECT | a function used in conjunction with measurements. It sets up the analyzer state for cable TV tests, lets you select a specific channel number, and centers the specific channel on the analyzer screen. |
| | The channel is selected based on one of the five channel formats available in the Setup menu as described in Chapter 2 of the <i>HP 8591C Cable TV Analyzer Cable TV</i> <i>Measurements User's Guide</i> : STD, AIR, HRC, IRC, or T. |
| | Key Access: CHANNEL MEAS or TV RECEIVER |
| CHANNEL SPACING | allows the user to enter the channel spacing to set up the analyzer when using the measurement functions under the Power Menu softkey. Key Access: (MEAS/USER) |



| Clear | |
|----------------|--|
| | clears the current screen title or prefix. This softkey is accessed under the YZ # Spc Clear softkey when using |
| | Change Title or Change Prefix |
| | |
| | Key Access: (CAL), (CONFIG), (DISPLAY), (RECALL), or (SAVE) |
| CLEAR PARAM | clears all of the pulse parameters in the gate utility. It sets the value of pulse width, pulse repetition interval, and reference edge to zero and turns off the coupling of the pulse parameters to the resolution bandwidth, video bandwidth, and sweep time. |
| | Key Access: (Sweep) |
| CLEAR TESTS | allows you to clear all selected tests for the current channel. |
| | Key Access: SYSTEM MEAS |
| CLEAR | |
| TST PLAN | allows you to clear the current test plan. |
| | Key Access: SYSTEM MEAS |
| CLEAR | |
| WRITE A | erases any data previously stored in trace A and continuously displays any signals during the sweep of the analyzer. This function is activated at power on and by pressing (PRESET). |
| | Key Access: (TRACE) |
| CLEAR | |
| WRITE B | erases any data previously stored in trace B and continuously displays any signals detected during the sweep of the analyzer. This function is activated at power on and by pressing (PRESET). |
| | Key Access: (TRACE) |
| | |

| CLEAR | |
|---------------------|---|
| WRITE C | erases any data previously stored in trace C and continuously displays any signals detected during the sweep of the analyzer. This function is activated at power on and by pressing (PRESET). |
| | Kov Accoss: (TRACE) |
| | Rey Access. [INACL] |
| CNT RES AUTO MAN | allows the resolution of the marker counter to be selected manually or auto-coupled. The marker counter has a resolution range of 10 Hz to 100 kHz. |
| | Key Access: (MKR FCTN) |
| CNTL A | |
| 0 1 | makes the auxiliary-interface control line A output high or low (TTL). |
| | Key Access: (AUX CTRL) |
| CNTL B | |
| 0 1 | makes the auxiliary-interface control line B output high or low (TTL). |
| | Key Access: (AUX CTRL) |
| CNTL C | |
| 0 1 | makes the auxiliary-interface control line C output high or low (TTL). |
| | Key Access: (AUX CTRL) |
| CNTL D | |
| 0 1 | makes the auxiliary-interface control line D output high or low (TTL). |
| | Key Access: (AUX CTRL) |

| COARSE TUNE DAC | displays the analog output of the YTO coarse-tune DAC located on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
|--------------------|---|
| | Key Access: CAL |
| CONF TEST | initiates a variety of tests to check the major functions of the analyzer. |
| | Key Access: CAL |
| CONFIG | accesses the softkey menu used for printer and plotter configurations, the time and date display functions, changing the current prefix, memory card configuration functions, disposing of user-defined variables and programs from analyzer memory, changing the analyzer address or the baud rate, displaying the installed options on screen, and changing the format of the MONITOR output. |
| CONT | |
| MEAS | sets the functions in the Power Menu so that they make the measurement at the end of every sweep. |
| | Key Access: (MEAS/USER) |
| CONTINUS FFT | initializes the fast Fourier transform (FFT) function, puts the analyzer in continuous sweep and performs an FFT at the end of each sweep. If the FFT function is already active, it puts the analyzer in continuous sweep and performs FFTs. |
| | Key Access: (MEAS/USER) |

(COPY)

initiates an output of the screen data, without an external controller, to a previously specified graphics printer or plotter.

The printer or plotter must have already been selected using <u>CONFIG</u> and either Plot Config (for a plotter) or Print Config (for a printer). To obtain a print, press <u>CONFIG</u>, COPY DEV PRNT PLT (so that PRNT is underlined), then Print Config, Prn Port Config.

Use **BAUD RATE** to change the baud rate of the analyzer, if necessary. Press (CONFIG), **Print Config**, **Prn Port Config**, **PRN PORT SER PAR** (so that SER is underlined), then press **BAUD RATE**.

For Option 041, use **PRINTER ADDRESS** to change the HP-IB address of the printer, if necessary. Press CONFIG, Print Config, Prn Port Config,

PRN PORT HPIB PAR (so that HPIB is underlined), then press **PRINTER ADDRESS**.

If the analyzer is connected to an HP PaintJet printer and you want a color printout, press **PAINTJET PRINTER**. If the analyzer is connected to an HP PaintJet printer and you want a black and white printout, press **HP B&W PRINTER**. More printer information can be found in the *HP 8590 E-Series and L-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Programmer's Guide*.

If you want the softkey labels to be printed with the analyzer display printout when using (COPY), press **PRT MENU ON OFF** so that ON is underlined.

Press (COPY) and the process will begin. The screen remains frozen (no further sweeps taken) until the data transfer to the printer is complete. The analyzer works with many Hewlett-Packard and Epson compatible printers. The plotting process is similar to the printing process. On the spectrum analyzer, press (CONFIG), Plot Config. To obtain a plot, press (CONFIG), COPY DEV PRNT PLT (so that PLT is underlined), then Plot Config, Plt Port Config.

Use BAUD RATE to change the baud rate of the analyzer, if necessary. Press CONFIG, Plot Config, Plt Port Config, PLT PORT SER PAR (so that SER is underlined), then press BAUD RATE.

For Option 041, use PLOTTER ADDRESS to change the HP-IB address for the plotter, if necessary. Press CONFIG, Plot Config, Plt Port Config, PLT PORT HPIB PAR (so that HPIB is underlined), then press PLOTTER ADDRESS.

With PLTS/PG 1 2 4, you can choose a full-page,

half-page, or quarter-page plot. Press PLTS/PG 1 2 4 to underline the number of plots per page desired. If two or four plots per page are chosen, a softkey function is displayed that allows you to select the location of the plotter output on the paper. If two plots per page are selected, PLT __LOC __ is displayed. If four plots per page are selected, PLT __LOC __ is displayed. Press the softkey until the rectangular marker is in the desired section of the softkey label. The upper and lower sections of the softkey label graphically represent the position of the page where the plotter output will be located.

NOTE

The HP 7470A Plotter does not support two plots per page output. If you use an HP 7470A Plotter with the HP 8591C cable TV analyzer, you can select one or four plots per page, but not two plots per page.

For a multipen plotter, the pens of the plotter draw the different components of the screen as follows:

| Pen Number | Description |
|---------------|---|
| 1 | Draws the annotation and graticule. |
| 2 | Draws trace A. |
| 3 | Draws trace B. |
| 4 | Draws trace C and the display line. |
| 5 | Draws user-generated graphics and the lower limit line. |
| 6 | Draws the upper limit line. |

To plot, press **Previous Menu**, **COPY DEV PRNT PLT** (PLT should be underlined), and (COPY).

Printing is usually faster than plotting, but plotting provides higher resolution output. The analyzer works with plotters such as the HP 7440A.

COPY DEV PRNT PLT changes between a printer and plotter. For example, if you have been printing and want to do a plot, press COPY DEV PRNT PLT to underline PLT before pressing (COPY). Key Access: (CONFIG)

CORRECT ON OFF

controls use of some of the correction factors. When ON is underlined, correction factors are used and CORR appears on the display. When OFF is underlined, correction factors are not used. Turning the correction factors off degrades amplitude accuracy.

NOTE

Correction factors must be on for the analyzer to meet its specified performance.

Key Access: (CAL)

CPL RBW

ON OFF

automatically selects the optimum resolution bandwidth for an unmodulated pulse if the pulse width has been entered. If the pulse width has not been determined, the resolution bandwidth will not be coupled to the pulse parameters and a warning message will occur. If a resolution bandwidth is entered manually, the coupling will be turned off.

Key Access: (SWEEP)

| CPL SWP ON OFF | automatically selects the optimum sweep time if the pulse repetition interval has been entered. If the pulse repetition interval has not been determined, the sweep time will not be coupled and an error message will occur. If a sweep time is entered manually, the coupling will be turned off. Sweep time coupling only applies to the frequency domain window. |
|----------------------|---|
| | Key Access: (SWEEP) |
| CPL VBW ON OFF | automatically selects the optimum video bandwidth if the gate length has been entered. If the gate length has not been determined, the video bandwidth will not be coupled to the gate length and an error message will occur. If a gate length is entered manually, the coupling will be turned off. |
| | Key Access: (SWEEP) |
| Create TstPlan | allows you to access the functions used for building a test plan. |
| | Key Access: SYSTEM MEAS |
| CROSSMOD | initiates the Cross Modulation measurement. Key Access: CHANNEL MEAS |
| CRT HORZ POSITION | changes the horizontal position of the signal on the analyzer display. Press CAL STORE if you want the analyzer to use this position permanently, so that it is not lost when the power is turned off. |
| | Key Access: [CAL] |

| CRT VERT POSITION | changes the vertical position of the signal on the analyzer display. Press CAL STORE if you want the analyzer to use this position permanently, so that it is not lost when the power is turned off. |
|----------------------|---|
| | Key Access: (CAL) |
| CSO/CTB | initiates the Composite Second Order/Composite Triple Beat measurement. |
| | Key Access: CHANNEL MEAS |
| DACS | changes the DAC numbers of the span, DAC YTO coarse-tune, DAC YTO fine-tune, and YTO FM tune DAC located on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: CAL |
| DATA TO PRINTER | allows you to retrieve the test data from a RAM card file and send it to a printer to be printed. (The printer needs to be connected to the analyzer.) |
| DATEMODE MDY DMY | changes the display of the date from a month-day-year format to a day-month-year format. It is set to a month-day-year format by pressing DEFAULT CONFIG . |
| | Key Access: (CONFIG) |

| dBm | |
|---------------------|---|
| CDM | changes the amplitude units to dBm for the current setting (log or linear). |
| | Key Access: (AMPLITUDE) |
| dBmV | changes the amplitude units to dBmV for the current setting (log or linear). |
| | Key Access: (AMPLITUDE) |
| dBuV | changes the amplitude units to $dB\mu V$ for the current setting (log or linear). |
| | Key Access: (AMPLITUDE) |
| DEFAULT CAL DATA | accesses the factory-default correction factors. A special pass code is required for use. If the message Self cal needed appears when DEFAULT CAL DATA is pressed, the CAL FREQ and CAL AMPTD routines need to be run. The CAL FREQ and CAL AMPTD must be run to ensure specifications. The calibration results must then be saved by pressing CAL STORE . See Chapter 8 for more information. |
| | Key Access: (CAL) |
| DEFAULT CONFIG | resets the analyzer configuration to the state it was in when it was originally shipped from the factory and performs an instrument preset. See Table 4-2 for the default user-configuration values set by pressing DEFAULT CONFIG . |

| Configuration | Default Value |
|---------------------------------------|-------------------------|
| Analyzer address Option 021 | 18 |
| Copy device | p rin ter |
| CRT position Horizontal and Vertical | 10, 48 |
| Printer address | 1 |
| Plotter address | 5 |
| Baud rate | 1200 |
| External preamp | 0 dB |
| Save lock internal states or traces | Off |
| Prin ter | black and white printer |
| Print menu | on |
| Plots per page | 1 |
| Time/date display | on |
| Date mode | month-day-year format |
| Prefix | blank |
| Analyzer state at power on | instrument preset |

Table 4.2. Default Configuration Values

Key Access: (CONFIG)

DEFAULT

SYNCrestores the factory default values of the horizontal and
vertical synchronization constants for the rear panel
MONITOR output. CRT SYNC DEFAULT can be used
to exit from the NTSC or PAL modes to return to the
normal monitor output and use the default synchronization
constants.Define
CouplingOption 107 only.
accesses the time gate utility menu for coupling analyzer
settings to the input pulsed signal parameters. It also
accesses the pulsed signal parameter entry menus.

Key Access: (SWEEP)

| Define Gate | <i>Option 107 only.</i> accesses the menu for turning on and defining the gate, from within the time gate utility. |
|----------------|---|
| | Key Access: (SWEEP) |
| Define Time | <i>Option 107 only.</i> accesses the menu for manipulating the time domain window in the gate utility. It will automatically make the time window active and turn off the gate. The trigger marker can be activated from this menu. |
| | Key Access: (SWEEP) |
| DELETE FILE | function allows you to delete an item from analyzer memory or a file from the memory card. |

NOTE

Deleting items beginning with an underscore from analyzer memory is not recommended and may have unexpected results. Items beginning with an underscore are used by the analyzer.

Key Access: (RECALL) or (SAVE)

DELETE POINT deletes an amplitude-correction factor that was previously selected by SELECT POINT.

Key Access: (CAL)

 DELETE

 SEGMENT

 deletes the limit-line entry for the selected segment number.

 Press
 SELECT

 Segment
 number to select the limit-line entry for deletion.

Key Access: (DISPLAY)

| Demod | accesses the softkeys controlling demodulation functions, |
|---------------------|--|
| | speaker volume, squelch level, FM gain, and dwell time. |
| | Key Access: (AUX CTRL) |
| DEPTH | |
| MOD | initiates the Depth of Video Modulation measurement. |
| | Key Access: CHANNEL MEAS |
| DEMOD | |
| AM FM | allows selection of amplitude (AM) or frequency (FM) demodulation. |
| | Key Access: (AUX CTRL) |
| DEMOD | |
| ON OFF | turns the AM or FM demodulation on and off. |
| | Key Access: (AUX CTRL) |
| DESKJET 310/550C | selects the HP DeskJet 310 or 550C for color printing. Use this function if you have one of these color printers. This softkey can be accessed by pressing <u>CONFIG</u> , then Set Colr Printer . |
| | Key Access: (CONFIG) |
| DESKJET 540 | selects the HP DeskJet 540 for color printing. Use this function if you have this color printer. This softkey can be accessed by pressing (<u>CONFIG</u>), then Set Colr Printer . |
| | rront-Panel Key Access [CONFIG] |

DETECTOR selects between positive peak, sample, and negative PK SP NG peak detection. When sample detection is selected, SMPL appears in the upper-left corner of the screen. In sample mode, the instantaneous signal value at the present display point is placed in memory. Sample detection is activated automatically for noise level markers, during video averaging, and for FFT measurements. Key Access: (TRACE) DIF GAIN initiates the differential gain/differential phase measurement. DIF PHAZ Key Access: CHANNEL MEAS (DISPLAY) accesses softkeys that include the HOLD key and limit line menus, and activate the display line, threshold and analog+ display mode. They allow title and prefix entry, and control the display of the graticule and screen annotation. DISPLAY displays the current correction-factor data generated by the CAL DATA frequency and amplitude self-calibration routines. This is a service diagnostic function and is for service use only. Key Access: (CAL) DISPLAY saves the current analyzer display image on the memory \rightarrow CARD card for viewing or copying later. It does not save the analyzer state or trace, so it cannot be used to restore an instrument to the conditions at the time the image was saved. Key Access: (SAVE)
DISPLAY
CNTL I displays the status of the auxiliary connector input (control
line I) on the analyzer screen (high = 1 or low = 0, in TTL).
Key Access: (AUX CTRL) Dispose
User Mem accesses the softkeys ERASE MEM CARD, ERASE DLP MEM,
ERASE STATEALL, ERASE TRACEALL, and
ERASE MEM ALL, which are used to erase the user
programs and variables that are in analyzer memory. NOTE

Use DELETE FILE to selectively delete stored programs or variables from analyzer memory.

Key Access: (CONFIG)

| DL DELTA | |
|----------|---|
| ON OFF | displays the change, in percent, between a reference display line and the display line activated by pressing DL DELTA ON OFF. The percentage is displayed in the active function block. |
| | Key Access: (DISPLAY) |
| DO FRQ | |
| RESP TST | compares a spectrum stored at a reference point to a spectrum at any other point in the system where you want to evaluate the frequency response. The signal level and amplitude slope should be approximately the same as the reference trace. |
| | Key Access: CHANNEL MEAS |

| DONE | |
|---------------------|--|
| DONE | allows you to indicate that you have finished making your selection. |
| | (The DONE softkey should not be confused with the DONE programming command.) |
| | Key Access: SYSTEM MEAS |
| DROOP | disables the reset of the peak detector on the A16 processor/video assembly after each analog-to-digital conversion. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| DSP LINE ON OFF | activates an adjustable horizontal line that is used as a visual reference line. |
| | Key Access: (DISPLAY), (MKR FCTN), or (PEAK SEARCH) |
| DUMP TO PRINTER | allows you to have the test results sent to a printer when using Measure Now or Timed Measure. The printer needs to be connected to the analyzer. |
| | Key Access: SYSTEM MEAS |
| DUMP TO RAM CARD | allows you to store the test results on a RAM card when using Measure Now or Timed Measure . The RAM card must be formatted and ready for use to store the test data. Up to 40 files of test data can be stored on a 128 kilobyte or 32 kilobyte RAM card. No more than 40 files of test data can be stored on a RAM card, however. |
| | |

Key Access: SYSTEM MEAS

| DUMP TO | allows you to send test results to the analyzer screen. Upon | | |
|----------|--|--|--|
| SCREEN | being sent to the screen, the following options are available: | | |
| | • NEXT PAGE displays the next page of test results. | | |
| | • PREV PAGE displays the previous page of test results. | | |
| | • GOTO PAGE displays selected page number. | | |
| | • PRINT REPORT prints a test results report. | | |
| | • STORE TO CARD stores test results to a RAM card. | | |
| | Key Access: SYSTEM MEAS | | |
| DWELL | | | |
| TIME | sets the dwell time for the marker pause, during which demodulation can take place in nonzero span sweeps. The dwell time can be set from 2 milliseconds to 100 seconds. | | |
| | Key Access: (AUX CTRL) | | |
| EDGE POL | | | |
| POS NEG | determines whether the gate triggers on the positive-going or negative-going edge of the signal at the GATE TRIGGER INPUT connector (on the rear panel of the analyzer). | | |
| | Key Access: (SWEEP) | | |
| Edit | | | |
| Amp Cor | allows you to edit the current amplitude-correction factors table by accessing the SELECT POINT, SELECT FREQ, | | |
| | SELECT AMPLITUD, DELETE POINT, Edit Done, and | | |
| | PURGE AMP COR softkeys. | | |
| | Key Access: CAL | | |

| EDIT CAT ITEM | starts the DLP editor function and loads the highlighted item from the catalog of analyzer user memory to be displayed and edited. Refer to the <i>HP 8590 E-Series and L-Series</i> <i>Spectrum Analyzers and HP 8591C Cable TV Analyzer</i> <i>Programmer's Guide</i> for more information. |
|------------------|---|
| | Key Access: (RECALL) or (SAVE) |
| Edit Done | can be accessed through both the amplitude-correction menu and the change prefix menu. When accessed from the amplitude-correction menu, the amplitude-correction factors table is erased from the analyzer screen and the amplitude-correction menu is restored on-screen. Use Edit Done when all the amplitude-correction factors have been entered. |
| | Key Access: (Cal), CONFIG, (DISPLAY), (RECALL), or (SAVE) |
| EDIT DONE | can be accessed through the limit-line menu. Pressing EDIT DONE erases the limit-line table from the analyzer screen and restores the menu accessed by the Limit Lines softkey. Use EDIT DONE when all the limit-line values have been entered. |
| | Key Access: (DISPLAY) |
| EDIT FLATNESS | allows flatness-correction constants to be viewed or modified. This is a service calibration function and is for service use only. |
| | Key Access: (CAL) |

| EDIT LAST | starts the DLP editor function and allows the most recent item that was being edited, in the DLP editor buffer, to be accessed again. The item will not be in the analyzer user memory until it is processed by the SAVE EDIT softkey. The DLP editor memory remains intact when the instrument is preset and when it is powered off. Refer to the <i>HP 8590 E-Series and L-Series Spectrum Analyzers and</i> <i>HP 8591C Cable TV Analyzer Programmer's Guide</i> for more information. |
|------------------|--|
| | Key Access: (RECALL) or (SAVE) |
| Edit Limit | allows you to edit the current limit-line tables by accessing Edit Upper, Edit Lower, Edit Up/Low, and Edit Mid/Delt. Use PURGE LIMITS under any of the above edit menus to dispose of the current limit-line table. |
| | Key Access: (DISPLAY) |
| Edit Lower | allows you to view or edit the lower limit-line table. |
| | Key Access: (DISPLAY) |
| Edit Mid/Delt | allows you to view or edit the upper and lower limit-line tables simultaneously. |
| | Key Access: (DISPLAY) |
| Edit Up/Low | allows you to view or edit the upper and lower limit-line tables simultaneously. |
| | Key Access: (DISPLAY) |
| Edit Upper | allows you to view or edit the upper limit-line table. Key Access: (DISPLAY) |
| | |

| EDIT | | |
|---------------------|--|--|
| UPR LWR | selects upper or lower limit-line tables. It switches to the limit-line table that is not currently being edited. | |
| | Key Access: (<u>DISPLAY</u>) | |
| Editor | accesses the menu of down loadable program (DLP) editor softkeys. Refer to the <i>HP 8590 E-Series and L-Series</i> Spectrum Analyzers and <i>HP 8591C Cable TV Analyzer</i> Programmer's Guide for more information. | |
| | Key Access: (RECALL) or (SAVE) | |
| EMI BW Menu | accesses the functions that set the analyzer resolution bandwidth to the values required for electromagnetic interference (EMI) testing. (200 Hz EMI BW <i>is only</i> <i>available with Option 130</i> .) | |
| | Key Access: (BW) | |
| ((ENTER)) | Pressing $(Hz/\mu V/\mu s)$, which is also the ENTER key, terminates and enters into the analyzer a numerical value that has been entered from the front panel using the keypad, knob, or step up/down keys. | |
| ENTER BY CHANNEL | allows you to enter channels individually or to edit the test plan. A maximum of 158 channels (channel numbers 0, 2 through 158) can be entered into a test plan. | |
| | Key Access: SYSTEM MEAS | |
| ENTER BY RANGE | allows you to enter channels as a range of channel numbers. Unlike ENTER BY CHANNEL, you cannot use ENTER BY RANGE to edit or delete a channel. A maximum of 158 channels (channel numbers 0, 2 through 158) can be entered into a test plan. Key Access: SYSTEM MEAS | |
| | | |

ENTER LOCATION allows you to enter a location code. You do not have to enter a location code; it is only for your convenience to help you keep track of where the test data originated. The location code can be up to seven digits long.

Key Access: SYSTEM MEAS

ENTER TEMP

allows you to enter the temperature. You can enter a number from -999 to +999. The temperature is entered without Celsius or Fahrenheit units. The analyzer does not measure the temperature. You must provide the temperature reading.

You do not have to enter a temperature; it is only for your convenience. If you do not enter a temperature, the default temperature entry will either be the temperature that was previously entered or, if a temperature reading was not previously entered, a "0."

On the RAM card, the temperature entry is stored in the test data file. The temperature entry is printed when the test data is printed.

In analyzer memory, the temperature entry is not stored with the current test plan in plans 1 through 5. The temperature entry is stored in nonvolatile analyzer memory, however. Storing the temperature entry in nonvolatile memory means that once a temperature is entered, it will be used as the temperature until a new temperature is entered or the HP 85721A program is deleted from analyzer memory.

Key Access: SYSTEM MEAS

| ENTER PRI | accesses the menu for entering the value of the pulse repetition interval (PRI). |
|-------------------|--|
| | Key Access: (SWEEP) |
| ENTER REF EDGE | accesses the menu used to define the edge of the pulse that will be used as the time reference. |
| | Key Access: (SWEEP) |
| ENTER WIDTH | accesses the menu for entering the value of the pulse width. Press MARKER ON to activate a delta marker. |
| | Key Access: (SWEEP) |
| ERASE DLP MEM | allows you to dispose of the DLPs, all traces defined by TRDEF, and all VAREF variables that are in analyzer memory. |
| | |

NOTE

Use **DELETE FILE** to selectively delete stored programs or variables from analyzer memory.

Key Access: (CONFIG)

ERASE MEM ALL

allows you to purge all user state registers, all user trace registers, all mode registers, the editor buffer, Group delay normalization, all DLP memory, and all microprocessor stack data.

NOTE

ERASE MEM ALL ignores the state of the **SAV LOCK**. So, even if STATES and TRACES are locked, they will still be erased by **ERASE MEM ALL**.

Key Access: (CONFIG)

ERASE

MEM CARD

allows you to dispose of any data or programs stored on the memory card by formatting it.

NOTE

Use DELETE FILE to selectively delete stored programs or variables from analyzer memory.

Key Access: (CONFIG)

ERASESTATEALLallows you to purge all the user state registers 1 through 9.

NOTE

- If SAVE LOCK ON OFF is set to (ON), this function is disabled.
- Use **DELETE FILE** to selectively delete stored programs or variables from analyzer memory.

| Kev | Access: | |
|-------|---------|----------|
| ITC A | ALLESS. | (CONFIG) |

ERASE TRACEALL

allows you to purge all the user trace registers 0 through TRCMEM.

NOTE

- If **SAVE LOCK ON OFF** is set to (ON), this function is disabled.
- Use DELETE FILE to selectively delete stored programs or variables from analyzer memory.

Key Access: (CONFIG)

EP LQ570 SML LRG sets the analyzer to be compatible with an Epson LQ-570 compatible printer.

Key Access: (CONFIG)

| EP MX80 SML LRG | sets the analyzer to be compatible with an Epson MX80 compatible printer. |
|--------------------|--|
| | Key Access: (CONFIG) |
| EXECUTE TITLE | executes a programming command displayed in the screen title area of the analyzer. |
| | Key Access: CAL |
| EXIT | exits the EDIT FLATNESS softkey menu. This is a service calibration function and is designed for service use only. |
| | Key Access: (CAL) |
| Exit Catalog | returns the analyzer to the state it was in before the current catalog function was invoked. It clears the catalog display and returns to a normal analyzer display. |
| | Key Access: (RECALL) or (SAVE) |
| EXIT SHOW | removes the screen annotation left after pressing SHOW OPTIONS . |
| | Key Access: CONFIG |
| EXIT UTILITY | exits the gate utility returning to the gate control menu. The analyzer is returned to normal operation. The state of the analyzer becomes the state of the window that was active when exiting the gate utility. |
| | Key Access: (SWEEP) |

| EXT AMP GAIN | allows you to select the gain value of an external preamplifier for the carrier-to-noise test. The default value is 20 dB for the HP 85905A preamplifier. | |
|-------------------|---|--|
| | The preamplifier selection is stored in nonvolatile analyzer memory. This means that the analyzer retains the selection, even when power is turned off, until a new selection is made. | |
| | Press CABLE TV ANALYZER, Setup, Analyzer Input, EXT AMP YES NO to access EXT AMP GAIN. | |
| | Key Access: CABLE TV ANALYZER | |
| EXT AMP NZ FIG | allows you to select the noise figure value of an external preamplifier for the carrier-to-noise test. The default value is 7 dB for the HP 85905A preamplifier. | |
| | The preamplifier selection is stored in nonvolatile analyzer memory. This means that the analyzer retains the selection, even when power is turned off, until a new selection is made. | |
| | Press CABLE TV ANALYZER, Setup, Analyzer Input to access EXT AMP NZ FIG. | |
| | Key Access: CABLE TV ANALYZER | |
| EXT AMP YES NO | allows you to select an external preamplifier for the carrier-to-noise test. The preamplifier selection is stored in nonvolatile analyzer memory. This means that the analyzer retains the selection, even when power is turned off, until a new selection is made. | |
| | Press CABLE TV ANALYZER, Setup, Analyzer Input to access EXT AMP YES NO. | |
| | Key Access: CABLE TV ANALYZER | |

EXT PAD YES NO

can be used to compensate for the amplitude loss caused by using a 50 Ω to 75 Ω external matching pad. (A 50 Ω to 75 Ω external matching pad is used to match the signal from a 75 Ω impedance system to a spectrum analyzer with a 50 Ω impedance input.) The external pad function compensates for the 5.8 dB amplitude loss of the external pad by setting the spectrum analyzer's reference level offset to 5.8 dB.

If you are using a spectrum analyzer with 50 Ω input impedance but are not using an external matching pad, the cable TV analyzer mode will compensate for most of the impedance mismatch, but amplitude measurements can have up to ± 0.2 dB error caused by the uncompensated mismatch over the frequency range.

This function is set to NO when the HP 85721A is loaded into analyzer memory.

In spectrum analyzer memory, the setting of the external pad function entry is not stored with the current test plan in plans 1 through 5. The setting of the external pad function is stored in nonvolatile spectrum analyzer memory, however. Storing the setting in nonvolatile memory means that once the external pad function is set to YES, it will remain set to YES until you change the setting or delete the HP 85721A program from analyzer memory.

Press CABLE TV ANALYZER, Setup, Analyzer Input to access EXT PAD YES NO.

Key Access: CABLE TV ANALYZER

EXTERNAL

activates the trigger condition that allows the next sweep to start when an external voltage (connected to EXT TRIG INPUT on the rear panel) passes through approximately 1.5 V. The external trigger signal must be a 0 V to +5 V TTL signal.

Key Access: (TRIG)

| EXTERNAL | |
|----------|---|
| PREAMPG | adds a positive or negative preamplifier gain value, which is subtracted from the displayed signal. |
| | Key Access: (AMPLITUDE) |
| FFT | |
| MARKERS | if the FFT mode is already active, pressing FFT MARKERS |
| | only activates the FFT markers. If the FFT mode is not |
| | active, then the FFT marker will be activated and the |
| | FFT annotation will be displayed but an FFT will not be |

performed.

FFT Menu Key Access: (MEAS/USER)

accesses the menu of keys to initiate and evaluate a Fourier transform of the analyzer displayed data. Activating other analyzer functions will automatically exit the FFT function or will corrupt the display. See Table 4-3.

Table 4.3. Compatibility of FFT with Other Functions

| Function | Compatibility Info |
|-------------------------|----------------------|
| Catalog functions | corrupts the display |
| DLP editor | corrupts the display |
| Marker table | exits FFT |
| N dB points measurement | exits FFT |
| Peak table | exits FFT |
| Power menu measurements | exits FFT |
| Show options | corrupts the display |
| Time gate functions | exits FFT |
| TOI measurement | exits FFT |
| Windows | exits FFT |

Key Access: (MEAS/USER)

FFT OFF

exits the FFT mode and menus, returning the analyzer to normal operation.

Key Access: (MEAS/USER)

FFT STOP FREQ

allows the user to enter the stop frequency for the desired FFT span. This sets the analyzer sweep time by the relationship:

FFT stop freq = $400 / (sweep time \times 2)$

NOTE

When using the FFT stop frequency function, the knob and step key increments are not optimal for positioning the signal. To use the knob or step keys to position the signal, select sweep time as the active function.

| | Key Access: (MEAS/USER) |
|------------------|--|
| FINE TUNE DAC | displays the output of the YTO fine-tune DAC, which is produced on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: CAL |
| FLAT | draws a zero-slope line between the coordinate point of the current segment and the coordinate point of the next segment, producing limit-line values equal in amplitude for all frequencies between the two points. If the amplitude values of the two segments differ, the limit line "steps" to the frequency value of the second segment. |
| | Key Access: (DISPLAY) |
| Flatness Data | provides access to the softkeys used for viewing or editing the flatness-correction constants. This is a service calibration function and is for service use only. |
| | Key Access: (CAL) |

| FM COIL DRIVE | displays the output of the FM coil driver produced on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
|------------------|---|
| | Key Access: (CAL) |
| FM DEV | initiates the FM Deviation measurement. |
| | Key Access: CHANNEL MEAS |
| FM GAIN | adjusts the FM deviation display. The center graticule represents zero deviation. The top graticule is the positive deviation set by FM GAIN. The bottom graticule is the negative deviation set by FM GAIN. The range for FM gain is from 10 kHz to 500 kHz. The default value is 100 kHz. |
| | Key Access: (AUX CTRL) or (CAL) |
| FM OFFSET | adjusts the horizontal trace for center-screen with no modulation on the carrier. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| FM SPAN | displays the FM_SPAN signal from the span dividers on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| FORMAT CARD | formats a card in logical interchange format (LIF). This is the same as $\tt ERASE \ MEM \ CARD$. |
| | Key Access: (CONFIG) |

| FREE BUN | |
|-----------------------|--|
| | activates the trigger condition that allows the next sweep to start as soon as possible after the last sweep. |
| | Key Access: (Trig) |
| FREQ DIAG | displays, in real-time, frequency diagnostic information for the LO section. This is a service diagnostic function and is for service use only. |
| | Key Access: CAL |
| FREQ DISC NORM OFF | indicates the status of the frequency discriminator as a function of LO span. This is a service diagnostic function and is for service use only. |
| | Key Access: CAL |
| FREQ OFFSET | allows the user to input a frequency offset value that is added to the frequency readout, to account for frequency conversions external to the analyzer. |
| | Key Access: (FREQUENCY) |
| (FREQUENCY) | activates the center-frequency or start-frequency function and accesses the menu that has the frequency functions. The center frequency or start frequency value appears below the graticule on the screen. |

Although the analyzer allows entry of frequencies greater than the specified frequency range, using frequencies greater than the frequency span of the analyzer is not recommended.

Gate Option 107 only. Control accesses the menu of gate control functions and the entrance to the gate utility menus. The gate can be controlled independently or from within the gate utility. The gate utility makes it easier to set up and manipulate the gate. Key Access: (SWEEP) Option 107 only. GATE CTL determines if the gate is enabled on the edge of the trigger EDGE LVL input or on a threshold level of the input signal. If the gate control is set to EDGE, the edge of the input signal triggers the timer for the gate delay. When the gate control is set to LVL, the gate follows the positive level of the signal connected to GATE TRIGGER INPUT. Key Access: (SWEEP) Option 107 only. GATE sets the duration of the delay after an edge trigger before DELAY the gate switch closes. The gate delay can be set from 1 μ s to 65.535 ms in 1 μ s steps. Key Access: (SWEEP) Option 107 only. GATE sets the duration of the gate. The gate length can be set LENGTH from 1 μ s to 65.535 ms in 1 μ s steps. Key Access: (SWEEP) Option 107 only. GATE turns on or off the gate, the time-gated analyzer capability. ON OFF The gate can be turned on outside, or from within, the gate utility. The gate utility makes it easier to set up and manipulate the gate. The gating function allows you to make carrier to noise, in-channel frequency response, and composite second order (CSO) distortion measurements without turning off modulation (Option 107 required). Key Access: (SWEEP)

GATE UTILITY

Option 107 only.

accesses the softkey functions used for the time-gated analyzer capability. It creates related time domain and frequency domain windows to set up the time gate and make measurements.

Note that if the gate utility menus are exited by pressing one of the front panel keys they may be re-entered by pressing (SWEEP) twice.

Some analyzer functions are altered or are not available when the gate utility is active. See Table 4-4. From within the gate utility press **EXIT UTILITY** to return to normal analyzer operation. Press **SWEEP** twice, and **EXIT UTILITY** to return to the gate utility and then exit it properly.

Table 4.4. Commands Altered/Not Available within the Gate Utility

| Command | Description of Change |
|------------------------|---|
| % AM | measurement function is not available |
| Adjacent channel power | measurement function is not available |
| Calibration | no calibration functions can be accessed |
| Channel power | measurement function is not available |
| FFT | measurement function is not available |
| N dB points | measurement function is not available |
| Occupied bandwidth | measurement function is not available |
| Peak zoom | routine is not available |
| Res BW | turns off resolution BW coupling to pulse width |
| Span | must be zero in the time domain window |
| Sweep time | turns off sweep time coupling to PRI |
| TOI | measurement function is not available |
| Video BW | turns off video BW coupling to gate length |

Key Access: (SWEEP)

| GHIJKL | accesses the softkey menu used for selecting screen title or prefix characters G through L. |
|----------|---|
| | Key Access: (CAL), (CONFIG), (DISPLAY), (RECALL), or (SAVE) |
| GND REF | |
| DETECTOR | displays the output of the analog-ground reference produced on the A16 processor/video assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| GOTO | |
| PAGE | allows you to select a specific page of on-screen test results to view. |
| GRAT | |
| ON OFF | turns the screen graticule on and off. This is helpful when alternative graphics are drawn on the screen through a remote controller and during plotting, when a graticule is not required. |
| | Key Access: (DISPLAY) |
| GRPH MKR | |
| ON OFF | turns the graph marker ON or OFF. This softkey is available when ACPGRAPH is ON (for ACP or ACP extended measurements), or when PWRGRAPH is ON (for channel power measurements). |
| | Front-Panel Access Key: (MEAS/USER) |
| HOLD | |
| | deactivates the active function and blanks the active function text from the display. No data can be accidentally entered using the knob, step keys, or keypad. Activating another function will turn off the hold function. |
| | Key Access: (DISPLAY) |

| HP B&W DJ540 | selects a black and white print. Use this function if you have a black and white HP DeskJet 540 printer. Press HP B&W Printer and then select the HP B&W PRINTER softkey. Start printing by pressing COPY DEV PRNT PLT (PRNT) and (COPY). |
|-------------------|--|
| | Front-Panel Key Access (<u>CONFIG</u>) |
| HP B&W PRINTER | selects a black and white print. Use this function if you have a black and white HP printer, or if you are using an HP PaintJet printer, but want to have a black and white print. |
| | Key Access: CONFIG |
| HRC | allows you to select the harmonically related carriers (HRC) tune configuration. HRC is the channel tuning configuration in which the frequency of each video carrier is a multiple of 6 MHz. The HRC frequency assignments are defined in the Federal Communications Commission's (FCC) channel identification plan, part 76.612. |
| | If you do not select a channel tuning configuration, the default configuration is the standard (STD) configuration. The choices for the channel tuning configuration are: standard (STD), off-the-air (AIR), incrementally related carriers (IRC), harmonically related carriers (HRC), and T format. |
| | The channel tuning configuration selection is selected independently of the test plan and is not stored with the test plan. The analyzer retains the channel tuning configuration selection, even when power is turned off, until a new selection is made. |
| | Press CABLE TV ANALYZER, Setup, Channel Tuning to access HRC. |
| | Key Access: CABLE TV ANALYZER |

| ним | |
|--------------------|---|
| | initiates the Hum measurement. |
| | Key Access: CHANNEL MEAS |
| IDNUM | is used when the instrument is powered on the first time. It inputs the analyzer model number and option information. This is a service calibration function and is for service use only. Refer to the service guide for more information. |
| | Key Access: (CAL) |
| INIT FLT | is used when the instrument is powered on the first time. It sets the defaults for spectrum analyzer flatness including the start and stop frequencies and the step size. All of the correction values are set to zero. This is a service calibration function and is for service use only. Refer to the service guide for more information. |
| | Key Access: (CAL) |
| INPUT Ζ 50Ω 75Ω | adjusts the voltage readout by 1.76 dB to correct for the difference between voltage and power measurements in a 75 Ω system versus a 50 Ω system. The impedance you select is for computational purposes only, since the actual impedance of 75 Ω is set by internal hardware. The preset value can be changed by using a service function. Select the computational input impedance by pressing INPUT Z 500 750 or by entering 75 or 50 using the numeric keypad. |
| | Key Access: (AMPLITUDE) |
| INTERNAL CARD | selects between spectrum analyzer memory and the memory card for the save and recall functions. Key Access: (RECALL) or (SAVE) |

$\begin{array}{r} \text{INTERNAL} \\ \rightarrow & \text{STATE} \end{array}$

recalls the saved analyzer state from the selected state register. Recalling a state from the analyzer memory displays the time and date when the state data was stored.

Key Access: (RECALL)

Internal

 \rightarrow Trace

accesses a softkey menu that allows you to either select the trace in which the trace data is to be recalled (trace A, trace B, or trace C), recall the current limit-line tables, or recall amplitude correction factors.

Key Access: (RECALL)

IRC

allows you to select the incrementally related carriers (IRC) channel tuning configuration. IRC is the configuration in which all channels except channels 5 and 6 are at the Federal Communication Commission's (FCC) standard channel frequencies. The IRC frequency assignments are defined in the FCC channel identification plan, part 76.612.

If you do not select a channel tuning configuration, the default configuration is the standard (STD) configuration. The choices for the channel tuning configuration are: standard (STD), off-the-air (AIR), incrementally related carriers (IRC), harmonically related carriers (HRC), and T format.

The channel tuning configuration selection is selected independently of the test plan and is not stored with the test plan. The analyzer retains the channel tuning configuration selection, even when power is turned off, until a new selection is made.

Press CABLE TV ANALYZER, Setup, Channel Tuning to access IRC.

Key Access: CABLE TV ANALYZER

| LAST SPAN | changes the analyzer frequency span to the previous span setting. |
|-------------------|---|
| | Key Access: (SPAN) |
| Limit Lines | accesses the limit-line menus. |
| | Key Access: (DISPLAY) |
| LIMIT LINES | When accessed by SAVE, pressing LIMIT LINES stores the current limit-line tables in analyzer memory or on the memory card. When accessed by RECALL, pressing LIMIT LINES recalls limit-line tables from analyzer memory or the memory card. See "To Save a Limit-Line Table or Amplitude Correction Factors" or "To Recall Limit-Line Tables or Amplitude Correction Factors" in Chapter 3 of the HP 8591C Cable TV Analyzer Spectrum Analyzer Reference User's Guide for more information. |
| | Key Access: (RECALL) or (SAVE) |
| LIMITS FIX REL | allows you to choose fixed or relative type of limit lines. The fixed (FIX) type uses the current limit line as a reference with fixed frequency and amplitude values. The relative (REL) setting causes the current limit-line value to be relative to the displayed center frequency and reference-level amplitude values. When limit lines are specified with time, rather than frequency, the REL setting only affects the amplitude values. The current amplitude values will be relative to the displayed reference-level amplitude, but the time values will always start at the left edge of the graticule. |

| LIMITS FRQ TIME | selects whether limit lines will be entered using frequency or sweep time to define the segments. Limit lines can be created by the user to test trace data. They can be specified as a table of limit-line segments of amplitude versus frequency, or of amplitude versus time. |
|----------------------|---|
| | Key Access: (DISPLAY) |
| LINE | activates the trigger condition that allows the next sweep to be synchronized with the next cycle of the line voltage. |
| | Key Access: (TRIG) |
| LISTEN ON OFF | <i>Option 102 only.</i> demodulates the aural carrier so that it is audible through the built-in speaker. |
| | Key Access: CHANNEL MEAS |
| LMT DISP Y N AUTO | displays any portion of the limit lines that are currently within the analyzer display boundary. If Y (yes) is underlined the limit lines are displayed. If N (no) is underlined they are not displayed. If AUTO is underlined, the display of the limit lines is dependent on LMT TEST. The limit lines will be displayed while the limit test function is turned on, otherwise they will be turned off. |
| | Limit lines cannot be displayed while using the analog+ display mode. Limit testing can be done but the limit lines will not be displayed. |
| | Key Access: (DISPLAY) |

| LMT TEST ON OFF | turns the limit-line testing and (if LMT DISP AUTO is selected) turns the display of the limit lines on and off. When limit-line testing is enabled, every measurement sweep of trace A is compared to the limit lines. If trace A is at or within the bounds of the limit lines, LIMIT PASS is displayed. If trace A is out of the limit-line boundaries, LIMIT FAIL is displayed. Key Access: (DISPLAY) |
|--------------------|--|
| LOAD FILE | loads a file from the memory card into analyzer memory. |

NOTE

Use of the LOAD FILE softkey is not recommended for recalling limit-line tables or amplitude-correction factors stored in analyzer memory.

Key Access: (RECALL) or (SAVE)

(LOCAL)

Pressing CONFIG (LOCAL) after the analyzer has been placed in the remote mode places the analyzer in the local mode and enables front-panel control. During remote operation, "R" appears in the lower-right corner of the screen indicating remote and talk. A "T" or "L" may appear during remote operation, indicating talk or listen. Pressing the CONFIG key removes the "R" symbol in the lower-right corner.

Key Access: (CONFIG)

| MAIN | |
|--------------------|---|
| COIL DR | displays the output produced by the main-coil driver on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| Main Menu | returns the analyzer to the main menu. Key Access: SYSTEM MEAS or SWEEP |
| MAIN | |
| SPAN | displays the main-coil-span signal, MC_SPAN, from the span dividers on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| MAN TRK ADJUST | <i>Option 011 only.</i> allows the user to adjust the frequency of the tracking-generator oscillator manually using the step keys or knob. The tracking adjust is tuned to maximize the amplitude of the trace. |
| | Key Access: (AUX CTRL) |
| MARKER # ON OFF | turns the selected marker on or off. One of the four markers must first be selected by the SELECT 1 2 3 4 key, otherwise marker 1, or the last selected marker, will be turned on or off. When a new marker is turned on by pressing MARKER # ON OFF, the MK TRACE AUTO ABC function always switches to AUTO and the marker is put on the trace that is selected by the AUTO mode. |
| | Key Access: (MKR) |
| | |

MARKER Δ

activates a second marker at the position of the first marker. (If no marker is present, two markers appear at the center of the display.) The amplitude and frequency of the first marker is fixed, and the second marker is under your control. Annotation in the active function block and in the upper-right corner of the screen indicates the frequency and amplitude differences between the two markers. The display mode must not be changed between log and linear while using a delta marker.

NOTE

If there are already four markers when MARKER Δ is pressed, a nonactive marker disappears, the active marker becomes a reference marker, and the delta marker becomes the active marker.

Key Access: (MKR) or (PEAK SEARCH)

MARKER ∆→SPAN

sets the start and stop frequencies to the values of the delta markers. The start and stop frequencies will not be set if the delta marker is off.

Key Access: $(MKR \rightarrow)$

| MARKER→ AUTO FFT | functions exactly like the CONTINUS FFT softkey, if the spectrum analyzer is already in zero span. If the analyzer is not in zero span it activates a marker that must be placed on the signal that will have an FFT performed on it and MARKER — AUTO FFT must be pressed again. The resolution bandwidth setting must be wide enough to include the displayed modulation signals. Refer to Chapter 5 "Measuring Amplitude Modulation Using the Fast Fourier Transform" for more information. |
|---|---|
| | Press FFT OFF to return the analyzer to normal operation. The analyzer state prior to pressing MARKER \rightarrow AUTO FFT can be recalled from register 8. |
| | Key Access: (MEAS/USER) |
| $\begin{array}{l} \texttt{MARKER} \\ \rightarrow \texttt{CF} \end{array}$ | changes the analyzer settings so that the frequency at the marker becomes the center frequency. |
| | Key Access: $(MKR \rightarrow)$ or (PEAK SEARCH) |
| MARKER →CF STEP | changes the center-frequency step size to match the value of the active marker. Press (FREQUENCY) then CF STEP AUTO MAN to view the step size. If marker delta is active, the step size will be set to the frequency difference between the markers. |
| | Key Access: $(MKR \rightarrow)$ |

| MARKER→ FFT STOP | changes the FFT stop frequency to whatever the current value of the FFT marker frequency is, within the limitations of the available sweep times. This puts the marker on the right side of the graticule. |
|---------------------|---|
| | Key Access: (MEAS/USER) |
| MARKER→ MID SCRN | changes the frequency at the middle of the FFT display to whatever the current value of the FFT marker frequency is, within the limitations of the available sweep times. This puts the marker in the middle of the graticule. |
| | When using the FFT function the MARKER \rightarrow MID SCRN |
| | softkey replaces the MARKER \rightarrow CF softkey in the (PEAK SEARCH) and (MKR \rightarrow) menus. |
| | Key Access: (Meas/user), (MKR \rightarrow), or (peak search) |
| MARKER→ MINIMUM | moves the active marker to the minimum detected amplitude value. |
| | Key Access: (\overline{MKR}) |
| MARKER →PK-PK | finds and displays the frequency and amplitude differences between the highest and lowest trace points. Pressing MARKER \rightarrow PK-PK performs the routine similar to pressing the following keys: (PEAK SEARCH), MARKER Δ , and MARKER \rightarrow MINIMUM. |
| | Key Access: $(MKR \rightarrow)$ |

| MARKER →REF LVL | changes the analyzer settings so that the amplitude at the active marker becomes the reference level. |
|--------------------|---|
| | Key Access: $(MKR \rightarrow)$ |
| MARKER —START | changes the start frequency so that it is equal to the frequency of the active marker. This moves the active marker to the left edge of the display. |
| | Key Access: $(MKR \rightarrow)$ |
| MARKER →STOP | changes the stop frequency so that it is equal to the frequency of the active marker. This moves the active marker to the right edge of the display. |
| | Key Access: $(MKR \rightarrow)$ |
| MARKER ALL OFF | turns off all of the markers, including markers used for marker track and demodulation. Marker annotation is also removed. |
| | Key Access: MKR |
| MARKER AMPTD | keeps the active marker at the requested amplitude on the screen. |
| | Key Access: (MKR) |
| MARKER NORMAL | activates a single frequency marker at the center frequency on the active trace if an on-screen marker is not already displayed. If there is an on-screen marker before the MARKER NORMAL function is enabled, a frequency marker is enabled at the position of the first marker. |
| | Key Access: MKR |

| MARKER ON | <i>Option 107 only.</i> activates a marker in the gate utility. |
|----------------|---|
| | Key Access: (SWEEP) |
| MAX HOLD A | maintains the maximum level for each trace point of trace A. Updates each trace point if a new maximum level is detected in successive sweeps. |
| | Key Access: (TRACE) |
| MAX HOLD B | maintains the maximum level for each trace point of trace B. Updates each trace point if a new maximum level is detected in successive sweeps. |
| | Key Access: (TRACE) |
| MAX MXR LVL | lets you change the maximum input mixer level in 10 dB steps from 39 dBmV to -51 dBmV. |
| | Key Access: (AMPLITUDE) |
| MEAS OFF | turns off the measurement functions under the Power Menu softkey and restores the analyzer to the state prior to initiating the power measurement. |
| | Key Access: (MEAS/USER) |
| (MEAS/USER) | switches between the User Menu and the menu containing N dB PTS ON OFF, % AM ON OFF, TOI ON OFF, Power Menu, and FFT Menu. If no keys have been defined in the user menu, No User Menu is displayed. See the HP 8590 E-Series and L-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Programmer's Guide for more information about defining keys in the user menu |
| | |

Measure Now

accesses the functions that allow you to initiate the testing immediately (instead of initiating testing at a later time) and send the test results to the screen, a printer, or to a RAM card.

Pressing Measure Now will display a screen that allows the option of selecting CTB testing. Then, the DUMP TO SCREEN, DUMP TO PRINTER, DUMP TO CARD, and Main Menu softkeys are accessed.

Key Access: SYSTEM MEAS

MEM LOCKED

indicates that the save lock function is on. It replaces the ERASE STATEALL and ERASE TRACEALL softkeys under the CONFIG key, and it replaces STATE \rightarrow INTERNAL and TRACE \rightarrow INTERNAL under the SAVE key when SAV LOCK ON OFF is ON. Pressing SAV LOCK ON OFF returns the menu to its unlocked state.

Key Access: (SAVE) or (CONFIG)

| | | | | | | | | | | | | | | | r | | | |
|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|---|--|--|--|
| | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | - | | | |
| | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | |

maintains the minimum level for each trace point of trace C. Updates each trace point if a new minimum level is detected in successive sweeps.

Key Access: (TRACE)

MIN HOLDputs the analyzer into the sample detection mode and
continuously places the minimum level at each trace point of
trace A into trace B. Once activated, the function continues
until turned off with MIN HOLD OFF.

Key Access: CHANNEL MEAS

| MK COUNT | |
|--------------------|---|
| ON OFF | turns on the marker counter when ON is underlined. If no marker is active before MK COUNT ON OFF is pressed, a marker is activated at center screen. Press |
| | MK COUNT ON OFF (so that OFF is underlined), to turn the |
| | marker counter off. Press CNT RES AUTO MAN to change the marker counter resolution to an uncoupled value. |
| | Key Access: (MKR FCTN) |
| MK NOISE | |
| ON OFF | reads out the average noise level, referenced to a 1 Hz noise power bandwidth, at the marker position. If no marker is present, a marker appears at the center of the screen. The root-mean-square noise level, normalized to a 1 Hz noise power bandwidth, is read out. The sample detector is activated. |
| | Key Access: (MKR FCTN) |
| MK PAUSE | |
| ON OFF | stops the analyzer sweep at the marker position for the duration of the dwell time. The dwell time can be set from 2 milliseconds to 100 seconds. |
| | Key Access: (MKR FCTN) |
| MK READ F T I P | selects the marker readout to be displayed in signal frequency, sweep time, the inverse of the sweep time, or the period that is the inverse of the frequency. When the instrument is in zero span the frequency type readout cannot be selected. |
| | Key Access: (MKR) |

MK TABLE

ON OFF provides a list of the four available markers that is updated at the end of each sweep or when a marker is used. The marker data can be displayed in different formats. It can be in absolute or delta frequency and amplitude formats. There is also a delta display line format, which can be accessed using the TABLE ΔDL NRM softkey.

The marker table is not saved with the **SAVE** and **RECALL** keys.

Key Access: (MKR FCTN)

MK TRACE AUTO ABC

assigns a marker to a trace. Pressing MK TRACE AUTO ABC will activate a marker on trace A if there are no markers turned on. If a marker is currently active, press MK TRACE AUTO ABC until A, B, or C are underlined. The active marker will be moved to the selected trace.

Selecting the AUTO mode will move the marker to the trace that is automatically selected. The selection order is to first look for a trace in the clear-write mode, in the order of trace A, then trace B, then trace C. If no traces are currently being written, it will select a trace in the view-store mode, again in the order of trace A, B, then C.

Key Access: MKR

MK TRACK ON OFF

moves the signal that is nearest to the active marker to the center of the screen and keeps the signal there. MKR-TRK or CNTR-TRK appears in the upper-right corner of the display. An (*) may appear in the upper-right corner of the display while the analyzer is verifying that it has the correct signal.

Key Access: (MKR FCTN)

| (MKR) | |
|---------------------|--|
| | accesses the marker control softkeys that select the type and number of markers and turn them on and off. Markers are diamond-shaped characters that identify points of traces and allow the traces to be manipulated and controlled on the screen. During manual operation, four markers may appear on the display simultaneously; only one can be controlled at a time. The marker that is controlled is called the "active" marker. Pressing MKR activates the MARKER NORMAL softkey. |
| $(MKR \rightarrow)$ | (read "marker to") accesses the softkeys used for the transfer of marker information directly into other functions. |
| (MKR FCTN) | accesses the marker function softkeys. These softkeys can be used to access the marker table and to turn on marker functions for tracking the signal and counting its frequency. Noise markers and the marker pause function are also accessed under (MKR FCTN). |
| MNOPQR | accesses the softkey menu for selecting screen title or prefix characters M through R. |
| | Key Access: (CAL), (CONFIG), (DISPLAY), (RECALL), or (SAVE) |
| MODE | changes the softkey menus for the analyzer mode and other modes of operation when SPECTRUM ANALYZER (located under PRESET) and PRESET SPECTRUM are selected, respectively. Other modes are available using the downloadable measurement personalities. The HP 85721A Cable Television Measurements Personality is an example of these modes. Consult the documentation accompanying the personality for information about these other modes of operation. |
| Monitor TV | allows you to access the TV receiver mode. This feature is the same as the TV RECEIVER mode of operation except that it is accessed from a front-panel hardkey. |
|--------------------|---|
| | Press AUX CTRL to access Monitor TV. |
| | Key Access: (AUX CTRL) |
| N dB PTS ON OFF | automatically places two markers at points N dB from the highest point on the highest displayed signal, and determines the frequency difference between the two markers. N dB is the active function and the value of N is set by the user. The measurement defaults to 3 dB when it is first turned on. The measurement runs continuously re-executing at the end of each sweep. |
| | Key Access: (MEAS/USER) |
| NEW EDIT | Refer to the <i>HP 8590 E-Series and L-Series Spectrum</i> <i>Analyzers and HP 8591C Cable TV Analyzer Programmer's</i> <i>Guide</i> for more information. starts the DLP editor function, clearing the DLP editor memory to create a new item in the analyzer 2500 byte DLP editor memory. The item will not be in the analyzer user memory until it is processed by the SAVE EDIT softkey. The DLP editor memory buffer remains intact when the instrument is preset and when it is powered off. |
| | Key Access: (RECALL) or (SAVE) |
| (NEXT) | switches the active window between the two displayed windows, if the windows display mode has been turned on. The active window is marked by solid lines. |
| | |

| NEXT | | | |
|------------------|--|--|--|
| PAGE | allows you to select the next page of on-screen test results. | | |
| | Press CABLE TV ANALYZER, SYSTEM MEAS, | | |
| | Measure Now, Yes or No, DUMP TO SCREEN to access | | |
| | NEXT PAGE . | | |
| | Key Access: SYSTEM MEAS | | |
| NEXT | | | |
| PEAK | places the marker on the next highest peak. The signal peak must exceed the threshold value. (Also see the PEAK EXCURSN and THRESHLD ON OFF softkey descriptions.) | | |
| | Key Access: (AUX CTRL) or (PEAK SEARCH) | | |
| NEXT PK LEFT | moves the marker to the next peak to the left of the current marker. The signal peak must exceed the threshold value. If there is no peak to the left, the marker will not move. (Also see the PEAK EXCURSN and THRESHLD ON OFF softkey descriptions.) | | |
| | Key Access: (AUX CTRL) or (PEAK SEARCH) | | |
| NEXT PK RIGHT | moves the marker to the next peak to the right of the current marker. The signal peak must exceed the threshold value. If there is no peak to the right, the marker will not move. (Also see the PEAK EXCURSN and THRESHLD ON OFF softkey descriptions.) | | |
| | Key Access: [PEAK SEARCH] | | |

| NO | allows you to select not to do the function. If you press CLEAR TST PLAN and then press NO, you have selected not to clear the current test plan from analyzer memory. If you enter a channel by using ENTER BY CHANNEL or ENTER BY RANGE and then press NO, the channel or range of channels will be treated as channels that are not scrambled. NO is accessed either by pressing CLEAR TST PLAN or by entering a channel number with ENTER BY CHANNEL or |
|----------------------|---|
| | ENTER BY RANGE . |
| | Key Access: SYSTEM MEAS |
| No User Menu | is displayed if key number 1 has not been defined by the user. Key number 1 can be defined by remote programming commands (KEYCMD or KEYDEF). |
| | Key Access: (MEAS/USER) |
| NORMLIZE ON OFF | subtracts trace B from trace A and adds the result to the display line. The result is displayed in trace A. |
| | Key Access: (TRACE) |
| NORMLIZE POSITION | displays the display line and makes the display line function active. The trace data is normalized with respect to the display line even if the value of the display line is changed. |
| | Key Access: (TRACE) |
| NTSC | allows you to trigger on the NTSC video format. Key Access: (TRIG) |
| | |

| OCC BW % POWER | allows the user to enter the percent of the power desired when using the occupied bandwidth measurement under the Power Menu softkey. |
|----------------------|--|
| | Key Access: (MEAS/USER) |
| OCCUPIED BANDWDTH | integrates the power of the displayed and puts markers at the frequencies containing a selected percent of the power. The measurement defaults to 99% of the occupied bandwidth power. |
| | Key Access: (MEAS/USER) |
| PAINTJET PRINTER | selects a color print (for use with an HP PaintJet printer only). The traces are displayed in orange (trace A), blue (trace B), and red (trace C). The graticule, screen annotation, and user information are displayed in black. |
| | Key Access: (CONFIG) |
| PAL | allows you to trigger on the PAL video format. |
| | Key Access: (Trig) |
| PAL-M | allows you to trigger on the PAL-M video format. Key Access: (TRIG) |
| | |

| PARAM AUTO MAN | lets the user choose between automatically or manually setting the parameters used for the measurement functions under the Power Menu softkey. |
|-------------------|---|
| | Key Access: (MEAS/USER) |
| PEAK EXCURSN | sets the minimum amplitude variation of signals that the marker can identify as a peak. |
| | Key Access: (PEAK SEARCH) |
| Peak | |
| Menu | accesses the same softkeys that are available when (PEAK SEARCH) is pressed (see the key description for |
| | (PEAK SEARCH) below). Pressing Peak Menu instead of (PEAK SEARCH) allows you to use the peak-search functions without initiating a new peak search. |
| | Key Access: $(MKR \rightarrow)$ |
| (PEAK SEARCH) | automatically places a marker on the highest amplitude of a trace, displays the marker's amplitude and frequency. It also accesses the menus of marker peak functions including the |

peak table functions.

| PEAK ZOOM | finds the highest displayed signal and narrows the span to a value selected by the user. |
|--------------------|--|
| | Key Access: (SPAN) |
| PICTURE ON OFF | allows you to turn the TV picture on and off. This feature is the same as the TV RECEIVER mode of operation except that it is accessed from a front-panel hardkey. |
| | Press (AUX CTRL), Monitor TV to access PICTURE ON OFF. |
| | Key Access: (AUX CTRL) |
| PIX &SND ON OFF | allows you to view the TV picture and listen to the sound simultaneously. This feature allows you to quickly identify picture quality problems such as noise, distortion, hum, and ingress. Select a channel by pressing CHANNEL SELECT, entering a TV channel number, then pressing PIX &SND ON OFF. To exit this mode, press any softkey |
| | Press TV BECEIVER to access PIX &SND ON OFF |
| | Key Access: TV RECEIVER |

PK MODE <>DL NRM

select which peaks will be listed in the peak table to include all peaks, or to exclude the peaks that are either above or below the display line. See table below. The display line is activated if it was not currently being displayed.

| Selection | Peaks Listed |
|-----------|---------------------------------|
| NRM | all peaks listed |
| >DL | peaks above display line listed |
| < DL | peaks below display line listed |

Key Access: (PEAK SEARCH)

| PK SORT FRQ AMP | switches the peak table sorting routine between listing the peaks in order by descending amplitude or by ascending frequency. |
|--------------------|--|
| | Key Access: (PEAK SEARCH) |
| PK TABLE ON OFF | displays a list, of up to ten signal peaks, that is updated at the end of each sweep. The peaks can be sorted in order by descending amplitude or by ascending frequency. Peaks above or below the display line can be excluded from the table. The peak table function works with trace A only. The peak table is not saved with the <u>SAVE</u> and <u>RECALL</u> keys. |
| | Key Access: (peak search) |

PLAN 1

allows you to select test plan 1 as the test plan number for either saving the current test plan or recalling a previously saved test plan.

Saving the current test plan overwrites any test plan previously saved in plan 1. When you save the current test plan in test plan 1, it is stored in trace register 6.

Key Access: SYSTEM MEAS

PLAN 2 allows you to select test plan 2 as the test plan number for either saving the current test plan or recalling a previously saved test plan.

The operation of **PLAN 2** is similar to **PLAN 1** except that the test plan is stored in and recalled from trace register 7.

Key Access: SYSTEM MEAS

PLAN 3 allows you to select test plan 3 as the test plan number for either saving the current test plan or recalling a previously saved test plan.

The operation of **PLAN 3** is similar to **PLAN 1** except that the test plan is stored in and recalled from trace register 8.

Key Access: SYSTEM MEAS

| FLAN 4 | allows you to select test plan 4 as the test plan number for either saving the current test plan or recalling a previously saved test plan. |
|--------------------|---|
| | The operation of PLAN 4 is similar to PLAN 1 except that the test plan is stored in and recalled from trace register 9. |
| | Key Access: SYSTEM MEAS |
| PLAN 5 | allows you to select test plan 5 as the test plan number for either saving the current test plan or recalling a previously saved test plan. |
| | The operation of PLAN 5 is similar to PLAN 1 except that the test plan is stored in and recalled from trace register 10. |
| | Key Access: SYSTEM MEAS |
| Plot Config | accesses the menu used to address the plotter and to select plotter options. See the (\overline{COPY}) key for more information. |
| | Key Access: CONFIG |
| PLOTTER ADDRESS | <i>Option 041 only.</i> changes the HP-IB address of the plotter. The plotter address is set to 5 when DEFAULT CONFIG is pressed. |
| | Key Access: CONFIG |

| PLT LOC | selects the position of the plotter output. The highlighted portion of the softkey label indicates where the plot is to be output on the page. This softkey function appears only if two or four plots per page are selected when PLTS/PG 1 2 4 is pressed. |
|----------------------|---|
| | Key Access: (CONFIG) |
| PLT MENU ON OFF | allows the softkey labels to be plotted along with the analyzer display. This function operates when the COPY key is used in a plot configuration. The PLT MENU ON OFF function is set to OFF when DEFAULT CONFIG is pressed. |
| | Key Access: (CONFIG) |
| Plt Port Config | Option 041 or 043 only. accesses the menus used by Option 041 (HP-IB and Parallel Interface) and Standard Option 043 (RS-232 and Parallel Interface) to select plotter options. See the COPY key for more information. Front-Panel Key Access: CONFIG |
| PLT PORT HPIB PAR | <i>Option 041 only.</i> allows you to select between HP-IB or parallel plotter ports. |
| PLT PORT SER PAR | allows you to select between serial or parallel plotter ports. The BAUD RATE softkey appears in this menu only when serial is selected. |

PLTS/PG 124 allows you to plot a full-page, half-page, or quarter-page output. Selecting two plots per page requires a plotter that has the rotate command (RO). The plotter will be set to a full-page output when **DEFAULT CONFIG** is pressed. Key Access: (CONFIG) PLT->LJT allows you to plot a full-page, half-page, or quarter-page ON OFF output to an HP LaserJet printer. This softkey can be accessed by pressing (CONFIG), then Plot Config. Key Access: (CONFIG) POINT specifies a limit value for one coordinate point, so that a POINT segment specifies a limit value for a single frequency or time. Key Access: (DISPLAY)

Power
Menuaccesses functions that make transmitter power
measurements. The measurements are designed for analog
radio or continuous carrier digital radio signals.
Key Access: (MEAS/USER)POWER ON
IP LASTdetermines the state of the analyzer when the analyzer is
powered on.

NOTE

If you have a downloadable program or "personality" installed in analyzer memory, the following changes apply to the operation of the POWER ON function: When using a downloadable program or personality, the last state of the personality is not recalled. We recommend that if you are using a downloadable program or personality, you set **POWER ON IP LAST** to IP. If

POWER ON IP LAST is set to LAST, you must press (**PRESET**) whenever you power on the analyzer.

Key Access: (CONFIG)

PRESET

provides a convenient starting point for making most measurements. Pressing (PRESET) displays softkeys used for accessing the operating modes available for your analyzer. See Table 4-6 and Table 4-5 for the conditions established by pressing (PRESET).

The instrument preset function performs a processor test, but does not affect CAL data. Pressing **PRESET** clears both the input and output buffers, but does not clear trace B. The amplitude values of trace C are set to the reference level. Amplitude-correction factors are turned off. Limit-line testing is turned off, but the limit-line tables remain in analyzer memory. The status byte is set to 0. Instrument preset affects all operating modes. (See the key description for **MODE** for more information about other operating modes.) Pressing **PRESET** erases all "on time" functions—ONCYCLE, ONDELAY, ONEOS, ONMKR, ONSRQ, ONSWP, ONTIME, and TRMATH. These are remote programming commands. See the *HP 8590 E-Series and L-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Programmer's Guide* for more information.

NOTE

Turning the analyzer on performs an instrument preset. Turning on the analyzer also fetches CAL data; completes a processor test; clears trace B, trace C, and both the input and output buffers; turns off amplitude correction factors; turns off limit-line testing; and sets the status byte to 0. The last state of the analyzer (before it was switched off) is recalled, unless IP has been set by the POWER ON function.

Table 4.5. Model Specific Preset Conditions

| Model | Center Frequency | Span | Start Frequency | Stop Frequency | Sweep Time |
|----------|---------------------|---------|--------------------|-------------------|---------------------|
| HP 8591C | 900 MHz | 1.8 GHz | 0 Hz | 1.8 GHz | 20 ms auto-coupled |

Table 4.6. Common Preset Conditions

| $A - B \rightarrow A$ | off | | | |
|---------------------------------------|--|--|--|--|
| Amplitude correction factors | off | | | |
| Amplitude units | default values | | | |
| Annotation and graticule display | on | | | |
| Attenuation | 10 dB auto-coupled | | | |
| Center frequency | Refer to Table 4–5. | | | |
| CF step size | 10% of span | | | |
| Coupled functions | all set to AUTO | | | |
| Mass storage device card or internal | INTERNAL | | | |
| Detector | positive peak | | | |
| Display line level | 2.5 graticule divisions below reference level, display off | | | |
| Frequency offset | 0 Hz | | | |
| Limit-line testing | off | | | |
| LIMIHI and LIMILO | cleared | | | |
| Log scale | 10 dB/division | | | |
| Marker counter | off | | | |
| Marker counter resolution | au to -coupled | | | |
| Markers | off | | | |
| Mixer level | 39 d Bm V | | | |
| Operating mode | analyzer | | | |
| Reference level | 49 dBmV in power-on units | | | |
| Reference level offset | O dB | | | |
| Reference level position | top 8th graticule | | | |
| Resolution bandwidth | 3 MHz auto-coupled | | | |
| Span | Refer to Table 4-5. | | | |

| SRQ mask | octal 50 |
|---------------------|---|
| Start Frequency | Refer to Table 4-5. |
| Stop Frequency | Refer to Table 4-5. |
| State registers 1—8 | unaffected |
| Sweep | con tinuo us |
| Threshold level | one graticule above baseline, display off |
| Title | cleared |
| Trace A | clear-write |
| Trace B | s to re-blank |
| Trace C | store-blank, at reference level |
| Trace registers | unaffected |
| Trigger | free run |
| VBW/RBW ratio | 0.3 |
| Video averaging | off |
| Video bandwidth | 1 MHz auto-coupled |

Table 4.6. Common Preset Conditions (continued)

PRESET SPECTRUM

allows the analyzer mode only to be preset. Table 4-7 lists the conditions affected by the PRESET function. Other operating modes will not be affected. See the description of the (MODE) key.

In addition, pressing **PRESET SPECTRUM** erases user-generated graphics and blanks the active-function block that is on the analyzer screen.

Pressing PRESET SPECTRUM disposes of ONEOS, ONSWP, and TRMATH. These are remote programming commands; see the HP 8590 E-Series and L-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Programmer's Guide for more information.

Key Access: (MODE) or (PRESET)

| $A - B \rightarrow A$ | off |
|--------------------------------|---------------------|
| Analog+ display mode | off |
| Annotation and graticule | on |
| Attenuation | coupled |
| Center frequency | Refer to Table 4-5. |
| Center frequency step size | 10% of span |
| Coupled functions | all set to AUTO |
| Detector | positive peak |
| Display line | off |
| Frequency offset | 0 Hz |
| Limit-line testing | off |
| Scale | log 10 dB/div |
| Marker counter | off |
| Marker counter resolution | 2 kHz auto-coupled |
| Markers | off |
| Measure | SA analyzer |
| On end-of-sweep command ONEOS | cleared |
| On-sweep command ONSWP | cleared |

Table 4.7. Preset Conditions for All Models

| Reference level | 49 dBmV in power-on units |
|---|-------------------------------|
| Reference level offset | O dB |
| Reference level position | top 8th graticule |
| Resolution bandwidth | 3 MHz coupled |
| Span | Refer to Table 4-5. |
| Start frequency | Refer to Table 4-5. |
| Stop frequency | Refer to Table 4-5. |
| State registers 1—8 | unaffected |
| S weep | continuous |
| Sweep time | Refer to Table 4-5. (coupled) |
| Threshold | off |
| Trace A | clear-write |
| Trace B | store-blank |
| Trace C | store-blank |
| Trace math command TRMATH | cleared |
| Trace registers | unaffected |
| Trigger | free |
| Video averaging | off |
| Video bandwidth | 1 MHz (coupled) |
| Video bandwidth to resolution bandwidth ratio | 0.3 |

Table 4.7. Preset Conditions for All Models (continued)

| Prev | |
|------|--|
| Menu | allows you to access the previous system measurement |
| | menu. |
| | Key Access: SYSTEM MEAS |
| PREV | |
| PAGE | allows you to select the previous page of on-screen test results. |
| | Press CABLE TV ANALYZER, SYSTEM MEAS, |
| | Measure Now, Yes or No, DUMP TO SCREEN to access |
| | PREV PAGE. |
| | Key Access: SYSTEM MEAS |

| Print Config | accesses the softkey functions that are used to address the printer, select a black and white print or a color print, and reset the printer. See the COPY key for more information. |
|--------------------|---|
| | Key Access: (CONFIG) |
| PRINT REPORT | allows you to print a report of the currently displayed test results. |
| | Press CABLE TV ANALYZER, SYSTEM MEAS, |
| | Measure Now, Yes or No, DUMP TO SCREEN to access PRINT REPORT. |
| | Key Access: SYSTEM MEAS |
| PRINTER ADDRESS | <i>Option 041 only.</i> allows you to change the HP-IB address of the printer. The printer address is set to 1 by pressing DEFAULT CONFIG . |
| | Key Access: (CONFIG) |
| PRINTER SETUP | resets the printer, sets the printer to 60 lines per page, and skips line perforations. This function enables you to obtain up to three printouts per page. The printer paper should be at the top of the form before using this function. The PRINTER SETUP function may not work with printers that are not recommended (see Chapter 9 for recommended printers). |
| | Key Access: (CONFIG) |

| Prn Port Config | Option 041 or 043 only. accesses the menus used by Option 041 (HP-IB and Parallel Interface) and Standard Option 043 (RS-232 and Parallel Interface) to select printer options. See the COPY key for more information. Front-Panel Key Access: CONFIG |
|----------------------|---|
| PRN PORT HPIB PAR | <i>Option 041 only.</i> allows you to select between HP-IB or parallel printer ports. |
| PRN PORT SER PAR | allows you to select between serial or parallel printer ports. The BAUD RATE softkey appears in this menu only when serial is selected. |
| PRT MENU ON OFF | allows the softkey labels to be printed along with the analyzer display. This function operates when the <u>COPY</u> key is used in a print configuration. The PRT MENU function is set to ON when DEFAULT CONFIG is pressed. |
| Pulse Param | accesses the gate utility menus for entering the pulse parameters: reference edge, pulse width, and pulse repetition interval. If pulse parameters have previously been entered, pressing Pulse Param causes a list of the current values to be displayed. |
| | Key Access: (SWEEP) |

| PURGE | |
|---------|---|
| AMP COR | clears the current amplitude-correction factors table. |
| | Pressing PURGE AMP COR displays the message: If |
| | you are sure, press key again to purge data. |
| | Pressing PURGE AMP COR a second time clears the |
| | amplitude-correction data. Press SAVE AMP COR |
| | to save amplitude-correction factors, and then press |
| | PURGE AMP COR to clear the current amplitude-correction |
| | factors table. |
| | Key Access: (CAL) |
| PURGE | |
| LIMITS | clears the current limit-line table from analyzer memory. |
| | Pressing PURGE LIMITS displays the message: If you |
| | are sure, press key again to purge data. Press |
| | PURGE LIMITS again if you wish to clear the current |
| | limit-line table. Press SAVE LIMIT to save the current |

limit-line table, and then press **PURGE LIMITS** to clear the

current limit-line table. Key Access: (DISPLAY)

| PWRGRAPH ON OFF | turns the channel power graph ON or OFF. With the PWRGRAPH ON, the channel powergraph is calculated and displayed and the numeric results are not displayed. The value of the channel power is displayed at the selected marker frequency. This graph function is used after doing a channel power measurement with the CHANNEL POWER softkey. |
|--------------------|---|
| | Key Access: (MEAS/USER) |
| PWR SWP ON OFF | Option 011 only. activates (ON) or deactivates (OFF) the power-sweep function, which sweeps the output power of the tracking generator over the selected power-sweep range. The value of the power-sweep range is displayed in the active-function block when PWR SWP ON OFF is turned on. |
| | Key Access: (AUX CTRL) |
| RECALL | accesses softkey menus that allow you to recall data from the memory card or analyzer memory. |
| RECALL AMP COR | recalls an amplitude-correction factors table from the current mass-storage device (analyzer memory or memory card). Key Access: (CAL) |

Recall Data

allows you to access DATA TO PRINTER and

DATA TO SCREEN. DATA TO PRINTER sends the test results from a file on a RAM card to a printer. DATA TO SCREEN sends the test results from a file on a RAM card to the screen.

Press CABLE TV ANALYZER, SYSTEM MEAS,

More 1 of 2 to access Recall Data.

Key Access: SYSTEM MEAS

RECALL FRM CRD

allows you to recall a previously saved test plan from a RAM card.

Recalling a test plan overwrites the current test plan. If you do not want the current test plan to be overwritten, save the current test plan before recalling a test plan.

Press CABLE TV ANALYZER, SYSTEM MEAS, Recall TstPlan to access RECALL FRM CRD.

Key Access: SYSTEM MEAS

RECALLFor channel-measurement mode it recalls a trace into trace BFRM INTfor use as a reference trace or to recall a system frequency
response trace.

This function simplifies recalling a trace from the analyzers internal memory. A trace can also be recalled from a memory card using procedures provided in the analyzer user's guide.

For system-measurement mode it allows you to recall a previously saved test plan (from plans 1 through 5) from analyzer memory.

Recalling a test plan overwrites the current test plan. If you do not want the current test plan to be overwritten, save the current test plan before recalling a new one. See "Step 3. Save the current test plan in internal memory" in Chapter 3 of the *HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide* for more information.

The test plans are saved in trace registers 6 through 10.

The channel numbers, tests, scrambling status, channel tuning configuration, quiet line number, frequency response line number, test signal line number, and field selection are saved in the test plan; the location code, the temperature, the selection of CTB testing, and the setting for an external pad (EXT PAD YES NO) are stored in nonvolatile analyzer memory and are retained until changed.

Key Access: CHANNEL MEAS or SYSTEM MEAS

| RECALL LIMIT | recalls limit-line tables from the current mass-storage device (analyzer memory or memory card). |
|--------------------|--|
| | Key Access: (DISPLAY) |
| Recall TstPlan | allows you to access the softkeys used to recall a previously saved test plan from analyzer memory, or from a RAM card. |
| | Recalling a test plan overwrites the current test plan. If you do not want the current test plan to be overwritten, save the current test plan before recalling a new one. |
| | Press CABLE TV ANALYZER, SYSTEM MEAS to access |
| | Recall TstPlan. |
| | Key Access: SYSTEM MEAS |
| REF LVL | allows the reference level to be changed. |
| | Key Access: (AMPLITUDE) |
| REF LVL OFFSET | adds an offset value to the displayed reference level. |
| | Key Access: (AMPLITUDE) |
| RES BW AUTO MAN | changes the analyzer 3 dB resolution bandwidth. |
| | Key Access: (AUTO COUPLE) or BW |

| RPG TITLE | provides additional characters for the Change Title function. |
|--------------------|---|
| | Key Access: (CAL) or (DISPLAY) |
| SAV LOCK ON OFF | locks all the current internal state and trace registers against further data storage, when ON is underlined. |

NOTE

When SAV LOCK ON OFF is set to ON, none of the state registers, (1-8), can be overwritten. The analyzer automatically updates state register nine with the last state. ERASE MEM ALL ignores the state of the SAV LOCK. So, even if STATES and TRACES are locked, they will still be erased by ERASE MEM ALL.

Key Access: (SAVE)

(SAVE)

accesses softkey menus that allow you to store state data, trace data, limit-line tables, and amplitude-correction factors on a memory card or in analyzer memory. The SAVE function also allows you to save state data, trace data, limit-line tables, amplitude-correction factors, and program data on the memory card. In addition, pressing <u>SAVE</u> accesses the softkey menus used to catalog the saved data in analyzer memory or on the memory card.

| SAVE AMP COR | saves the current amplitude-correction factors table to the current mass-storage device (analyzer memory or memory card). |
|--------------------|---|
| | Key Access: (CAL) |
| SAVE EDIT | Refer to the HP 8590 E-Series and L-Series Spectrum Analyzers and HP 8591C Cable TV Analyzer Programmer's Guide for more information. passes the text from the DLP editor memory through the parser to execute as analyzer commands. |
| | Key Access: (RECALL) or (SAVE) |
| SAVE LIMIT | saves the current limit-line tables in the current mass-storage device (analyzer memory or memory card). |
| | Key Access: (DISPLAY) |
| SAVE TO INT MEM | allows you to save the current test plan in analyzer memory. |
| | Saving a test plan overwrites any previously saved test plan of the same test plan number. |
| | Up to five test plans can be saved in analyzer memory. The test plans are saved in nonvolatile analyzer memory, in trace registers 6 through 10. |
| | The channel numbers, tests, scrambling status, channel tuning configuration, quiet line number, frequency response line number, test signal line number, and field selection are saved in the test plan; the location code, the temperature, the selection of CTB testing, and the setting for an external pad (EXT PAD YES NO) are stored in nonvolatile analyzer memory and are retained until changed. |
| | Pressing CABLE TV ANALYZER, SYSTEM MEAS, More 1 of 2 Save TetPlan accesses |
| | SAVE TO INT MEM. |
| | Key Access: SYSTEM MEAS |

| SAVE TO MEM CRD | allows you to store the on-screen test results to a RAM card. |
|--------------------|--|
| | Press CABLE TV ANALYZER, SYSTEM MEAS, |
| | Measure Now, Yes or No, DUMP TO SCREEN to access |
| | STORE TO CARD. |
| | Key Access: SYSTEM MEAS |
| SAVE TRB TO INT | stores trace B in internal memory. |
| | This function simplifies saving either a reference trace or a test result to internal memory. This trace can also be saved to a memory card using procedures provided in the analyzer user's guide. |
| | Key Access: CHANNEL MEAS |

Save allows you to access SAVE TO INT MEM and TstPlan SAVE TO MEM CRD. Pressing CABLE TV ANALYZER, SYSTEM MEAS, More 1 of 2, Create TstPlan accesses Save TstPlan. Key Access: SYSTEM MEAS SCALE scales the vertical graticule divisions in logarithmic units LOG LIN when LOG is underlined. When the SCALE LOG LIN function is the active function, the logarithmic units per division can be changed. Values may range from 0.1 to 20 dB per division. When LIN is underlined, the vertical scale is in linear mode that has a range of 1 kW to 1 pW. The reference-level value is set to the top of the screen and the bottom graticule becomes zero volts. (Each division of the graticule is one-eighth of the reference level in volts.) Key Access: (AMPLITUDE)

SECAM-L

triggers on the SECAM-L video formats.

Key Access: (TRIG)

SELECT selects one of the four possible markers. A marker can be 1234 turned on once it is selected. A marker that has already been turned on will become active when it is selected. If a marker has already been turned on and assigned to a specific trace it will become active on that trace and the MK TRACE AUTO ABC softkey will have the appropriate trace letter underlined. Key Access: MKR SELECT allows you to enter either the amplitude value for the AMPLITUD displayed (upper or lower) limit-line segment or the amplitude value for the current amplitude-correction point. Enter the amplitude value for the selected frequency or time by using the data keys. Change an amplitude value by using the step keys or the knob. Press (BK SP) to correct errors. Key Access: (CAL) or (DISPLAY) SELECT allows you to enter the delta amplitude value. The middle DLT AMPL amplitude value and the delta amplitude value create an upper and lower limit-line segment. The default value is 0. Key Access: (DISPLAY)

| SELECT FREQ | allows you to enter the frequency value for a limit-line |
|----------------|--|
| | segment or for an amplitude-correction point. |
| | Key Access: (CAL) or (DISPLAY) |
| SELECT | |
| LWR AMPL | allows you to enter the amplitude value for the lower limit-line segment. |
| | Key Access: (<u>DISPLAY</u>) |
| SELECT | |
| MID AMPL | allows you to enter the middle amplitude value. The middle amplitude value and the delta amplitude value create upper and lower limit-line segments. |
| | Key Access: (<u>DISPLAY</u>) |
| SELECT | |
| POINT | allows you to create or edit an amplitude-correction factor data point. |
| | Key Access: (CAL) |
| SELECT | |
| PREFIX | allows you to select an already existing prefix of a cataloged file and changes the current prefix to this selected prefix. |
| | Key Access: (RECALL) or (SAVE) |
| SELECT | |
| SEGMENT | allows you to create or edit a limit-line segment. |
| | Key Access: (DISPLAY) |
| SELECT | |
| TIME | allows you to enter the time value for a limit-line segment. |
| | Key Access: (DISPLAY) |

| SELECT TYPE | accesses the softkey menu used to select the limit-line type of line. |
|--------------------|---|
| | Key Access: (DISPLAY) |
| SELECT UPR AMPL | allows you to enter the amplitude value for the upper limit-line segment. |
| | Key Access: (DISPLAY) |
| Service Cal | accesses several service calibration functions. The service calibration functions are designed for service use only. More detailed descriptions of the service functions are available in the service documentation. Service documentation can be obtained by ordering Option 915 through your HP Sales and Service office. For a listing of all available service calibration functions, refer to "Service Functions" at the beginning of this chapter. |
| | Key Access: CAL |
| Service Diag | accesses several service diagnostic functions. The service diagnostic functions are designed for service use only. More detailed descriptions of the service diagnostic functions are available in the service documentation. Service documentation can be obtained by ordering Option 915 through your HP Sales and Service office. For a listing of all available service diagnostic functions, refer to "Service Functions" at the beginning of this chapter. |

Key Access: CAL

| SET ATTN ERROR | sets the calibration attenuator-error factors (this is not the same as the input attenuator). This is a service calibration function and is for service use only. |
|---------------------|--|
| | Key Access: (CAL) |
| Set B&W Printer | accesses the softkeys for setting up black and white HP and Epson compatible printers. |
| | Key Access: CONFIG |
| Set Colr Printer | accesses the softkey for setting up the HP paintjet printer. |
| | Rey Access. (CONFIG) |
| DATE | allows you to set the date of the real-time clock. Enter the date in the YYMMDD format using the number keypad and press ENTER. Valid year (YY) values are 00 through 99. Valid month (MM) values are from 01 to 12, and valid day values are from 01 to 31. |
| | Key Access: (CONFIG) |
| SET TIME | allows you to set the time of the real-time clock. Enter the time in 24 hour, HHMMSS format, using the number keypad and pressing (ENTER). Valid hour (HH) values are from 00 to 23. Valid minute (MM) and second (SS) values are from 00 to 59. |
| | Key Access: (CONFIG) |

| Setup | accesses the menu used to set up parameters specific to the power measurements |
|--------------------|---|
| | If another front-panel key is pressed, exiting the power menus, press the (MEAS/USER) key twice to return to the last power menu that was being used. |
| | Key Access: (MEAS/USER) |
| Setup | (For Cable TV Analyzer) allows you to access the softkeys used to set up defaults for channel tuning configuration, input configuration, non-interfering RF and video measurement configurations, and channel to test for system beats. |
| | The channel tuning configuration softkeys include STD, AIR, IRC, HRC, T, USR CH O CNTR FRQ, and USR CH O SPAN. |
| | The analyzer input configuration softkeys include EXT PAD YES NO (50 Ω spectrum analyzers only), |
| | EXT AMP YES NO, and TTL PWR @ INPUT. |
| | The non-interfering RF and video measurement configuration softkeys include GATING YES NO and PAUSE YES NO . |
| | The SYS BTS TST CH# softkey is used to specify the channel you wish to use for testing beats. |
| | These settings are stored in nonvolatile analyzer memory. This means that the analyzer retains the selections, even when power is turned off, until a new selection is made. |
| | Press CABLE TV ANALYZER to access Setup. |
| | Key Access: CABLE TV ANALYZER |
| (<u>SGL SWP</u>) | changes the sweep control to single sweep if the analyzer is in the continuous sweep mode. It executes a sweep after the trigger condition is met. |

| SHOW | |
|---------|---|
| OPTIONS | displays the number and description of the options installed in your analyzer, the instrument model number of the analyzer, the last five digits of the analyzer serial number, and the firmware revision. |
| | Key Access: (CONFIG) |
| SIGNAL | |
| ID | activates an FFT marker that must be put on a signal to verify that it is not being displayed at the wrong frequency due to aliasing. |
| | Key Access: (MEAS/USER) |
| SINGLE | |
| FFT | initializes the fast Fourier transform (FFT) function. |
| | Key Access: (MEAS/USER) |

| SINGLE MEAS | sets the functions in the Power Menu so that they make the measurement on a single sweep. |
|-----------------|--|
| | Key Access: (MEAS/USER) |
| SLOPE | draws a straight line between the coordinate point of the current segment and the coordinate point of the next segment, producing limit-line values for all frequencies between the two points. |
| | Key Access: (DISPLAY) |
| SOUND ON OFF | allows you to turn the TV sound on and off. This feature is the same as the TV RECEIVER mode of operation except that it is accessed from a front-panel hardkey. |
| | Press (AUX CTRL), Monitor TV to access SOUND ON OFF. |
| | Key Access: (AUX CTRL) |
| (SPAN) or SPAN | activates the SPAN function and accesses the frequency-span functions. Key Access: (SPAN) |

| SPAN | |
|----------------------|---|
| Z00M | finds the highest signal peak on-screen. If a marker is not already on the peak, it places a marker on it, turns on the marker-track function, and activates the span function. Pressing SPAN ZOOM performs the routine similar to pressing the following keys: (PEAK SEARCH), (MKR FCTN), MK TRACK ON OFF (ON), and (SPAN). |
| | Key Access: (<u>Span</u>) |
| SPEAKER ON OFF | turns the internal speaker on and off. The volume from the speaker is controlled by the -panel volume control knob and FM GAIN (when using FM demodulation). There is no output from the speaker unless demodulation is turned on. Pressing (PRESET) sets SPEAKER ON OFF to ON. |
| | Key Access: (AUX CTRL) |
| SPECTRUM ANALYZER | sets the analyzer to the spectrum analyzer operating mode and accesses the PRESET SPECTRUM softkey function. |
| | Key Access: (MODE) or (PRESET) |
| SQUELCH | adjusts the squelch level. The squelch level mutes weak signals and passes strong signals. The squelch level affects the audio output only. |
| | Key Access: (AUX CTRL) |
| SRC ATN MAN AUTO | <i>Option 011 only.</i> allows you to select between automatic and manual adjustment of the tracking generator's switching attenuator. |
|---------------------|--|
| | Key Access: (AUX CTRL) |
| SRC PWR OFFSET | Option 011 only. offsets the displayed power of the tracking generator (SRC). Offset values may range from -100 dB to $+100 \text{ dB}$. |
| | Key Access: (AUX CTRL) |
| SRC PWR ON OFF | <i>Option 011 only.</i> activates (ON) or deactivates (OFF) the output power of the tracking generator (SRC). |
| | Key Access: (AUX CTRL) |
| SRC PWR STP SIZE | Option 011 only. allows the user to set the step size of the source-power level, source-power offset, and power-sweep range functions. The step size may be values from -32.7 dB to 32.7 dB. The default setting is one vertical scale division. |
| | Key Access: (AUX CTRL) |
| START FREQ | sets the frequency at the left side of the graticule. |
| | |

Key Access: $(\ensuremath{\mathsf{FREQUENCY}})$

START TIME

allows you to specify the time and the date when testing is to begin for timed measurements. The time is specified in 24-hour, hour, minute format; the date is specified in year, month, and day format. If you do not specify a time, the current time is used. If you do not specify a date, the current date is used. If you specify a time earlier than the current time, the current time is used. If you specify a date earlier than the current date, the current date is used. If the date or time setting of the analyzer is incorrect, see "If the wrong date or time is displayed" in Chapter 4 of the *HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide*.

Pressing CABLE TV ANALYZER, SYSTEM MEAS,

Timed Measure accesses START TIME.

Key Access: SYSTEM MEAS

STD

allows you to select the standard (STD) channel tuning configuration. The standard configuration means that the channels are at the frequencies that the FCC defines to be the standard channel frequencies. The standard frequency assignments are defined in the FCC channel identification plan, part 76.612.

If you do not select a channel configuration, the default configuration is the standard (STD) configuration. The choices for the channel tuning configuration are: standard (STD), off-the-air (AIR), incrementally related carriers (IRC), harmonically related carriers (HRC), and T format.

The channel tuning configuration selection is selected independently of the test plan and is not stored with the test plan. The analyzer retains the channel tuning configuration selection, even when power is turned off, until a new selection is made.

Press CABLE TV ANALYZER, Setup, Channel Tuning to access STD.

Key Access: CABLE TV ANALYZER

STOP TIME allows you to specify the time and the date to end the testing for a timed measurement. The time is specified in 24-hour, hour, minute format; the date is specified in year, month, and day format. If you enter a stop time that is 24 hours or greater, the stop time will be set to 23 hours and 59 minutes. Pressing CABLE TV ANALYZER, SYSTEM MEAS, Timed Measure accesses STOP TIME.

Key Access: SYSTEM MEAS

| STATE | |
|----------------------|---|
| \rightarrow CARD | saves the current analyzer state on the memory card. |
| | Key Access: (SAVE) |
| STATE | |
| \rightarrow INTRNL | saves the current analyzer state in the selected state register. |
| | Key Access: (SAVE) |
| STOP | |
| FREQ | sets the frequency at the right side of the graticule. |
| | Key Access: (FREQUENCY) |
| STOR PWR | |
| ON UNITS | sets the default settings for the units used in the linear and the logarithmic modes when the instrument is powered on. The settings for the units can be changed during normal instrument operation but they will return to the default settings when the instrument is powered on again. This is a service calibration function. Refer to the service guide for more information. |
| | Key Access: (CAL) |

| STP GAIN ZERO | disables the two 20 dB step-gain amplifiers on the A12 amplitude control assembly. This is a service diagnostic function and is for service use only. |
|-------------------|--|
| | Key Access: CAL |
| STUVWX | accesses the softkey menu used for selecting screen title or prefix characters S through X. |
| | Key Access: (CAL), (CONFIG), (DISPLAY), (RECALL), or (SAVE) |
| (SWEEP) | accesses the sweep-time menu and accesses the sweep-time softkey functions SWP TIME AUTO MAN and SWEEP CONT SGL . |
| SWEEP CONT SGL | switches the analyzer between the continuous-sweep mode and the single-sweep mode. |
| | Key Access: (SWEEP) or (TRIG) |
| SWEEP DELAY | Option 107 only. delays the start of sweep with respect to the gate trigger edge in the time domain window of the gate utility. The time segment being displayed can be shifted up to 65 msec after the gate trigger edge in 1 μ sec increments. |
| | Key Access: (SWEEP) |

| SWEEP RAMP | displays the RAMP signal from the sweep-ramp generator that is located on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
|----------------------|---|
| | Key Access: (CAL) |
| SWEEP TIME DAC | displays the output of the sweep-time DAC (SWP_DAC) from the sweep-ramp generator that is on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. |
| | Key Access: (CAL) |
| SWP CPLG SR SA | <i>Option 011 only.</i> selects stimulus-response (SR) or spectrum-analyzer (SA) auto-coupled sweep time. |
| | Key Access: (AUX CTRL) |
| SWP TIME AUTO MAN | selects the length of time in which the analyzer sweeps the displayed frequency span. |
| | Key Access: (AUTO COUPLE) or (SWEEP) |
| SYNC CRD TV TRIG | provides sweep triggering on the selected line of a video picture field and accesses the softkey menu used to select the line number of the video picture field and the type of video picture frame. |
| | Key Access: (Trig) |

| SYNC NRM NTSC | changes the rear panel MONITOR output between normal internal monitor horizontal and vertical synchronization constants or the NTSC video compatible format. |
|--------------------|--|
| CYNC | Key Access: (CONFIG) |
| NRM PAL | changes the rear panel MONITOR output between normal internal monitor horizontal and vertical synchronization constants or the PAL video compatible format. |
| | Key Access: (CONFIG) |
| SYS BTS TST CH# | allows you to select a channel to perform system beats testing. The default is channel 4. |
| | If YES is selected after selecting Measure Now, system beats will be measured in the channel specified by SYS BTS TST CH#. The results are displayed on the first page of the measurement data. |
| | System beats testing is selected independently of the test plan and is not stored with the test plan. The analyzer retains the channel selection, even when power is turned off, until a new selection is made. |
| | Pressing CABLE TV ANALYZER, Setup accesses SYS BTS TST CH#. |

Key Access: CABLE TV ANALYZER

SYSTEM FRQ RESP initiates the System Frequency Response measurement. Key Access: CHANNEL MEAS

SYSTEM MEAS

Accesses the system measurement mode. When the analyzer is in the system measurement mode, MONITOR is displayed in the upper right corner of the analyzer display.

Pressing SYSTEM MEAS accesses the following:

ENTER LOCATION, ENTER TEMP, Recall TstPlan, Measure Now, Timed Measure, Create TstPlan, Recall Data, CATALOG CARD, VIEW LAST, and VIEW CH SEL.

Accessing the system measurement mode automatically changes the input impedance setting of a spectrum analyzer to 75 Ω . (The setting of the input impedance is usually controlled by INPUT Z 50 75.) For spectrum analyzer's with 50 Ω impedance, changing the input impedance with INPUT Z 50 75 minimizes the impedance mismatch, but does not eliminate the impedance mismatch unless a 50 Ω to 75 Ω external matching pad is used.

Press (MODE), CABLE TV ANALYZER to access SYSTEM MEAS .

Key Access: SYSTEM MEAS

T

allows you to select the T channel tuning configuration.

If you do not select a channel tuning configuration, the default configuration is the standard (STD) configuration. The choices for the channel tuning configuration are: standard (STD), off-the-air (AIR), incrementally related carriers (IRC), harmonically related carriers (HRC), and T format.

The channel tuning configuration selection is selected independently of the test plan and is not stored with the test plan. The analyzer retains the channel tuning configuration selection, even when power is turned off, until a new selection is made.

Press CABLE TV ANALYZER, Setup, Channel Tuning to access ${\tt T}$.

Key Access: CABLE TV ANALYZER

TAKE REF TRACE allows you to take a reference trace measurement at any point in the system that you want to use as the "reference." For example, typical reference spectrums can be set up at the output of the headend or AML hub. Thirty-two or more reference traces can be stored.

Key Access: CHANNEL MEAS

| Timed Measure | allows you to access the functions that you can use to |
|------------------|---|
| | initiate the testing at a later time. |
| | When you press Timed Measure, you access a |
| | menu with the following softkeys: START TIME, |
| | STOP TIME, TIME INTERVAL, DUMP TO PRINTER, |
| | DUMP TO RAM CARD, and Main Menu. |
| | Pressing CABLE TV ANALYZER SYSTEM MEAS accesses |
| | Timed Measure . |
| | Key Access: SYSTEM MEAS |
| TTMF | |
| INTERVAL | allows you to specify the time interval between the starting of testing for a timed measurement. The time is specified in 24-hour, hour, minute format. The maximum time interval is 99 hours. |
| | If you do not specify a time interval, the time interval is set to zero. When the time interval is set to zero, testing will begin immediately after the previous tests have been completed. |
| | When specifying a time interval, you need to specify a time interval that is at least long enough to allow the analyzer to complete the testing before beginning testing again. Here are some general guidelines for estimating how long it takes the analyzer to test one channel: |
| | • It takes approximately 1 minute to test a channel that is not scrambled with all of the tests selected for that channel. |
| | • It takes approximately 1 minute and 30 seconds to test a scrambled channel with all of the tests selected for that channel. |
| | So the time interval that you specify (in minutes) should, at the least, exceed the number of channels that you are testing plus 10 minutes. |
| | |

Press CABLE TV ANALYZER, SYSTEM MEAS, Timed Measure to access TIME INTERVAL.

Key Access: SYSTEM MEAS

TODAY'S DATE

allows you to select the current date as the start date or stop date for timed measurements.

When you press **TODAY'S DATE**, the analyzer uses the date setting of the spectrum analyzer. If the date or time setting of the analyzer is incorrect, see "If the wrong date or time is displayed" in Chapter 4 of the *HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide*.

Press CABLE TV ANALYZER, SYSTEM MEAS,

Timed Measure, START TIME to access TODAY'S DATE as the start date for the timed measurement. Press Timed Measure, STOP TIME to access TODAY'S DATE as the stop date for the timed measurement.

Key Access: SYSTEM MEAS

| TODAY'S TIMEDATE | allows you to select the current time and date as the start time and date for timed measurements. |
|----------------------|---|
| | When you press TODAY'S TIMEDATE , the analyzer uses the time and date setting of the analyzer. If the date or time setting of the analyzer is incorrect, see "If the wrong date or time is displayed" in Chapter 4 of the <i>HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide</i> . |
| | Press CABLE TV ANALYZER, SYSTEM MEAS, |
| | Timed Measure, START TIME to access |
| | TODAY'S TIMEDATE. |
| | Key Access: SYSTEM MEAS |
| T WINDOW RES BW | <i>Option 107 only.</i> changes the resolution bandwidth in the time domain window of the gate utility. |
| | Key Access: (SWEEP) |
| T WINDOW SWP TIME | <i>Option 107 only.</i> changes the sweep time in the time domain window of the gate utility and re-scales the gate markers. |

Key Access: (SWEEP)

| TABLE ∆DL NRM | switches the marker table between the normal marker formats and the delta display line format. The marker information can be displayed in absolute amplitude and frequency. It can also be displayed in delta amplitude and frequency using one marker as the reference. These are normal marker modes. |
|--------------------|--|
| | Key Access: (MKR FCTN) |
| THRESHLD ON OFF | sets a lower boundary to the active trace. The threshold line "clips" signals that appear below the line when this function is on. The boundary is defined in amplitude units that correspond to its vertical position when compared to the reference level. |
| | Key Access: DISPLAY |
| Time Date | accesses the softkey menu used to set and display the real-time clock. |
| | Key Access: (CONFIG) |
| TIMEDATE ON OFF | turns the display of the real-time clock on or off. Pressing DEFAULT CONFIG sets TIMEDATE ON OFF to ON. |
| | Key Access: (CONFIG) |

| TOI | |
|----------|--|
| ON OFF | finds the third-order intercept of the two highest amplitude signals and the two associated distortion products. |
| | Key Access: (MEAS/USER) |
| (TRACE) | accesses the trace softkeys that allow you to store and manipulate trace information. Each trace is comprised of a series of data points that form a register where amplitude information is stored. The analyzer updates the information for any active trace with each sweep. If two traces are being written to, they are updated on alternating sweeps. |
| TRACE | |
| ABC | selects the softkey menu used for trace A, trace B, or trace C functions. Press TRACE A B C until the letter of the desired trace is underlined. |
| | Key Access: (TRACE) |
| TRACE A | sets up trace A for recalling previously-saved trace data into trace A or saving trace data from trace A. |
| | Key Access: (RECALL) or (SAVE) |
| TDACE D | |
| I RACE D | sets up trace B for recalling previously-saved trace data into trace B or saving trace data from trace B. |
| | Key Access: (RECALL) or (SAVE) |
| TRACE C | sets up trace C for recalling previously-saved trace data into trace C or saving trace data from trace C. |
| | Key Access: (RECALL) or (SAVE) |

| $Trace \rightarrow Card$ | begins the process used to save trace data, limit-line tables or amplitude-correction factors on the memory card. |
|--------------------------|--|
| | Key Access: (SAVE) |
| Trace → Intrnl | accesses a softkey menu that allows you to select the item to be stored in analyzer memory: the trace to be saved (trace A, trace B, or trace C), limit-line tables, or amplitude-correction factors. |
| | Key Access: (SAVE) |
| Track Gen | <i>Option 011 only.</i> displays softkey menus for use with a built-in tracking generator. |
| | Key Access: (AUX CTRL) |
| TRACKING PEAK | <i>Option 011 only.</i> activates a routine that automatically adjusts the tracking adjustment to obtain the peak response of the tracking generator on the analyzer display. |

NOTE

For tracking peak to function properly, the tracking generator must be connected to the analyzer.

Before making a stimulus-response measurement, care must be taken to maximize the tracking adjustment of the tracking generator to ensure maximum available dynamic range.

Key Access: (AUX CTRL)

(TRIG)

accesses softkeys that let you select the sweep mode and trigger mode. (Also see "Screen Annotation" in Chapter 1 of this guide.)

NOTE

With some delayed trigger functions (for example, external or TV triggering), the softkey menu is not updated until after the trigger has occurred.

| TRIG MKR ON OFF | <i>Option 107 only.</i> activates a marker that indicates the time from the gate trigger to the current marker position. The trigger marker is in the time domain window of the gate utility. It is turned off when the menu is exited or if another active function is activated. |
|--------------------|---|
| | Key Access: (SWEEP) |
| TTL PWR © INPUT | allows you to measure the total power at the input of the analyzer. |
| | This function is used in conjunction with an external preamplifier. |
| | By measuring the total input power, you ensure that the preamplifier's specifications are not exceeded, which could contribute to distortion products when measuring CSO/CTB. |
| | Press CABLE TV ANALYZER, Setup, Analyzer Input to access TTL PWR @ INPUT. |
| | Key Access: CABLE TV ANALYZER |
| TV LINE # | selects the line number of the video picture field. The values |

allowed are 1 to 1012. (PRESET) sets the TV line number to 17.

Key Access: (TRIG)

TV RECEIVER

allows you to access the TV receiver mode. In this mode, you can view a TV picture and listen to the sound simultaneously by pressing CHANNEL SELECT , entering a

TV channel number, then pressing **PIX &SND ON OFF**. To exit this mode, press any softkey.

Press (MODE) to access TV RECEIVER .

Front-Panel Access Key: (MODE)

TV

Standard allows the analyzer to trigger on the standard video formats; NTSC, PAL, PAL-M, or SECAM-L.

Key Access: (TRIG)

TV SYNC NEG POS

selects the polarity of the modulation of the video format. NTSC uses the negative or positive modulation video format. NTSC, PAL, and PAL-M use negative modulation. SECAM-L uses positive modulation.

Key Access: (TRIG)

| TV TRIG EVEN FLD | selects an even video field of an interlaced video format to trigger on. |
|---------------------|--|
| | Key Access: (TRIG) |
| TV TRIG ODD FLD | selects an odd video field of an interlaced video format to trigger on. |
| | Key Access: (TRIG) |
| TV TRIG VERT INT | selects a vertical interval to trigger on. Triggering occurs on the next pulse edge. If it triggers on an even field, triggering will not alternate between odd and even fields. If it triggers on an odd field, triggering will alternate between odd and even fields. The vertical interval is used for non-interlaced video formats. |
| | Key Access: (TRIG) |
| UPDATE TIMEFREQ | Option 107 only. switches between the time domain and frequency domain windows in the gate utility. The window is updated to reflect the current state when it is activated. The WINDOWS (NEXT) key will also switch between the windows. |
| | Key Access: (SWEEP) |
| User Menus | accesses a menu available for your use for user-defined programs and key functions. |
| | Key Access: (MEAS/USER) |

| USR CH O CNTR FRQ | allows you to define the center frequency for a user defined channel. This channel is then accessed by selecting channel 0 (zero). The setting defined here is stored in nonvolatile analyzer memory. This means that the analyzer retains the setting, even when power is turned off, until a new setting is made. | | |
|----------------------|--|--|--|
| | Press CABLE TV ANALYZER, Setup, Channel Tuning to access CHNL O CNTR FRQ. | | |
| | Rey Access. GADLE IV ANALIZER | | |
| USR CH O SPAN | allows you to define the span for a user defined channel. This channel is then accessed by selecting channel 0 (zero). The setting defined here is stored in nonvolatile analyzer memory. This means that the analyzer retains the setting, even when power is turned off, until a new setting is made. | | |
| | Press CABLE TV ANALYZER, Setup, Channel Tuning to | | |

access CHNL O SPAN.

Key Access: CABLE TV ANALYZER

| V | changes the amplitude units to V for the current setting (log or linear). |
|-------------------|---|
| | Key Access: (AMPLITUDE) |
| VBW/RBW RATIO | selects the ratio between the video and resolution bandwidths. If signal responses near the noise level are visually masked by the noise, the ratio can be set to less than 1 to smooth this noise. The knob and step keys change the ratio in a 1, 3, 10 sequence. Pressing (PRESET) and AUTO ALL sets the ratio to 0.300 X. The ratio can be changed using the step keys or the knob. |
| | Key Access: (BW) |
| VID AVG ON OFF | initiates a digital averaging routine that averages displayed signals and noise. This function does not affect the sweep time, bandwidth, or other analog characteristics of the analyzer. |
| | Key Access: (BW), (TRACE), or (MEAS/USER) |

VID BW AUTO MAN

changes the analyzer post-detection filter from 30 Hz to 3 MHz in a 1, 3, 10 sequence. *For Option 130 only.* Option 130 provides additional narrow resolution bandwidths of 30 Hz, 100 Hz, and 300 Hz.

NOTE

Coupling the video bandwidth function also couples the video bandwidth to resolution bandwidth ratio function. If you want to auto-couple the video bandwidth to a nonstandard ratio, you must set the video bandwidth to auto-couple before setting the video-bandwidth/resolution-bandwidth ratio.

| Kev | Access: | AUTO COUPLE | or | (BW) |
|-----|---------|-------------|----|------|
| ncy | necess. | AUTO COUFLE | 01 | |

VIDEO

activates the trigger condition that allows the next sweep to start if the detected RF envelope voltage rises to a level set by the display line.

Key Access: (TRIG)

VIEW A

holds and displays the amplitude data that is in the trace A register. The trace A register is not updated as the analyzer sweeps. If trace A is deactivated by pressing STORE BLANK A, the stored data can be retrieved by pressing VIEW A.

Key Access: (TRACE)

VIEW B holds and displays the amplitude data that is in the trace B register. The trace B register is not updated as the analyzer sweeps. If trace B is deactivated by pressing STORE BLANK B, the stored data can be retrieved by pressing VIEW B. Key Access: (TRACE) VIEW C holds and displays the amplitude data that is in the trace C register. The trace C register is not updated as the analyzer sweeps. If trace C is deactivated by pressing STORE BLANK C, the stored data can be retrieved by pressing VIEW C. Key Access: (TRACE) VIEW CH SEL allows you to view the selected channels of the current test plan. Key Access: SYSTEM MEAS allows the user to measure the relative power of any VIEW co-channel signal that is present. CO-CHNL Key Access: CHANNEL MEAS VIEW INGRESS initiates the View Ingress measurement. Key Access: CHANNEL MEAS

| VIEW LAST | allows you to view the last test results on the analyzer screen. Upon being sent to the screen, the following options are available: |
|----------------|---|
| | • NEXT PAGE displays the next page of test results. |
| | • PREV PAGE displays the previous page of test results. |
| | • GOTO PAGE displays selected page number. |
| | • PRINT REPORT prints a test results report. |
| | • STORE TO CARD stores test results to a RAM card. |
| | Press CABLE TV ANALYZER, SYSTEM MEAS, |
| | More 1 of 2 to access VIEW LAST. |
| | Key Access: SYSTEM MEAS |
| Volts | changes the amplitude units to volts. |
| | Key Access: (AMPLITUDE) |
| Watts | changes the amplitude units to watts. |
| | Key Access: (AMPLITUDE) |
| WINDOWS OFF | turns off the windows display mode and returns to the normal full-sized instrument display. The state of the last active window will become the instrument state when the windows display is turned off. |
| | Key Access: WINDOWS (ON) |

WINDOWS (ON)

activates the windows display mode and accesses the menu of window zone functions. The windows display function splits the screen into two separate displays. Only one of these displays is active at a time. The currently active window will have a solid line around the graticule rather than a broken line.

Most functions can be used from within the windows display mode. Some functions, like editing limit lines and showing the options, require a full-sized display. They will temporarily exit the windows display format. When the function is finished the instrument will return to a windows display. Other functions will permanently exit the windows display and it will be necessary to restart the windows display mode by pressing WINDOWS (ON). See Table 4-8.

This key can also be used to toggle the TV picture and sound On and Off when in the channel measurement mode and after a channel has been selected. Pressing any softkey returns you to the analyzer display.

| Function | Description | |
|------------------------|---|--|
| % AM | measures percent amplitude modulation | |
| Adjacent Channel Power | measures adjacent channel power | |
| Calibration Functions | self-calibration routines | |
| Channel Power | measures channel power | |
| Confidence Test | built-in self test routine | |
| Dispose User Memory | deletes user's items from analyzer memory | |
| FFT | initiates FFT on zero span input | |
| Gate Utility | accesses time gate functionality | |
| Instrument Preset | returns analyzer to preset state | |
| Marker Table | lists all of the active markers | |
| N dB Points | measures N dB bandwidth | |
| Occupied Bandwidth | measures occupied bandwidth | |
| Peak Table | lists displayed signal peaks | |
| Peak Zoom | initiates the peak zoom routine | |
| тоі | makes a third-order intercept measurement | |

Table 4-8. Functions That Exit the Windows Display Format

X FINE TUNE DAC displays the output of the YTO extra-fine-tune DAC (FM_TUNE) that is on the A7 analog interface assembly. This is a service diagnostic function and is for service use only. Key Access: (CAL) allows you to confirm that you want to perform the function. If you press CLEAR TST PLAN and then press YES, you have selected to clear the current test plan from analyzer memory. If you enter a channel by using ENTER BY CHANNEL or ENTER BY RANGE and then press YES, the channel or range of channels will be treated as scrambled channels. If you press Measure Now and then press YES. CTB testing will be performed.

YES

YES is accessed by pressing CLEAR TST PLAN, or entering a channel number with ENTER BY CHANNEL or ENTER BY RANGE, or by pressing Measure Now.

Key Access: SYSTEM MEAS

| YZ _# Spc Clear | accesses the softkey menu used for selecting the characters Y, Z, underscore (_), #, space, or for clearing the screen title. |
|---------------------------|---|
| | Key Access: (CAL), (CONFIG), (DISPLAY), (RECALL), or (SAVE) |
| ZERO MARKER | zeros the value of the delta marker to establish a new reference. The ZERO MARKER softkey is in the pulse parameter entry menus of the gate utility. |
| | Key Access: SWEEP |
| ZERO SPAN | changes the frequency span to zero and turns off marker track if it is on. Key Access: (SPAN) |

| ZONE CENTER | allows the zone markers to be moved in frequency without changing the zone span. |
|------------------|--|
| | Key Access: WINDOWS (ON) |
| ZONE PK LEFT | searches for the next frequency peak outside and to the left of the zone markers on the upper window trace and then moves the zone so that it is centered around the new peak. |
| | Key Access: WINDOWS (ON) |
| ZONE PK RIGHT | searches for the next frequency peak outside and to the right of the zone markers on the upper window trace and then moves the zone so that it is centered around the new peak. |
| | Key Access: WINDOWS ON |
| ZONE SPAN | allows the span of the zone markers to be changed without changing the center frequency. |
| | Key Access: WINDOWS (ON) |
| (<u>zoom</u>) | switches between the split-screen windows display and a full size display of the window that is currently active. Once the ZOOM function is active the (NEXT) key can be used to switch between the two windows while remaining zoomed (full sized). |

5

Key Menus

Key Menus

This chapter contains the key menu diagrams for the HP 8591C Cable TV Analyzer.

Each key menu diagram is arranged alphabetically according to the front-panel key name.

NOTE

The softkeys in the key menus that specific to the HP 85721A Cable TV RF/Video Measurements Personality are described in the HP 8591C Cable TV Analyzer Cable TV Measurements User's Guide.

Key Menus Cable TV Analyzer Functions













Key Menus Cable TV Analyzer Functions







Quick Reference

5-6

Key Menus Cable TV Analyzer Functions



| FREQUENCY | |
|---|-------|
| CENTER FREQ START FREQ STOP FREQ CF STEP AUTO MAN FREQ OFFSET | |
| | pu16c |





5-8

Key Menus Cable TV Analyzer Functions





Quick Reference






















| PU128 |
|-----------|
| SWP PU128 |







5 - 20





Index

- 10 MHz reference DAC setting, 4-19
 10 MHz REF OUTPUT, 3-7
 +10V REF DETECTOR, 4-4, 4-6
 -10V REF DETECTOR, 4-4, 4-6
 120 kHz EMI BW, 4-7
- 2 200 Hz EMI BW, 4-7 2v REF DETECTOR, 4-4, 4-6
- $3\,$ dB bandwidth measurement, 4-77
- 9 9 kHz EMI BW, 4-7

A A <-> B, 4-7 $A-B \rightarrow A \text{ ON OFF}, 4-7$ $A \rightarrow C, 4-7$ ABCDEF, 4-8 ABORT, 4-8 accessories shipped with the cable TV analyzer, 2-5 ACPGRAPH , 4-8 active function, 3-15 active function block, 3-4 active function clearing, 4-60addressing printer, 4-94 address of analyzer, 4-13 address of plotter, 4-85 adjacent channel power, 4-8 extended dynamic range, 4-9 adjacent channel power graph, 4-8 adjacent channel power ratio, 4-8 ADJ CHAN POWER, 4-8 ADJ CHAN PWR extd, 4-9 AIR, 4-9 ALC MTR INT XTAL, 4-9 ALC TEST, 4-4, 4-10 aliasing signal identification, 4-110 ALL DLP \rightarrow CARD, 4-10 ALL TESTS, 4-10 alpha characters, 4-8, 4-60, 4-117, 4-137

% AM ON OFF, 4-6 Amp Cor, 4-10 AMP COR, 4-10 AMP COR ON OFF, 4-11 amplifier, select external, 4-52 AMPLITUDE, 3-4, 3-17, 4-11 amplitude accuracy, 4-33 amplitude and frequency self-calibration routine, 4-19 amplitude correction factors cataloging, 4-21delete point, 4-38 edit, 4-43 edit done, 4-44 menus, 4-10 on or off, 4-11 purge, 4-96 recall table, 4-97 save table, 4-102 select amplitude, 4-105 select frequency, 4-106 select point, 4-106 select sweep time, 4-106 storing, 4-10 amplitude demodulation, 4-39 amplitude menus, 4-11 amplitude modulation, 4-6 amplitude scale, 4-104amplitude self-calibration, 4-18 amplitude units, 4-11 dBm, 4-36 dBmV, 4-36 $\mathrm{dBuV},\ 4\text{-}36$ Volts, 4-135 Watts, 4-135 Amptd Units, 4-11 analog+ display, 4-11 ANALOG + ON OFF, 4-11 ANALYZER ADDRESS, 4-13 analyzer attenuator, 4-14 analyzer auto-coupled sweep time, 4-118 analyzer battery, 3-32 analyzer configuration default, 4-36 ANALYZER GAINS, 4-4, 4-13 Analyzer Input, 4-13 analyzer memory cataloging functions, 4-22 catalog variables, 4-24 delete file, 4-38

DLP cataloging, 4-22 erase user programs and variables, 4-41 program and variable cataloging, 4-21 saving to, 4-127 state and trace register status, 4-24analyzer mode, 4-92 analyzer state recalled, 4-88 annotation, 3-11 ANNOTATN ON OFF, 4-14 APND CAT ITEM, 4-14 ATTEN AUTO MAN, 4-14 attenuation coupling, 4-15 attenuator-error factors, 4-108 audio volume, 4-65 aural carrier, 4-25 AUTO ALL, 4-15 AUTO COUPLE, 4-15 automatic FFT, 4-69 automatic leveling control, 4-9 AUXB, 4-4, 4-16 Aux Conn Control, 4-15 AUX CTRL, 4-16 AUX IF OUTPUT, 3-7 auxiliary connector input, 4-41 auxiliary interface connector, 4-16 control line A, 4-28 control line B, 4-28 control line C, 4-28 control line D, 4-28 AUX INTERFACE, 3-7 AUX VIDEO OUTPUT, 3-7 $\mathbf{B} \ \mathbf{B} \to \mathbf{C}, \ 4\text{-}16$

B <-> C, 4.16
B-DL → B, 4.16
bandwidth functions, 4.18
bandwidth measurement, 4.77
bandwidth ratio
video bandwidth to resolution bandwidth, 4.132
battery
changing the battery, 3.29
BAUD RATE, 4.16
BINARY SPAN, 4.4, 4.17
black and white printing, 4.61
BLANK A, 4.17
BLANK B, 4.17
BLANK C, 4.17

BLANK CARD, 4-17 built-in tracking generator, 4-127 BW, 4-18

C CAL, 3-21, 4-18 CAL AMPTD, 2-14, 3-23, 4-18 CAL FETCH, 3-22, 4-19 CAL FREQ, 2-14, 3-23, 4-19 CAL FREQ & AMPTD, 2-14, 3-22, 4-19 calibrating the analyzer, 3-22 calibration, 4-18attenuator-error factors, 4-108 self-calibration routines, 3-21 CAL OUT, 3-4 CAL STORE, 2-15, 3-22, 4-19 CAL TIMEBASE, 4-3, 4-19 CAL TRK GEN, 2-16, 3-24, 4-20 card changing the battery, 3-29 inserting a memory card, 3-28 Card Config, 4-20 CARD \rightarrow DISPLAY, 4-20 CARD \rightarrow DLP, 4-20 CARD \rightarrow STATE, 4-21 Card \rightarrow Trace, 4-21 carrier level, 4-25 CARRIER LEVEL&&FRQ, 4-21 CARRIER/NOISE, 4-21 CATALOG ALL, 4-21 CATALOG AMP COR, 4-21 Catalog Card, 4-22 CATALOG CARD , 4-22 CATALOG DISPLAY, 4-22 CATALOG DLP, 4-22 cataloging the RAM or ROM card, $4\mathchar`-22$ Catalog Internal, 4-22 CATALOG LMT LINE, 4-22 CATALOG ON EVENT, 4-23 CATALOG PREFIX, 4-23 CATALOG REGISTER, 4-24 CATALOG STATES, 4-24 CATALOG TRACES, 4-24 CATALOG VARIABLS, 4-24 CENTER FREQ, 4-24 center frequency, 4-24 window zone, 4-138 center frequency step size, 4.24center frq, user defined channel, 4-131

CF STEP AUTO MAN, 4-24 change function values, 3-6, 3-8 change prefix edit done, 4-44 Change Prefix, 4-24 Change Title, 4-25 changing windows, 4-77 CHANNEL BANDWDTH, 4-25 channel format HRC, 4-25 IRC, 4-25 over the air, 4-25 standard, 4-25 channel number range, 4-46 CHANNEL POWER, 4-25 CHANNEL SELECT, 4-25 channel selection, view, 4-134 CHANNEL SPACING, 4-25Channel Tuning, 4-26 characters, 4-8, 4-60, 4-117, 4-137 check analyzer operation, 4-29 checking the fuse, 2-8 CISPR testing, 4-46 CLDI, 4-26 Clear, 4-27 clear display active function area, 3-6, 3-8 clearing the active function, 4-60 clearing the test plan, 4-27 CLEAR PARAM, 4-27 CLEAR TESTS , $4\mathchar`-27$ CLEAR TST PLAN , $4\mathchar`-27$ CLEAR WRITE A, 4-27 CLEAR WRITE B, 4-27 CLEAR WRITE C, 4-28 CNTL A 0 1, 4-28 CNTL B 0 1, 4-28 CNTL C 0 1, 4-28 CNTL D 0 1, 4-28 CNT RES AUTO MAN, 4-28 COARSE TUNE DAC, 4-4, 4-29 commands dispose all , 4-48, 4-49, 4-50 front panel execution, 4-51 common preset conditions, 4-90 confidence test, 4-29 CONFIG, 4-29 configuration plotter, 4-85, 4-86

printer, 4-94, 4-95 CONF TEST, 4-29 connector 10 MHz ref output, 3-7 AUXB, 4-16 aux IF output, 3-7 aux video output, 3-7 cal output, 3-4 earphone, 3-7 external keyboard, 3-7 ext ref in, 3-7 ext trig input, 3-7 gate output, 3-7 gate trigger input, 3-7 hi sweep in/out, 3-7 HP-IB interface, 3-7 monitor output, 3-7 sweep output, 3-7 TV trig output, 3-7 connector, auxiliary interface, 4-16 continuous sweep, 4-117 CONTINUS FFT, 4-29 CONT MEAS, 4-29 control functions, 3-4 control line A auxiliary interface, 4-28 control line B auxiliary interface, 4-28 control line C auxiliary interface, 4-28 control line D auxiliary interface, 4-28 controlling the gate GATE CTL EDGE LVL, 4-58 GATE ON OFF, 4-58 COPY, 3-4, 4-30 COPY DEV PRNT PLT, 4-33 correction factors, 3-21, 4-33 data display, 4-40 default, 4-36 self-calibration, 4-19 storing, 4-19 CORRECT ON OFF, 4-33 coupled functions menu, 4-15 coupled sweep time, 4-118 coupling all functions, 4-15 time gate utility, 4-27, 4-33, 4-34, 4-37 coupling video bandwidth, 4-133

CPL RBW ON OFF, 4-33 CPL SWP ON OFF, 4-34 CPL VBW ON OFF, 4-34 creating new programs, 4-77 CRT HORZ POSITION, 4-34 CRT VERT POSITION, 4-35

D DAC

extra fine-tune, 4-137 fine tune, 4-55 sweep time, 4-118 DACS, 4-4, 4-35 data controls, 3-6, 3-8 data keys, 3-4, 3-9 data recall, 4-97 DATA TO\PRINTER , 4-35 data transmission speed, 4-16 DATEMODE MDY DMY, 4-35 dBm, 4-36 dBmV, 4-36 dBuV, 4-36 deactivate function, 3-6, 3-8 deactivating the active function, 4-60 DEFAULT CAL DATA, 4-36 DEFAULT CONFIG, 4-36 default configuration, 4-109 DEFAULT SYNC, 4-37 Define Coupling, 4-37 Define Gate, 4-38 Define Time, 4-38 delay GATE DELAY, 4-58 delaying sweep, time gate utility, 4-117 DELETE FILE, 4-38 delete files from memory card, 4-17 DELETE POINT, 4-38 DELETE SEGMENT, 4-38 deleting the test plan, 4-27 delta display-line markers, 4-125 delta marker, 4-68 Demod, 4-39 DEMOD AM FM, 4-39 DEMOD ON OFF, 4-39 demodulation, 4-39 AM, 4-39 dwell time, 4-43 FM, 4-39 FM gain, 4-56

FM offset, 4-56FM span, 4-56 on and off, 4-39 speaker on off, 4-112 squelch, 4-112 DEPTH MOD, 4-39 deskjet, 4-39 DESKJET 310, 4-39 DESKJET 540, 4-39 DESKJET 550C, 4-39 deskjet printer, 4-39 detection mode negative peak, 4-40 positive peak, 4-40 sample, 4-40DETECTOR PK SP NG, 4-40 DIF GAIN/DIF PHAZ, 4-40 disable peak detector reset, 4-42 display analog+, 4-11cataloging, 4-22 two windows, 4-136 DISPLAY, 4-40 DISPLAY CAL DATA, 4-4, 4-40 DISPLAY CNTL I, 4-41 display compression due to monitor output format, 3-14 display line, 4-42 DISPLAY \rightarrow CARD, 4-40 display zone change span, 4-138 find left peak, 4-138 find right peak, 4-138 move center, 4-138 Dispose User Mem, 4-41 DJ540, 4-61 DL DELTA ON OFF, 4-41 DLP cataloging, 4-22 editor, 4-46 DLP editor function appending, 4-14 display and edit item, 4-44 edit last item, 4-45 DO FRQ RESP TST , 4-41 DONE, 4-42 DONE, 4-42 double display, 4-136 DROOP, 4-4, 4-42

DSP LINE ON OFF, 4-42 DUMP TO PRINTER, 4-42 DUMP TO RAM CARD, 4-42 DUMP TO SCREEN, 4-43 DWELL TIME, 4-43

 ${f E}$ earphone connector, 3-7 EDGE POL POS NEG, 4-43 edge triggering time gate utility, 4-43 edge triggering, gate control, 4-58 Edit Amp Cor, 4-43 EDIT CAT ITEM, 4-44 Edit Done, 4-44 EDIT DONE, 4-44 EDIT FLATNESS, 4-3, 4-44 editing DLPs, 4-46 EDIT LAST, 4-45 Edit Limit, 4-45 Edit Lower, 4-45 Edit Mid/Delt, 4-45 Editor, 4-46 Edit Up/Low, 4-45 Edit Upper, 4-45 EDIT UPR LWR, 4-46 EMI bandwidth 120 kHz, 4-7 200 Hz, 4-7 9 kHz, 4-7 EMI BW Menu, 4-46 ending a measurement, 4-8 ENTER, 4-46 ENTER BY CHANNEL , $4\mathchar`-46$ ENTER BY RANGE, 4-46 entering a channel, 4-46 entering a range of channels, 4-46 entering the location, 4-47 entering the start time, 4-114 entering the stop time, 4-115 entering the temperature, 4-47 entering the time interval for timed testing, 4-122 ENTER LOCATION , 4-47 ENTER PRI, 4-48 ENTER REF EDGE, 4-48 ENTER TEMP, 4-47 ENTER WIDTH, 4-48 epson format, 4-50, 4-51

EPSON LARGE, 4-108 epson lq570, 4-50 epson mx80 , 4-51 epson printer, 4-50, 4-51 EPSON SMALL, 4-108 ERASE ALL, 4-49 ERASE CARD, 4-49 ERASE DLP, 4-48 ERASE DLP MEM, 4-48 ERASE MEM ALL, 4-49ERASE MEM CARD, 4-49 ERASE STATE, 4-50 ERASE STATEALL, 4-50 ERASE TRACE, 4-50 ERASE TRACEALL, 4-50 erase user programs and variables, 4-41 ESD reducing damage caused by ESD, 2-3 exchange trace A and B, 4-7 exchange trace B and C, 4-16 execute a command front panel, 4-51 EXECUTE TITLE, 4-3, 4-51 EXIT, 4-51 Exit Catalog, 4-51 EXIT SHOW, 4-51 EXIT UTILITY, 4-51 expanding a window, 4-138 EXT AMP GAIN, 4-52 EXT AMP NZ FIG, 4-52 EXT AMP YES NO, 4-52EXTERNAL, 4-53 external keyboard connector, 3-7 external pad, 4-53 EXTERNAL PREAMPG, 4-54 EXT KEYBOARD, 3-7 EXT PAD YES NO, 4-53 EXT REF INPUT, 3-7 EXT TRIG INPUT, 3-7

F fast Fourier transform stop frequency, 4-55 fast Fourier transform function, 4-54 features front panel, 3-4 FFT markers, 4-70 FFT compatibility, 4-54

FFT marker frequency, 4-70 FFT MARKERS, 4-54 FFT measurement, 4-29, 4-110 FFT Menu, 4-54 FFT OFF, 4-54 FFT STOP FREQ, 4-55 fine-focus control, 3-10 FINE TUNE DAC, 4-4, 4-55 firmware date, 2-13 fixed, limit line type, 4-64 FLAT, 4-55 flatness factory defaults, 4-62 flatness correction constants, 4-44, 4-55Flatness Data, 4-3, 4-55 FM COIL DRIVE, 4-4, 4-56 FM DEV, 4-56 FM GAIN, 4-4, 4-56 FM OFFSET, 4-5, 4-56 FM SPAN, 4-5, 4-56FORMAT CARD, 4-56 format date, 4-35 Fourier transform See also fast Fourier transform, 4-54 Fourier transform measurement, 4-54 FREE RUN, 4-57 FREQ DIAG, 4-5, 4-57 FREQ DISC NORM OFF, 4-57 FREQ OFFSET, 4-57 frequency stop, 4-116 FREQUENCY, 3-4, 3-17, 4-57 frequency and amplitude self-calibration routine, 2-14, 4-19 frequency demodulation, 4-39 frequency offset, 4-57 frequency self-calibration, 4-19 frequency span, 4-111 front panel execution of programming commands, 4-51 front-panel features, 3-4-5 FRQ DISC NORM OFF, 4-5 function coupling, 4-15 function operations demodulate the aural carrier, 4-65 functions listen off, 4-65 listen on, 4-65 fuse, 2-8 fuse holder, 3-7

G gain

internal analyzer, 4-13 gain, amplifier, 4-52 gate coupling functions, 4-37 gate control, 4-38 edge or level triggering, 4-58Gate Control, 4-58 GATE CTL EDGE LVL, 4-58 gate definition, 4-38 GATE DELAY, 4-58 GATE LENGTH, 4-58 GATE ON OFF, 4-58 GATE OUTPUT, 3-7 GATE TRIGGER INPUT, 3-7 gate utility exit, 4-51 exiting, 4-59 GATE UTILITY, 4-59 gate utility coupling, 4-37 resolution bandwidth, 4-33sweep time, 4-34 video bandwidth, 4-34 GHIJKL, 4-60 GND REF DETECTOR, 4-5, 4-60 graph markers, 4-60 GRAT ON OFF, 4-60 GRPH MKR ON OFF, 4-60 guide conventions, 1-5

H hardkeys

(ENTER), 4-25 (FREQUENCY), 4-121 (MODE), 4-25, 4-121 harmonically related carrier tune configuration selecting, 4-61 highest peak finding next, 4-78 highest signal peak, 4-112 HIGH SWEEP IN/OUT, 3-7 HOLD, 3-6, 3-8, 4-60 horizontal signal positioning, 4-34 HP 8591C cable TV analyzer key menus, 5-3-21 $\,$ HP B&W PRINTER, 4-108 HP-IB address, 4-13 HP-IB interface connector, 3-7 HP Paintjet, 4-80 HRC , 4-61

 $\mathrm{HUM},\;4\text{-}62$

I identify the signal FFT, 4-110 IDNUM, 4-3, 4-62 impedance matching, 4-53 incrementally related carrier tune configuration selecting, 4-63 INIT FLT, 4-3, 4-62 initial inspection, 2-4 INPUT 75 Ω , 3-4 input and output auxiliary control, 4-15 input attenuation control, 4-14 input configuration, 4-13 input impedance, 4-62 INPUT Z 50 Ω 75 Ω , 4-62 inserting a memory card, 3-28 instrument preset, 4-89 instrument state, 3-4 intensity control, 3-4 interface connectors, 3-7 INTERNAL CARD, 4-62 INTERNAL \rightarrow STATE, 4-63 Internal \rightarrow Trace, 4-63

K key

guide conventions, 1-5 key menus, 5-2 knob, 3-4, 3-6, 3-8

L label, softkey, 3-4 laserjet plots, 4-87 LAST SPAN, 4-64 length GATE LENGTH, 4-58 level triggering, gate control, 4-58 limit lines cataloging, 4-22 delete segment, 4-38 displaying, 4-65 edit done, 4-44 edit lower table, 4-45 edit mid/delta amplitude, 4-45 edit table, 4-45 edit upper and lower tables, 4-45, 4-46 edit upper table, 4-45

fixed and relative, 4-64 flat type, 4-55 point, 4-87 purge, 4-96 recall table, 4-100 save table, 4-102 select amplitude, 4-105 select delta amplitude, 4-105 select frequency, 4-106 select frequency or time, 4-65select lower amplitude, 4-106 select middle amplitude, 4-106 select segment, 4-106 select time, 4-106 select type, 4-107 select upper amplitude, 4-107 slope type, 4-111 testing, 4-66 Limit Lines, 4-64 LIMIT LINES, 4-64 LIMITS FIX REL, 4-64 LIMITS FRQ TIME, 4-65 LINE, 4-65 linear scale, 4-104 LINE front-panel, 3-4 LINE switch, 3-4 line voltage selector switch, 2-6 LISTEN OFF, 4-65 LISTEN ON, 4-65

list of markers, 4-75 list of peaks, 4-83 LMT DISP Y N AUTO, 4-65 LMT TEST ON OFF, 4-66 LOAD FILE, 4-66 (LOCAL), 4-29 (LOCAL) hardkey, 4-66 local operation, 4-29, 4-66 location code range, 4-47 lock internal state and trace registers, 4-101 log scale, 4-104 lower limit line editing, 4-45 ${f M}$ MAIN COIL DR, 4-5, 4-67 Main Menu time gate utility, 4-67 MAIN SPAN, 4-5, 4-67 making a measurement, 3-17-19 MAN TRK ADJUST, 4-67 manual tracking adjustment, 4-67 marker delta display-line, 4-125 FFT, 4-54 gate utility trigger, 4-128 list of, 4-75 menu of controls, 4-76 menu of functions, 4-76 selection, 4-105 table, 4-75time gate utility, 4-72 trace selection, 4-75 zeroing delta marker in the gate utility, 4-137 MARKER Δ , 4-68 MARKER $\Delta {\rightarrow} \mathrm{SPAN}, \; 4{\text{-}}68$ MARKER ALL OFF, 4-71 MARKER AMPTD, 4-71 marker counter, 4-74 resolution, 4-28 marker functions, 3-4 marker noise, 4-74 MARKER NORMAL, 3-19, 4-71 MARKER ON, 4-72 marker pause, 4-74 dwell time, 4-43 marker readout modes, 4-74 MARKER \rightarrow AUTO FFT, 4-69 $\mathrm{MARKER} \rightarrow \mathrm{CF}, \ 4\text{-}69$ MARKER \rightarrow CF STEP, 4-69 MARKER \rightarrow FFT STOP, 4-70 MARKER \rightarrow MID SCRN, 4-70 MARKER \rightarrow MINIMUM, 4-70 MARKER \rightarrow PK-PK, 4-70 $\mathrm{MARKER} \rightarrow \mathrm{REF} \ \mathrm{IVL}, \ 4\text{-}71$ MARKER \rightarrow START, 4-71 MARKER \rightarrow STOP, 4-71 marker table mode, 4-125 marker to, 4-76 marker tracking, 4-75, 4-82 mass storage device analyzer memory, 4-62 memory card, 4-62

matching impedance, 4-53 MAX HOLD A, 4-72 MAX HOLD B, 4-72 maximum mixer level, 4-72 MAX MXR LVL, 4-72 MEAS OFF, 4-72 Measure Now, 4-73 MEAS/USER, 4-72 MEM LOCKED, 4-73 memory card battery, 3-29 card configuration, 4-20 catalog display images, 4-22 catalog DLPs, 4-22 catalog functions, 4-22 cataloging amplitude correction factor files, 4-21 catalog limit lines, 4-22 catalog states, 4-24 catalog traces, 4-24 changing the battery, 3-29 delete file, 4-17, 4-38 display recall, 4-20 formatting, 4-56 inserting a memory card, 3-28 program and variable cataloging, 4-21 recalling a DLP, 4-20 recalling a state, 4-21 recalling a trace, 4-21 saving states, 4-116 saving to memory card, 4-127 write-protect switch, 3-30 memory card reader, 3-4 menu and softkey overview, 3-15 menus, 5-2 message block, 3-4 messages **CHANNEL** ?, 4-25 **RECALL TRACE n**, 4-99 **REGISTER #**, 4-99 mid/delta limit line editing amplitude, 4-45 MIN HOLD C, 4-73 MIN HOLD OFF, 4-73 MIN HOLD ON, 4-73 minimum to marker, 4-70 mixer input level maximum, 4-72

MK COUNT ON OFF, 4-74 MK NOISE ON OFF, 4-74

MK PAUSE ON OFF, 4-74 MKR. 3-19, 4-76 MK READ F T I P, 4-74 MKR FCTN, 4-76 $MKR \rightarrow$, 4-76 MK TABLE ON OFF, 4-75 MK TRACE AUTO ABC, 4-75 MK TRACK ON OFF, 4-75 MNOPQR, 4-76 MODE, 3-4, 4-76 model specific preset conditions, 4-90 MONITOR OUTPUT, 3-7 display compression, 3-14 Monitor TV, 4-77 move trace A into C, 4-7 multipen plotter, 4-32 N dB PTS ON OFF, 4-77 negative peak detection mode, 4-40 NEW EDIT, 4-77 NEXT, 4-77 NEXT PAGE, 4-78 NEXT PEAK, 4-78 NEXT PK LEFT, 4-78 NEXT PK RIGHT, 4-78 NO, 4-79 noise figure, amplifier, 4-52 normal marker, 4-71 NORMLIZE ON OFF, 4-79 NORMLIZE POSITION, 4-79 No User Menu, 4-79 NTSC, 4-79 NTSC video format, 4-119 number/units keypad, 3-9 O OCC BW % POWER, 4-80 OCCUPIED BANDWDTH, 4-80 off-the-air tune configuration selecting, 4-9 ON, 4-136 ONCYCLE command, 4-23 ONDELAY command, 4-23 ONEOS command, 4-23 on event commands ONCYCLE, 4-23 ONDELAY, 4-23 ONEOS, 4-23 ONMKR, 4-23

ONPWRUP, 4-23 ONSRQ, 4-23 ONSWP, 4-23 ONTIME, 4-23 TRMATH, 4-23 ONMKR command, 4-23 on/off switch, 3-4 ONPWRUP command, 4-23 on-screen signal positioning horizontal, 4-34 vertical, 4-35 ONSRQ command, 4-23 ONSWP command, 4-23 ONTIME command, 4-23 operation local, 4-66 options displayed, 4-110 output power, 4-113 output screen data, 4-30 overview, menus and softkeys, 3-15 P paintjet printer, 4-108 PAINTJET PRINTER, 4-80 PAL, 4-80 PAL-M, 4-80 PAL video format, 4-119 PARAM AUTO MAN, 4-81 pass code, 4-36 peaknext, 4-78 next left, 4-78 next right, 4-78 peak detection mode positive/negative, 4-40

positive/negative, 4-40 PEAK EXCURSN, 4-81 Peak Menu, 4-81 peak response routine, 4-127 peak search outside of the zone, 4-138 PEAK SEARCH, 3-19, 4-81 peak table, 4-83 mode, 4-83 sorting, 4-83 peak to peak measurement, 4-70 PEAK ZOOM, 4-82 percent amplitude modulation, 4-6 phase lock on/off, 4-6 φ LOCK ON OFF, 4-4, 4-6

picture and sound, 4-82 picture on off, 4-82 PIX &SND ON OFF, 4-82 PK MODE <>DL NRM, 4-83 PK SORT FRQ AMP, 4-83 PK TABLE ON OFF, 4-83 PLAN 1, 4-84 PLAN 2, 4-84 PLAN 3, 4-84 PLAN 4, 4-85 PLAN 5, 4-85 Plot Config, 4-85 plot menu, 4-86 PLOTTER ADDRESS, 4-85 plotter configuration, 4-29 plotter output, 4-86 plotting, 4-30, 4-33 plotting to laserjets, 4-87 PLT _ _ LOC _ _, 4-86 PLT MENU ON OFF, 4-86 Plt Port Config, 4-86 PLT PORT HPIB PAR, 4-86 PLT PORT SER PAR, 4-86 PLTS/PG 1 2 4, 4-87 POINT, 4-87 point deletion, 4-38 positive peak detection mode, 4-40 power bandwidth, 4-80 power cable, 2-10 power input, 3-7 power measurement, 4-88 adjacent channel power, 4-8 adjacent channel power extended, 4-9 channel bandwidth, 4-25 channel power, 4-25 channel spacing, 4-25 continuous sweep mode, 4-29 occupied bandwidth, 4-80 power bandwidth, 4-80 turning off, 4-72 power measurement setup, 4-109 Power Menu, 4-88 POWER ON IP LAST, 4-88 power-on state of analyzer, 4-88 power requirements, 2-6 preamplifier gain, 4-54 predetermined correction factors, 4-36 prefix catalog of, 4-23

clearing, 4-27 entry, 4-24 selection, 4-106 PRESET, 3-4, 4-89 PRESET , 4-92 preset conditions common, 4-90 for all models, 4-92 model specific, 4-90 Prev Menu, 4-93 PREV PAGE, 4-93 print black and white, 4-61 Print Config, 4-94 printer HP PaintJet, 4-30 PRINTER ADDRESS, 4-94 printer configuration, 4-29 PRINTER SETUP, 4-94 printing, 4-30, 4-33 HP PaintJet, 4-80 print menu, 4-95 PRINT REPORT, 4-94 Prn Port Config, 4-95 PRN PORT HPIB PAR, 4-95 PRN PORT SER PAR, 4-95 PROBE PWR, 3-4 program cataloging, 4-21 program (DLP) editor, 4-46 program editing and creating, 4-46 programming command dispose all , 4-48, 4-49, 4-50 $\,$ front panel execution, 4-51 PRT MENU ON OFF, 4-95 pulse edge definition, 4-48 Pulse Param, 4-95 pulse parameters clearing, 4-27 gate length, 4-34 PRI, 4-34, 4-48 pulse width, 4-33, 4-48 reference edge, 4-48 time gate utility, 4-95 pulse repetition interval, 4-48 pulse width, 4-48 PURGE AMP COR, 4-96 PURGE LIMITS, 4-96 PWR SWP ON OFF, 4-97

 ${f R}$ RAM card cataloging, 4-22 recalling test data, 4-98 ratio video bandwidth to resolution bandwidth, 4-132 real-time clock, 4-125 set date, 4-108set time, 4-108 time and date display, 4-125 rear-panel battery information label, 3-32 rear-panel features, 3-6-8 RECALL, 3-4, 4-97 RECALL AMP COR, 4-97 Recall Data , 4-98 RECALL FRM CRD, 4-98 RECALL FRM INT, 4-99 RECALL FRM INT softkey, 4-99 recalling a display image from the memory card, 4-20 recalling a DLP from the memory card, 4-20 recalling a state from the memory card, 4-21 recalling a test plan from analyzer memory, $4\mathchar`-99$ recalling a trace from analyzer memory, 4-63 recalling a trace from the memory card, 4-21 recalling test data from a RAM card, 4-98 RECALL LIMIT, 4-100 Recall TstPlan , 4-100 receiver mode, select, 4-129 reference, 4-121 reference connector, 2-12 reference detector +10V, 4-6-10V, 4-6 2V, 4-6 ground, 4-60 reference level, 4-100reference level offset, 4-100 reference measurement, 4-121 reference trace, 4-103 REF LVL, 4-100 REF LVL OFFSET, 4-100 relative, limit line type, 4-64 remote command dispose all, 4-48, 4-49, 4-50 front panel execution, 4-51 remote operation, 4-29 RES BW AUTO MAN, 4-100 resolution bandwidth, $4\math{-}100$ resolution bandwidth coupling, 4-15 gate utility, 4-33

results to screen, 4-43 RF OUT 75 Ω, 3-4 ROM card cataloging, 4-22 RPG knob, 3-6, 3-8 RPG TITLE, 4-101 RS-232 interface connector, 3-7 S sample detection mode, 4-40, 4-73 SAVE, 3-4, 4-101 SAVE AMP COR, 4-102 save current display, 4-40 save current state memory card, 4-116 state register, 4-116 saved analyzer state, 4-63 SAVE EDIT, 4-102 SAVE LIMIT, 4-102 save lock on MEM LOCKED, 4-73 SAVE TO INT MEM, 4-102 SAVE TRB TO INT, 4-103 Save TstPlan, 4-104 saving a test plan in analyzer memory, 4-102 saving DLPs from analyzer memory to card, 4-10 saving key definitions from analyzer memory to card, 4-10 saving to a memory card, 4-127 saving to analyzer memory, 4-127 SAV LOCK ON OFF, 4-101 SCALE LOG LIN, 4-104 screen annotation, 3-11-14 on or off, 4-14 screen characters, 4-60, 4-76, 4-117, 4-137 screen data output, 4-30 screen graticule, 4-60 screen, send results to, 4-43 screen text font guide conventions, 1-5 screen title, 4-25 characters, 4-8 clearing, 4-27 RPG TITLE, 4-101 SECAM-L, 4-104 segment deletion, 4-38 SELECT 1 2 3 4, 4-105 select all tests, 4-10 SELECT AMPLITUD, 4-105

resolution of marker counter, 4-28

SELECT DLT AMPL, 4-105 SELECT FREQ, 4-106 selecting the tune configuration, 4-26 SELECT LWR AMPL, 4-106 SELECT MID AMPL, 4-106 SELECT POINT, 4-106 SELECT PREFIX, 4-106 SELECT SEGMENT, 4-106 SELECT TIME, 4-106 SELECT TYPE, 4-107SELECT UPR AMPL, 4-107 self-calibration routine CAL TRK GEN, 2-16 self-calibration routines, 2-14, 3-21-26 amplitude, 4-18 CAL, 4-18 correction factors, 4-19 frequency, 4-19 frequency and amplitude, 4-19 tracking generator, 4-20 Service Cal, 4-3, 4-107 Service Diag, 4-4, 4-107 service functions, 4-3 Service Cal, 4-3 Service Diag, 4-4 SET ATTN ERROR, 4-3, 4-108 Set B&W Printer, 4-108 Set Colr Printer, 4-108 SET DATE, 4-108 SET TIME, 4-108 setting the amplitude, 3-17 setting the center frequency, 3-17 setting the marker, 3-19 setting the span, 3-17 setting the start time, 4-114 setting the stop time, 4-115 setting the time interval for timed testing, 4-122 setup power measurement functions, 4-109 printer, 4-94 Setup, 4-109 SGL SWP, 4-109 SHOW OPTIONS, 4-110 SIGNAL ID, 4-110 signal peak highest, 4-112 signal tracking, 4-82 SINGLE FFT, 4-110 SINGLE MEAS, 4-111
single sweep, 4-109, 4-117 SLOPE, 4-111 softkey guide conventions, 1-5 softkey label, 3-4 softkey locations, 5-2 softkey overview, 3-15 softkeys, 3-4 CABLE TV ANALYZER, 4-25 START FREQ, 4-121 STOP FREQ, 4-121 softkeys, function LISTEN OFF, 4-65 LISTEN ON, 4-65 softkeys, test and function CHANNEL SELECT, 4-25 MIN HOLD ON OFF, 4-73 RECALL FRM INT, 4-99 SAVE TRB TO INT , 4-99, 4-103 TAKE REF TRACE, 4-121 sound on off , $4\mathchar`-111$ source power, 4-113 space character, 4-137 span 0 Hz, 4-137 LAST SPAN, 4-64 window zone, 4-138 SPAN, 3-4, 3-17, 4-111 span, user defined channel, 4-131 SPAN ZOOM, 4-112 SPEAKER ON OFF, 4-112 spectral density channel power, 4-25 SPECTRUM ANALYZER, 4-112 speed data transmission, 4-16 split screen, 4-136 SQUELCH, 4-112SRC ATN MAN AUTO, 4-113 SRC PWR OFFSET, 4-113 SRC PWR ON OFF, 4-113 SRC PWR STP SIZE, 4-113 standard timebase, 4-19 standard tune configuration

selecting, 4-114 start and stop frequencies, 4-68 START FREQ, 4-113 START TIME, 4-114 state of analyzer at power on, 4-88 state registers locking, 4-101 STATE \rightarrow CARD, 4-116 STATE \rightarrow INTRNL, 4-116 STD, 4-114 step keys, 3-4, 3-9 step size, center frequency, 4-24 step-size coupling, 4-15 stimulus response mode, 4-118 STOP FREQ, 4-116 stop frequency, fft, 4-55 stopping a measurement, 4-8 STOP TIME, 4-115 store correction factors, 4-19 store data to RAM card, 4-103 STORE TO CARD, 4-103 storing display, 4-40 storing test results printing, 4-42 storing the correction factors CAL STORE, 2-15 storing the test results on a RAM card, 4-42 STOR PWR ON UNITS, 4-3, 4-116 STP GAIN ZERO, 4-5, 4-117 STUVWX, 4-117subtract display line from trace B, 4-16 subtract trace B from A, 4-7 SWEEP, 4-117 sweep control continuous, 4-117 single, 4-109, 4-117 SWEEP CONT SGL, 4-117 SWEEP DELAY gate utility, 4-117 sweep modes, 3-13, 3-14 SWEEP OUTPUT, 3-7 SWEEP RAMP, 4-5, 4-118 sweep time, 4-118 sweep time coupling, 4-15 gate utility, 4-34 SWEEP TIME DAC, 4-5, 4-118 switch trace A and B, 4-7

SWP CPLG SR SA, 4-118 SWP TIME AUTO MAN, 4-118 synchronization constants, 4-37 SYNC NRM NTSC, 4-119 SYNC NRM PAL, 4-119 SYS BTS TST CH#, 4-119 SYSTEM FRQ RESP, 4-120 SYSTEM MEAS, 4-120

Т Т, 4-121 table of markers, 4-75 table of amplitude correction factors, 4-10 table of signal peaks, 4-83 TABLE Δ DL NRM, 4-125 TAKE REF TRACE, 4-121 temperature range, 4-47 temperature conditions, 2-14 test and function operations analyzer into sample detection mode, 4-73 recall trace B from internal memory, 4-99 select channel number, 4-25 set analyzer state for test, 4-25 store a reference trace, 4-121 store trace B in internal memory, 4-103 testing initiating testing later, 4-122 test plan clearing, 4-27 recalling a test plan from analyzer memory, 4-99 saving a test plan in analyzer memory, 4-102 Test Plan , 4-34 test plan, recall, 4-100 test plan, recall from card, 4-98 test results, view last, 4-135 tests and functions channel select, 4-25 frequency response setup, 4-121 minimum hold, 4-73 recall trace B from internal, 4-99 save trace B to internal memory, 4-103 tests, clear all, 4-27 third order intercept measurement, 4-126 THRESHLD ON OFF, 4-125 threshold line, $4\mathchar`-125$ Time Date, 4-125 TIMEDATE ON OFF, 4-125

Timed Measure , 4-122 time domain window definition, 4-38 gate utility, 4-38 time gate definition of, 4-38 delay, 4-58 gate on/off, 4-58 length, 4-58 triggering, 4.58time gate utility, 4-59 coupling, 4-37 delay sweep, 4-117 entering, 4-58 exit, 4-51 marker on, 4-72 pulse parameters, 4-95 time interval maximum time interval, 4-122 TIME INTERVAL , 4-122 time window resolution bandwidth, 4-124sweep time, 4-124 TODAY'S DATE , 4-123 TODAY'S TIMEDATE , $4\mathchar`-124$ TOI ON OFF, 4-126 total power, input, 4-128 trace blank trace A, 4-17 blank trace B, 4-17 blank trace C, 4-17 clear write trace A, 4-27 clear write trace B, 4-27 clear write trace C, 4-28 move trace B into C, 4-16 A <--> B, 4-7 $A-B \rightarrow A, 4-7$ A \rightarrow C, 4-7 $B-DL \rightarrow B, 4-16$ view mode, 4-133 TRACE, 4-126 trace A maximum hold, 4-72 TRACE A, 4-126 TRACE A B C, 4-126 trace B maximum hold, 4-72 TRACE B, 4-126trace C

minimum hold, 4-73 TRACE C. 4-126 trace modes, 3-13, 3-14 trace number, 4-99 trace register, 4-103 trace registers locking, 4-101 Trace \rightarrow Card, 4-127 Trace \rightarrow Intrnl, 4-127 Track Gen, 4-127 tracking generator, 4-127 frequency adjustment, 4-67 internal/external leveling, 4-9 output, 3-4 output power, 4-113 peak response, 4-127 power sweep, 4-97 power sweep range, 4-113 self-calibration, 4-20 self-calibration routine, 2-16, 3-24 source attenuator, 4-113 source power offset, 4-113 source power step size, 4-113 warm-up time, 3-24 tracking generator self-calibration routine, 2-16 TRACKING PEAK, 4-127 $\mathrm{TRIG},\;4\text{-}128$ trigger, 4-128 external, 4-53 free run, 4-57 line, 4-65 video, 4-133 triggering EDGE POL POS NEG, 4-43 gate: edge or level, 4-58 gate utility, 4-43 trigger modes, 3-13, 3-14 TRIG MKR ON OFF, 4-128 TRMATH command, 4-23 TTL PWR @ INPUT, 4-128 T tune configuration, 4-121tune configuration AIR, 4-9 HRC, 4-61 IRC, 4-63 selecting, 4-26 STD, 4-114 turning off markers, 4-71 turning off the FFT function, 4-54

turning off windows display, 4-135 turning on the cable TV analyzer for the first time, 2-12 turn off active function, 3-6, 3-8 TV LINE #, 4-128 TV picture using WINDOWS (ON), 4-136 TV RECEIVER, 4-129 TV Standard, 4-129 TV SYNC NEG POS, 4-129 TV trig even field, 4-130 NTSC video format, 4-79 odd field, 4-130 PAL-M video format, 4-80 PAL video format, 4-80 SECAM-L video format, 4-104 TV line number, 4-128 TV Standard, 4-129 TV sync, 4-129 vertical interval, 4-130 TV TRIG, 4-118 TV TRIG EVEN FLD, 4-130 TV TRIG ODD FLD, 4-130 TV TRIG OUTPUT, 3-7 TV TRIG VERT INT, 4-130 T WINDOW RES BW, $4\mathchar`-124$ T WINDOW SWP TIME, 4-124 two display windows, 4-136 type of limit line, 4-107

U underscore

character, 4-137 units, amplitude, 4-11 UPDATE TIMEFREQ, 4-130 upper and lower limit lines editing, 4-45 user defined center freq, 4-131 user defined span, 4-131 User Menus, 4-130 USR CH 0 CNTR FRQ, 4-131 USR CH 0 SPAN, 4-131

V V, 4-132 variable cataloging, 4-21 VBW/RBW RATIO, 4-132 vertical signal positioning, 4-35 VHS video output, 4-119 VID AVG ON OFF, 4-132 VID BW AUTO MAN, 4-133 VIDEO, 4-133 video averaging, $4\mathchar`-132$ video bandwidth, 4-133 video bandwidth coupling, 4-15 gate utility, 4-34 video bandwidth to resolution bandwidth ratio, $4\mathchar`-132$ VIEW A, 4-133 VIEW B, 4-134 VIEW C, 4-134 view channel selection, 4-134VIEW CH SEL, 4-134 VIEW CO-CHNL, 4-134 VIEW INGRESS, 4-134 VIEW LAST, 4-135 view last test results, 4-135 visual carrier, 4-25 VOL-INTEN, 3-4 VOLTAGE SELECTOR, 3-7 voltage selector switch, 2-6 Volts, 4-135 volume, 4-65 volume control, 3-4

W warmup, 2-14 warmup time, 3-21 Watts, 4-135 when self-calibration is needed, 3-26 windows NEXT, 4-77 ON, 4-136 switching between, 4-77 switching time and frequency windows, 4-130 ZOOM, 4-138 windows keys, 3-4 WINDOWS OFF, 4-135 write-protect switch, 3-30 ${\bf X}$ $\,$ X FINE TUNE DAC, 4-5, 4-137 $\,$

Y YES, 4-137 YZ_# Spc Clear, 4-137

Z ZERO MARKER, 4-137 ZERO SPAN, 4-137 ZONE CENTER, 4-138 ZONE PK LEFT, 4-138 ZONE PK RIGHT, 4-138 ZONE SPAN, 4-138 ZOOM, 4-138 zooming a window, 4-138