

Keysight E6640A EXM Wireless Test Set

Notice: This document contains references to Agilent. Please note that Agilent's Test and Measurement business has become Keysight Technologies. For more information, go to www.keysight.com.

Notices

© Keysight Technologies, Inc. 2014

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies, Inc. as governed by United States and international copyright laws.

Trademark Acknowledgements

Bluetooth is a trademark owned by Bluetooth SIG, Inc., U.S.A, and licensed to Keysight Technologies.

cdma2000 is a US registered certification mark of the Telecommunications Industry Association.

WiMAX is a US trademark of the WiMAX Forum.

Copyright 2014 Keysight Technologies Inc.

Licensed under the Apache License, Version 2.0 (the "License"); you may not use this file except in compliance with the License.

You may obtain a copy of the License at: <http://www.apache.org/licenses/LICENSE-2.0>

Unless required by applicable law or agreed to in writing, software distributed under the License is distributed on an "AS IS" BASIS WITHOUT WARRANTIES OR CONDITIONS OF ANY KIND, either express or implied.

See the License for the specific language governing permissions and limitations under the License.

Manual Part Number

E6640-90001

Print Date

October 28, 2014

Published in USA

Keysight Technologies, Inc.
1400 Fountaingrove Parkway
Santa Rosa, CA 95403

Warranty

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, KEYSIGHT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. KEYSIGHT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR ANY INFORMATION CONTAINED HEREIN. SHOULD KEYSIGHT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS DOCUMENT THAT CONFLICT WITH THESE TERMS, THE WARRANTY TERMS IN THE SEPARATE AGREEMENT WILL CONTROL.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as "Commercial computer software" as defined in DFAR 252.227-7014 (June 1995), or as a "commercial item" as defined in FAR 2.101(a) or as "Restricted computer software" as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or

contract clause. Use, duplication or disclosure of Software is subject to Keysight Technologies' standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

In This Guide...

This guide contains the following information:

1 Quick-Start

This chapter explains how to initialize the test set and generate and view a signal.

2 Front and Rear Panel Features

Refer to this chapter for information on front- and rear-panel key functionality, and display annotations.

3 Operating Tasks

This chapter describes the various operating tasks and the procedures for carrying out those tasks.

4 Test Set Operating System

This chapter describes the Microsoft Windows configuration and the settings used with the Keysight test set software.

5 Using Microsoft Windows

The information in this chapter provides some guidelines for using the Microsoft Windows feature capabilities with the test set.

6 Troubleshooting

This chapter details some basic steps that may solve any problems you are experiencing with either the test set or Microsoft Windows.

Warranty

This Keysight Technologies product is warranted against defects in material and workmanship for a period of one year from the date of shipment. During the warranty period, Keysight Technologies will, at its option, either repair or replace products that prove to be defective.

For warranty service or repair, this product must be returned to a service facility designated by Keysight Technologies. Buyer shall prepay shipping charges to Keysight Technologies and shall pay shipping charges to return the product to Buyer. However, Buyer shall pay all shipping charges, duties, and taxes for products returned to Keysight Technologies from another country.

Where to Find the Latest Information

Documentation is updated periodically. For the latest information about this test set, including firmware upgrades, application information, and product information, see the following URLs:

<http://www.keysight.com/find/e6640a>

Information on preventing test set damage can be found at:

<http://www.keysight.com/find/PreventingInstrumentRepair>

Contents

1 Quick-Start

About the Test Set	12
Initial Inspection	14
Shipping problems?	15
Test Set Location and Mounting Requirements	16
Locating the test set	16
Cooling and rack mounting	16
Turning on the test set the first time	17
Initializing the test set using a USB mouse and external monitor	17
Options and Licenses	21
Option 4FD	22
Anti-Virus Software and Firewalls	23
Shortening the Test Set Power-On Time	23
Instrument Information	24
Power requirements	24
Environmental Conditions (Operating)	25
Ventilation	26
Instrument Maintenance	26
Protecting against electrostatic discharge	28

2 Front and Rear Panel Features

Front-Panel Features	30
Controller Hardware Interface	31
TRX Module Hardware Interface	33
Reference Module Hardware Interface	35
Virtual Front Panel	36
When the Test Set is Controlled Remotely	37

Display Annotations	38
Rear-Panel Features	40
Front and Rear Panel Symbols	41

3 Operating Tasks

Alignments	44
“All” Alignment (daily use)	44
Weekly alignment procedure	44
TRX Selection	45
Using HiSLIP	45
Using Socket Ports	46
Using VXI-11/SICL	46
Using Telnet Connection	46
E6640A Calibration Status	47
Certificates of Calibration	48
Desktop Icons	50
TRX Configurator	50
Launch XSA	51
Configure Applications	51
Port Configuration	52
Full Duplex Port Setup (RFIO 1 and RFIO 2)	52
Separate Input and Output Port Setup (RF3 I/O and RF4 I/O)	54
Separate Input and Output Port Setup (RFIO 1 and RFIO 2)	57
Option 4FD: Full Duplex Port Setup	59
Option 4FD: Separate Input and Output Port Setup	60
LAN Address Configuration	61
Switched MIMO	62
Constraints on Radio Standards	62
MIMO Port Configuration	62
Source Sync Control	63
RF Input Port Mode	63
Half Duplex RF Input	63
Acquisition Triggers	64
Acquisition Integration	64
Acquisition RF Input Port	65
Switched MIMO Modulation Accuracy measurement SCPI commands	66

Switched MIMO Results	68
True MIMO	70
TRX Configurations	70
Source Settings for MIMO	72
Constraints on TRX Selection	73
Constraints on Radio Standards	73
MIMO Type	73
Remote Analyzer Configuration	73
MIMO Modulation Analysis measurement SCPI commands	76
MIMO Modulation Analysis Results	77
Tx Beam Forming	79
Tx Beam Forming Commands	79
Tx Beam Forming Results	80

4 Test Set Operating System

Keysight Software Installed	82
Measurement Software Applications	82
Customer Installation of Software	83
3rd party software verified by Keysight	83
Installation of other 3rd party software	83
User Accounts	84
Administrator login	84
User login	84
KeysightOnly user account	84
Keysight service user accounts	84
Customer creation of accounts	85
Keysight Licensing Options	86
Time-Limited Licenses (Trial Licenses)	87
Licensing New Measurement Application Software - After Initial Purchase	88
Installation procedure over USB	88
Transporting a License Between Test Sets	90
Procedure for Transporting a License, Neither Test Set Connected to the Internet	90
Windows Configuration	93
Settings that can be changed	93

Settings that must not be changed	94
Configuring Printers	96
Configuring LAN	96
Hostname	96
IP Address & Gateway	96
Windows Security	97
Windows Firewall	98
Automatic updates	99
Virus protection	100
Spyware protection	100
System Maintenance	101
Backup	101
System restore	101
Disk defragmenting	101
USB Connections	102
Hard Drive Partitioning and Use	103
Hard Drive Recovery Process	104
Updating the software	106
Configuring recovery prompt timing	112

5 Using Microsoft Windows Operating System

Navigating Windows Without a Mouse	114
Remote Desktop: Using the Test Set Remotely	115
Overview of Remote Desktop operation	115
Setting up Remote Desktop operation	115
How to locate the computer name of the test set	117
Running a Remote Desktop session	118
The Virtual Front Panel	128
Programmatic Control of the Test Set	129
Capturing/Printing Displays and Windows	130
Windows Shortcuts and Miscellaneous Tasks	131
Windows shortcuts (key combinations)	131
Windows taskbar auto-hide feature	133
Windows startup folder	135

6 Troubleshooting

Check the Basics	138
Problems with Microsoft Windows Operating System	140
Returning a test set for Service	141
Calling Keysight Technologies	141
Locations for Keysight Technologies	141
Read the Warranty	142
Service Options	142
Packaging the Test Set	143

Index

1 Quick-Start

This section explains how to initialize the test set and generate and view a signal.

The following topics can be found in this section:

"About the Test Set" on page 12

"Initial Inspection" on page 14

"Test Set Location and Mounting Requirements" on page 16

"Turning on the test set the first time" on page 17

"Options and Licenses" on page 21


"Anti-Virus Software and Firewalls" on page 23

"Shortening the Test Set Power-On Time" on page 23


"Instrument Information" on page 24

About the Test Set

The E6640A EXM Wireless Test set contains one to four keysight M9430A TRXs. Each TRX contains the full set of hardware elements illustrated below, and is run by its own instance of the XSA firmware application (a fully loaded test set shows four independent XSA windows on its monitor display).







The TRX includes a signal analyzer and a signal source, both of which interface with the front panel of the test set through an input/output matrix, and each of which is controlled by an independent list sequencer (the two sequencers exchange trigger signals to coordinate their operations). The input/output matrix is illustrated below. Ports RF3 I/O and RF4 I/O can be inputs or outputs, but not simultaneously. See also: **“TRX Selection”** on page 45 and **“Port Configuration”** on page 52.



Initial Inspection

Inspect the shipping container and the cushioning material for signs of stress. Retain the shipping materials for future use, as you may wish to ship the test set to another location or to Keysight Technologies for service.

Verifying the contents

Item	Deliverable	Description
Getting Started Guide		Provides first-time power on instructions, licensing information, operating system information, and general hardware information.
USB keyboard & mouse		A keyboard and mouse for communication with the controller and measurement applications. (These accessories are provided only if options E6640A-KYB and E6640A-MSE are ordered.)
Monitor cable adapter (Display Port to VGA)		An adapter for the front panel Display Port which makes it possible to use the test set with a monitor which has a VGA connector.
Power cable		Connection for test set power.

Shipping problems?

If the shipping materials are damaged or the contents of the container are incomplete:

- Contact the nearest Keysight Technologies office.
- Keep the shipping materials for the carrier's inspection.
- If you must return an test set to Keysight Technologies, use the original (or comparable) shipping materials. See **"Returning a test set for Service"** on page 141.

Test Set Location and Mounting Requirements

Locating the test set

Make sure that the fan inlet and exhaust vent areas on the sides, bottom, and back of the test set chassis are not obstructed. The minimal required clearance is 2 inches. Airflow restrictions cause additional airflow noise and cause the fans to speed up so they can draw in enough air for the required cooling resulting in excessive audible noise.

Cooling and rack mounting

Do not mount the test set side-by-side with any other instrument with side-by-side ventilation. Make sure the exhaust air from one instrument is directed away from the inlet of any other instrument (heated air directed into the inlet of an instrument can cause instrument failure due to excessive operating temperatures). The test set draws air in from the sides and bottom of the chassis and exhausts air from the back.



Turning on the test set the first time

NOTE


During the initial power-on process, the test set turns off and restarts several times. This *only* happens during the initial installation process.

Initializing the test set using a USB mouse and external monitor

Steps	Actions	Notes
1 Connect a mouse and keyboard	<ul style="list-style-type: none"> Connect the mouse and keyboard to two of the USB ports on the front panel of the test set. 	
2 Connect a monitor	<ul style="list-style-type: none"> Connect the monitor to the monitor port on the front panel of the test set. 	<ul style="list-style-type: none"> Model E6640A has no display; an external monitor is required. See “Virtual Front Panel” on page 36.
3 Power on the test set	<ol style="list-style-type: none"> Position the test set so you have easy access to the power cord and plug it in. Press the power switch (located near the top right corner of the PXI rack) to turn the test set on. 	<ul style="list-style-type: none"> See “Test Set Location and Mounting Requirements” and “Power requirements” on page 24 for more details. The test set can require more than 5 minutes to power-on. This is affected by Windows start-up requirements. The Keysight Technologies screen appears. <div data-bbox="850 1232 1185 1383" data-label="Image"> </div> <ul style="list-style-type: none"> At this time, it is safe to turn off the test set before starting the Setup Wizard.
4 Launch Microsoft Setup Wizard	<ol style="list-style-type: none"> Highlight Microsoft Windows and press Enter to start the pre-installed operating system. 	<div data-bbox="841 1505 1427 1789" data-label="Image"> </div>


Steps	Actions	Notes
	<p>a The analyzer performs the following steps:</p> <ul style="list-style-type: none"> • Windows Startup window • Black screen • Windows message window, “Please wait while windows prepares to setup” • Keysight window <p>b At this time, it is safe to turn off the instrument before initializing the software.</p>	
CAUTION		<ul style="list-style-type: none"> • After launching the setup, do not turn off the instrument or remove power before the Setup Wizard completes and the system restarts. Turning off the instrument may corrupt the system and the application may not operate.

- 5 Reboot and log on
- a Click **OK**.
- When the instrument restarts, the following message window appears:




- The system reboots.
- Several required processes continue.
- The system reboots.
- The XSA application initializes

- 6 Separate instances of the application are launched for each TRX installed in the PXI rack (they are identified as “TRX1” through “TRX4” in the upper left corner of the application window). The verification procedures which follow should be done for all TRXs.



Steps	Actions	Notes
7 Verify the installation	<p>a From the mouse right-click menu, select Utility, Virtual Front Panel. (All key-presses that follow refer to the keys shown on the Virtual Front Panel.)</p> <p>b Press System, Show, System.</p> <p>c Verify that the purchased application(s) appear in the list.</p>	<ul style="list-style-type: none"> If you require further assistance, contact the Keysight support team. Online assistance: http://www.keysight.com/find/assist
8 Attach a loopback cable	<p>a Use an RF cable to connect the front panel RFIO 1 and RFIO 2 ports for this TRX.</p>	<ul style="list-style-type: none"> The purpose is to use a loopback cable to verify that the test set's source can generate a signal and the test set's analyzer can measure it.
9 Generate and view a signal	<p>a Press Input/Output, RF Input, RF Input Port, RFIO 1.</p> <p>b Press Input/Output, More, RF Output & Test Set Config, RF Output Port, RFIO 2.</p> <p>c Press AMPTD Y Scale, Ref Value, 10, dBm.</p> <p>d Press FREQ, Center Freq, 2.4, GHz.</p> <p>e Press SPAN X Scale, Span, 8, MHz.</p> <p>f Press Source, Frequency, 2.4, GHz.</p> <p>g Press Source, RF Output, On.</p> <p>h Press Source, Amplitude, RF Power, 0, dBm.</p>	<ul style="list-style-type: none"> Select RFIO 1 as the input port Select RFIO 2 as the output port. Set the amplitude reference to 10 dBm Set the analyzer center frequency to 2.4 GHz. Set the analyzer span to 8 MHz. Set the source frequency to 2.4 GHz. Set the source RF output to On. Set the source amplitude to 0 dBm. To ensure accurate power levels from the source, see "Alignments" on page 44.

Steps	Actions	Notes
	<ul style="list-style-type: none"><li data-bbox="495 378 950 409">• The 2.4 GHz signal appears on the display.	 <p>The screenshot shows the Keysight TRX3 IQ Analyzer (Basic) - Complex Spectrum interface. The main display area shows a spectrum plot with a prominent peak at 2.4 GHz. The plot is titled 'Complex Spectrum' and has a center frequency of 2.40000000 GHz. The y-axis is labeled '10 dB/div' and 'Ref 10 dBm'. The x-axis is labeled 'Center 2.4 GHz' and 'Span 8 MHz'. A marker is placed on the peak, labeled 'Marker 1 X 2.400029197 GHz' with a value of '-0.58 dBm'. The plot shows a signal with a bandwidth of 12.000 MHz. The interface includes a top menu bar with options like 'RF ON', 'ALIGN OFF', and 'Trig: Free Run'. On the right side, there is a control panel with buttons for 'Source', 'RF Output', 'Amplitude', 'Modulation', and 'Frequency'. The 'Frequency' button is currently set to 2.40000000 GHz. The bottom of the interface shows 'Digital IF BW 12.000 MHz' and 'Capture Time 23.467 μs'.</p>

Options and Licenses

The options and licenses installed on the test set are listed on the **System > Show > System** display, as illustrated below. Certain features of the test set (such as its frequency range, its analysis bandwidth, the measurement applications it runs, and in some cases individual measurements within an application) are licensed features, and are unavailable if the appropriate license is not installed.

NOTE Hardware-related options such as frequency range and analysis bandwidth are specific to the TRX; measurement applications apply to any TRX in the same test set chassis.

Host ID	PCSERNO,GA47677499
mDNS Enabled	Yes
mDNS Host Name	A-E6640A-0111-2
mDNS Service Name	Keysight PCSERNO Wireless Connectivity Test Set - SC29...
Option	Name / Description
E6640A-504	Frequency Range, 3.8 GHz
E6640A-PCX	Intel(R) Core(TM) i7-3610QE CPU @ 2.30GHz, 4 GB
E6640A-W7X	Windows Embedded Standard 7, 64 bit OS
E6640A-506	Frequency Range, 6.0 GHz
E6640A-B40	Analysis bandwidth, 40 MHz
E6640A-RFO	RFIO
E6640A-SRC	Signal Source
N7699A-D24	Measurement Applications Temporary License - demo only
V9060A-2FP	Test Set Measurement Application
V9063A-2FP	Analog Demod Measurement Application
V9065A-2FP	Sequence Analyzer for MS Application
V9071A-2FP	GSM/EDGE Measurement Application
V9071A-3FP	EDGE Evolution Measurement Application
V9073A-2FP	IS97A/HSRPA Measurement Application

NOTE For many of the measurement applications, the license is listed here in separate "A" and "B" versions. For example, when the "V9071B-2FP" license is purchased and installed, the "V9071A-2FP" license is automatically installed along with it. This is not an error; both versions of the license should be present.

Option 4FD

Option 4FD modifies the RFIO interface of the E6640A EXM Wireless Test Set as follows:

- All four ports on each module are full-duplex ports (that is, any port can be used simultaneously as an RF input (to the test set's analyzer) and as an RF output (from the test set's source).
- Each TRX in the test set has the model designation M9431A, not M9430A.
- The port designations on the front panel are: RFIO 1, RFIO 2, RFIO 3, and RFIO 4.
- Performance characteristics for all four ports are same as those described for ports RFIO 1 and RFIO 2 in the E6640A EXM Wireless Test Set Data Sheet.
- Port settings under the **Input/Output** menu (and equivalent SCPI commands) are modified accordingly.

Any of the TRX's four RFIO ports can be configured as a full duplex port; it can also be configured as an input port (accepting a signal to be analyzed), with a separate port configured as an output port (providing an RF stimulus). The configuration details are described in "**Option 4FD: Full Duplex Port Setup**" on page 59 and "**Option 4FD: Separate Input and Output Port Setup**" on page 60.

Anti-Virus Software and Firewalls

No anti-virus software is shipped with the test set. It is recommended that you install anti-virus software if your test set is connected to the LAN. Check with your IT department to see what they recommend.

The test set is shipped with the Windows firewall enabled. Do not modify the default network settings as this may cause problems with the operating system of the test set.

Shortening the Test Set Power-On Time

The test set desktop includes a shortcut to the ConfigureApplications.exe application that lets you control your test set power-on configuration. You can set the configuration to pre-load only the applications that you typically run. This can significantly shorten the time it takes for your test set to power-on. You can also access this functionality by pressing: **System, Power On, Configure Applications**. (These key-presses refer to the Virtual Front Panel; see **"Virtual Front Panel"** on page 36.)

If (after power-on) you want to access an application that you did not pre-load, it will take a little longer to bring it up once you press the application softkey. This longer time-frame is only for the first access after power-on. After the initial access, mode (application) switching will operate as quickly as if the application was pre-loaded.

Instrument Information

Power requirements

The only physical installation of your Keysight test set is a connection to a power source. Line voltage does *not* need to be selected.

This test set does *not* contain customer serviceable fuses.

WARNING

This is a Safety Class 1 Product (provided with a protective earthing ground incorporated in the power cord). The mains plug shall only be inserted in a socket outlet provided with a protective earth contact. Any interruption of the protective conductor inside or outside of the product is likely to make the product dangerous. Intentional interruption is prohibited. (IEC 348 clauses 17.3.3c & 17.3.4)

Failure to ground the test set properly can result in personal injury. Before turning on the test set, you must connect its protective earth terminals to the protective conductor of the main power cable. Insert the main power cable plug into a socket outlet that has a protective earth contact only. DO NOT defeat the earth-grounding protection by using an extension cable, power cable, or autotransformer without a protective ground conductor.

CAUTION

This product is designed for use in Installation Category II and Pollution Degree 2 per IEC 61010 Second Edition and IEC 664 respectively. This instrument has autoranging line voltage input. Before switching on the instrument, be sure the supply voltage is within the specified range.

The Mains wiring and connectors shall be compatible with the connector used in the premise electrical system. Failure, to ensure adequate earth grounding by not using the correct components may cause product damage, and serious injury.

AC power cord

The test set is equipped with a three-wire power cord, in accordance with international safety standards. This cable grounds the test set cabinet when connected to an appropriate power line outlet. The cable appropriate to the original shipping location is included with the test set. See:

<http://www.keysight.com/find/powercords>

CAUTION

Always use the three-prong AC power cord supplied with this product. Failure to ensure adequate earth grounding by not using this cord can cause product damage.

This instrument has auto-ranging line voltage input; be sure the supply voltage is within the specified range and voltage fluctuations do not to exceed 10 percent of the nominal supply voltage:

100/120 VAC 50/60 Hz, 1000 W MAX
220/240 VAC 50/60 Hz, 1200 W MAX.

WARNING

If this product is not used as specified, the protection provided by the equipment could be impaired. This product must be used in a normal condition (in which all means for protection are intact) only.

Install the instrument so that the detachable power cord is readily identifiable and easily reached by the operator. The detachable power cord is the instrument disconnecting device. It disconnects the mains circuits from the mains supply before other parts of the instrument. The front panel switch is only a stand by switch and is not a LINE switch. Alternatively, an externally installed switch or circuit breaker (which is readily identifiable and is easily reached by the operator) may be used as a disconnecting device.

Environmental Conditions (Operating)

This product is designed for use in the following conditions:

- For indoor use only
- Altitude up to 3000 meters
- Maximum relative humidity 80% for temperatures up to 31 degrees C, decreasing linearly to 50% relative humidity at 40 degrees C.

CAUTION

This product is designed for use in Installation Category II and Pollution Degree 2 per IEC 61010 Second Edition and IEC 664 respectively.

Ventilation

CAUTION

VENTILATION REQUIREMENTS: When installing the product in a cabinet, the convection into and out of the product must not be restricted. The ambient temperature (outside the cabinet) must be less than the maximum operating temperature of the product by 40C for every 100 watts dissipated in the cabinet. If the total power dissipated in the cabinet is greater than 800 watts, then forced convection must be used.

Instrument Maintenance

Cleaning the instrument

WARNING

To prevent electrical shock, disconnect the test set from mains before cleaning. Use a dry cloth or one slightly dampened with water to clean the external case parts. Do not attempt to clean internally.

Cleaning the connectors

WARNING

Cleaning connectors with alcohol shall only be done with the instrument power cord removed, and in a well-ventilated area. Allow all residual alcohol moisture to evaporate, and the fumes to dissipate prior to energizing the instrument.

Battery Information

The test set uses a lithium battery located on the CPU board. This is not an operator replaceable part. See **“Returning a test set for Service”** on page 141. Replaceable parts must be approved or supplied by Keysight Technologies.

You can order the service documentation for the instrument through your Keysight Sales and Service office.

WARNING

Danger of explosion if battery is incorrectly replaced. Replace only with the same or equivalent type recommended. Discard used batteries according to the manufacturer's instructions.

Do not throw batteries away but collect as small chemical waste.



**DO NOT THROW BATTERIES AWAY BUT
COLLECT AS SMALL CHEMICAL WASTE.**

Protecting against electrostatic discharge

Electrostatic discharge (ESD) can damage or destroy electronic components (the possibility of unseen damage caused by ESD is present whenever components are transported, stored, or used).

Test equipment and ESD

To help reduce ESD damage that can occur while using test equipment:

WARNING

Do not use these first three techniques when working on circuitry with a voltage potential greater than 500 volts.

- Before connecting any coaxial cable to a test set connector for the first time each day, momentarily short the center and outer conductors of the cable together.
- Personnel should be grounded with a 1 M Ω resistor-isolated wrist-strap before touching the center pin of any connector and before removing any assembly from the test set.
- Be sure that all instruments are properly earth-grounded to prevent build-up of static charge.
- Perform work on all components or assemblies at a static-safe workstation.
- Keep static-generating materials at least one meter away from all components.
- Store or transport components in static-shielding containers.
- Always handle printed circuit board assemblies by the edges. This reduces the possibility of ESD damage to components and prevent contamination of exposed plating.

Additional information about ESD

For more information about ESD and how to prevent ESD damage, contact the Electrostatic Discharge Association (<http://www.esda.org>). The ESD standards developed by this agency are sanctioned by the American National Standards Institute (ANSI).

2 Front and Rear Panel Features

This section describes the following features:

"Front-Panel Features" on page 30

"Display Annotations" on page 38

"Rear-Panel Features" on page 40

"Front and Rear Panel Symbols" on page 41


Front-Panel Features

The test set consists of instruments loaded in a PXI rack, hidden behind a front panel. There is a controller (located on the far left), and one to four TRX (transmit/receive) sub-instruments, each of which includes an RF signal generator and an RF signal analyzer. A frequency-reference instrument, shared by all the TRX sub-instruments, is placed in the center.



Controller Hardware Interface

The controller’s front panel features are illustrated below:



#	Item Name	Description
1	Monitor Port	This connector supports a connection to a monitor with a DisplayPort input.
2	LED Indicators	<ul style="list-style-type: none"> DRIVE: indicates when the internal hard drive is being accessed. POWER: steady green to indicate that chassis power is on; blinking green to indicate the controller has entered the soft off state and is safe to power down; off to indicate that the controller is powered off; blinking red to indicate a power fault; steady red to indicate a thermal fault and resulting shutdown. USER: a user-definable indicator.
3	TRIG	This trigger line is not used by the E6640A.

2 Front and Rear Panel Features


#	Item Name	Description
4	LAN Connectors	Two TCP/IP Interface connectors that are used for remote test set operation. Choose the LAN 1 port to have an IP address assigned to the test set dynamically, using DHCP. If your local network does not support DHCP, choose the LAN 2 port to use a private static IP address. For more information, see "LAN Address Configuration" on page 61.
5	USB 3.0	The controller provides two USB 3.0 ports.
6	USB 2.0	The controller provides four USB 2.0 ports.

TRX Module Hardware Interface

There can be as many as four TRX modules in the PXI rack; each has the same hardware interface, as illustrated below.

NOTE

If the E6640A test set was purchased with Option 4FD, all four RF ports have full-duplex capability, therefore two of the ports are labeled and used differently than they are described below. For details, see **“Option 4FD”** on page 22, **“Option 4FD: Full Duplex Port Setup”** on page 59 and **“Option 4FD: Separate Input and Output Port Setup”** on page 60.




Item		Description
#	Name	
1	TRXn STATUS	After line power is applied to the PXI chassis, the STATUS LED for each TRX module lights orange when the TRX hardware is initialized, and green when the software controlling the TRX hardware is ready.
2	RFIO 2	This Type N connector is a full-duplex RF input/output port.
3	RFIO 1	This Type N connector is a full-duplex RF input/output port.
4	TRIG OUT 2	A trigger output used to synchronize other test equipment with the test set. Configurable from the Input/Output keys.

2 Front and Rear Panel Features

Item		Description
#	Name	
5	TRIG IN 1	A trigger input used to synchronize other test equipment with the test set. Configurable from the Input/Output keys.
6	RF4 I/O	This Type N connector is a half-duplex RF port which can be configured as an input or an output. (If Option 4FD is installed, see “Option 4FD” on page 22.)
7	RF3 I/O	This Type N connector is a half-duplex RF port which can be configured as an input or an output. (If Option 4FD is installed, see “Option 4FD” on page 22.)
8	TRIG OUT 1	A trigger output used to synchronize other test equipment with the test set. Configurable from the Input/Output keys.
9	TRIG IN 2	A trigger input used to synchronize other test equipment with the test set. Configurable from the Input/Output keys.

Reference Module Hardware Interface

The connectors which are exposed when the impact cover is in place are illustrated below:



Item		Description
#	Name	
1	10 MHz OUT	This BNC input port provides a timebase reference output from the M9300A Reference module.
2	REF IN	This BNC input port accepts a timebase reference input to the Ref In port of the M9300A Reference module.

Virtual Front Panel


The E6640A model does not include physical front-panel keys. However, if you have a PC mouse, monitor, and keyboard plugged in (or are working through a Remote Desktop interface), you can navigate front-panel functions using the virtual front panel (VFP) shown below. Access the VFP as follows:

1. Right-click the mouse.
2. Left-click Utility (1) in the menu, as shown below.
3. Left-click Virtual Front Panel (2) in the menu, as shown below.


NOTE

The PC mouse and monitor are required when using the E6640A. For ease in using the VFP, the PC keyboard is recommended.

:



On the VFP the keys labeled "Key 1" through "Key 7" function as the menu keys. Using the mouse to click on a combination of the VFP keys and the menu keys on the display screen, you can operate the instrument as if it had conventional front-panel keys. The VFP frame shows "Left" or "Right" at the upper left corner of the frame, so that it's clear which sub-instrument within the PXI rack is being controlled by it.



When the Test Set is Controlled Remotely

The test set can be controlled remotely, from another computer. (See [“Remote Desktop: Using the Test Set Remotely”](#) on page 115.) When the test set is being controlled remotely, a log-on screen is displayed on the monitor.

Under these circumstances, pressing the front-panel Power switch does not immediately turn instrument power off. Instead, the following message appears:

Other people are logged on to this computer. Shutting down windows might cause them to lose data. Do you want to continue shutting down?


Although pressing and holding down the Power button will force a power shutdown, this is not recommended; use the Windows shutdown function instead.



WARNING

Do not shut down power by pulling out the power plug; this will damage the power supply.

2 Front and Rear Panel Features


Display Annotations



Item		Function Key
#	Description	
1	Measurement bar; shows general measurement settings and information. Some measurements include limits which the data is tested against. A Pass/Fail indication may be shown in the lower-left corner. The following graphics indicate single/continuous measurement:  	All the keys in the test set Setup part of the front panel.
2	Active Function (measurement bar); when the current active function has a settable numeric value, it is shown here.	Currently selected front panel key.
3	The TRX sub-instrument associated with the application (identified as "TRX1" through "TRX4"). This is needed because the E6640A can include up to four independent TRX sub-instruments in one PXI rack.	
4	Banner; shows the name of the selected application that is currently running.	Mode
5	Measurement title; shows title information for the current measurement, or a title that you created for the measurement.	Meas View/Display, Display, Title













Item		Function Key
#	Description	
6	<p>Settings panel; displays system information that is not specific to any one application.</p> <ul style="list-style-type: none"> ▪ Input/Output status - green LXI indicates the LAN is connected. RLTS indicate Remote, Listen, Talk, SRQ ▪ Input impedance and coupling ▪ Selection of external frequency reference ▪ Setting of automatic internal alignment routine 	Local and System, I/O Config, Input/Output, Amplitude, System, and others
7	Active marker frequency, amplitude, or function value	Marker
8	Settings panel; time and date display	System, Control Panel
9	Trace and detector information	Trace/Detector, Clear Write (W) Trace Average (A) Max Hold (M) Min Hold (m) Trace/Detector, More, Detector, Average (A) Normal (N) Peak (P) Sample (S) Negative Peak (p)
10	Key labels that change based on the most recent key press	Softkeys
11	Displays information, warning, and error messages. Message area: single events. Status area: conditions.	
12	Measurement settings for the data currently being displayed in the graticule area. In the example above: center frequency, resolution band width, video band width, frequency span, sweep time and number of sweep points.	Keys in the test set Setup part of the front panel.

Rear-Panel Features



Item		Description
#	Name	
1	Line power input	The AC power connection. See the product specifications for more details.
2	FAN	Fan speed control (settings are HIGH and AUTO).
3	INHIBIT	Inhibit switch. The settings are DEF (default) and MAN (manual). <ul style="list-style-type: none"> When the switch is in the default position, the chassis is powered up by the front panel ON/Standby pushbutton. This is the recommended setting. When the switch is in the manual position, the chassis is powered up by the Inhibit input signal on the rear panel INHIBIT/VOLTAGE MON DB-9 connector.
4	INHIBIT/ VOLTAGE MON	This DB-9 connector provides access for testing of internal power supply voltages. The voltages (by Pin #) are: (1) Logic Gnd, (2) +5 VDC, (3) Rsrvd, (4) +3.3 VDC, (5) Inhibit [Low], (6) +12 VDC, (7) Rsrvd, (8) -12 VDC, (9) Logic Gnd.
5	10 MHz REF OUT	This BNC connector is not used by the E6640A test set. The front panel 10 MHz OUT connector should be used instead.
6	10 MHz REF IN	This BNC connector is not used by the E6640A test set. The front panel REF IN connector should be used instead.

Front and Rear Panel Symbols

	This symbol is used to indicate power ON (green LED).
	This symbol is used to indicate power STANDBY mode (yellow LED).
	This symbol indicates the input power required is AC.
	The instruction documentation symbol. The product is marked with this symbol when it is necessary for the user to refer to instructions in the documentation.
	The CE mark is a registered trademark of the European Community.
	The RCM Mark is a Compliance Mark according to the ACMA Labelling Requirement.
	This is a marking of a product in compliance with the Canadian Interference-Causing Equipment Standard (ICES-001). This is also a symbol of an Industrial Scientific and Medical Group 1 Class A product (CISPR 11, Clause 4).
	The CSA mark is a registered trademark of the Canadian Standards Association International.
	<p>South Korean Class A EMC Declaration A 급 기기 (업무용 방송통신기자재) 이 기기는 업무용 (A 급) 전자파적합기기로서 판매자 또는 사용자는 이 점을 주의하시기 바라 며 , 가정외의 지역에서 사용하는 것을 목적으로 합니다 .</p> <p>This equipment is Class A suitable for professional use and is for use in electromagnetic environments outside of the home.</p>
	This symbol indicates separate collection for electrical and electronic equipment mandated under EU law as of August 13, 2005. All electric and electronic equipment are required to be separated from normal waste for disposal (Reference WEEE Directive 2002/96/EC).
	Indicates the time period during which no hazardous or toxic substance elements are expected to leak or deteriorate during normal use. Forty years is the expected useful life of the product.
	This symbol on all primary and secondary packaging indicates compliance to China standard GB 18455-2001.
	To return unwanted products, contact your local Keysight office.

2 Front and Rear Panel Features

3

Operating Tasks

"Alignments" on page 44

"TRX Selection" on page 45

"E6640A Calibration Status" on page 47

"Desktop Icons" on page 50

"Port Configuration" on page 52

"LAN Address Configuration" on page 61

"Switched MIMO" on page 62

"True MIMO" on page 70

"Tx Beam Forming" on page 79

Alignments

Alignments are internal calibration adjustments which each TRX module must make to ensure that internal signal levels are properly maintained. To avoid interruptions, the alignments are not run automatically, either at startup or afterward; you must run them explicitly.

NOTE The alignment process requires a 45-minute preliminary warm-up period, after applying power or after restarting the XSA application (when the application is not running, power is not applied to the TRXs; therefore an application shutdown has the same effect on alignment as a power shutdown).

“All” Alignment (daily use)

This alignment of the source and analyzer in the TRX is sufficient to maintain specified performance, provided that (1) the TRX’s internal temperature has not drifted more than 5 degrees C since the previous alignment, (2) no more than 8 hours have elapsed since the previous “All” alignment, and (3) no more than 1 week has elapsed since the previous weekly alignment (see below). The “All” alignment typically takes less than 5 minutes to run.

The TRX’s internal temperature can be checked by sending the SCPI query
:CAL:TEMP:CURR?

To run the alignment, select **System > Alignments > Align Now > All**, or send the SCPI command :CAL:ALL.

The alignment requires a preliminary 45-minute warm-up period, as described in the note above.

Weekly alignment procedure

A more thorough alignment of the source and analyzer in the TRX is required on a weekly basis. This procedure is similar to the “All” alignment described above, but it also performs additional alignment functions related to slow-drifting parameters which do not change significantly over time periods shorter than a week.


There is no function on the **System > Alignments** menu which runs all parts of the weekly alignment procedure. To run the alignment, send the following sequence of SCPI commands:

```
:CAL:IF
*WAI
:CAL:RF
*WAI
:CAL:INT:SOUR
*WAI
```

The alignment requires a preliminary 45-minute warm-up period, as described in the note above.

TRX Selection

On the monitor or remote desktop view of the test set, each TRX is represented by a separate XSA window. You can identify which TRX a window controls by looking for the identifier (TRX1 through TRX4) at the top left corner of the window. (The TRXs are placed in numerical order in the PXIe chassis, beginning with TRX1 on the left.)



Although the test set as a whole is assigned only one IP address for the entire PXIe chassis, the individual TRXs within the chassis can be differentiated in network communication with the test set. The means of identifying a TRX depends upon the connection method, as outlined below.

Using HiSLIP

NOTE

HiSLIP (High-Speed LAN Instrument Protocol) is the recommended interface for connection to the E6640A, because of its superior performance characteristics.

Use hislip0 through hislip3:

- TRX1 :
TCPIP0::::hislip0::INSTR
- TRX2 :
TCPIP0::::hislip1::INSTR
- TRX3 :
TCPIP0::::hislip2::INSTR
- TRX4 :
TCPIP0::::hislip3::INSTR

Using Socket Ports

Use sockets 5025, 5125, 5225, and 5325:

- TRX1, 5025:
TCPIP0::::5025::SOCKET
- TRX2, 5125:
TCPIP0::::5125::SOCKET
- TRX3, 5225:
TCPIP0::::5225::SOCKET
- TRX4, 5325:
TCPIP0::::5325::SOCKET

Using VXI-11/SICL

Use inst0 through inst3:

- TRX1:
TCPIP0::::inst0::INSTR
- TRX2:
TCPIP0::::inst1::INSTR
- TRX3:
TCPIP0::::inst2::INSTR
- TRX4:
TCPIP0::::inst3::INSTR

Using Telnet Connection

Use telnet ports 5023, 5123, 5223, 5323:

- TRX1:
5023
- TRX2:
5123
- TRX3:
5223
- TRX4:
5323

E6640A Calibration Status

The E6640A EXM test set does not have a calibration certificate, because of its flexibility in configuration:

- The E6640A chassis can have 1, 2, 3, or 4 TRXs (M9430A) installed.
- Different TRXs can have different hardware configurations for frequency range and/or bandwidth.
- Additional TRXs can be added to an E6640A chassis after the initial instrument purchase. These would have different calibration due dates, and different warranty end dates, from the modules that were originally purchased with the E6640A.

There isn't a single instrument-level calibration (or calibration due-date) that covers the E6640A as a whole and all of the TRX modules (and the Reference module) within it. Each module is calibrated independently, and not necessarily at the same time that the others in the test set are calibrated.

The only way to track the calibration status of the E6640A is to track the calibration status of the reference module, and each TRX module, within it. The serial number of a particular module can be found from its XSA window, by selecting **System > Show > Hardware** on the Virtual Front Panel. The serial number is listed in the row which shows "Reference" or "M9430 TRX" under Assembly Name. The same row also shows the Cal Date and Cal Interval for the relevant module

Keysight TRX1 - Sequence Analyzer - List Sequencer (Prototype - Limited Sale Allowed)

Output:RFIO1

Hardware Information
 Keysight EXM Wireless Test Set
 Product Number: E6640A
 Serial Number: US00000001
 Instrument S/W Revision: E.14.50_R0017
 Revision Date: 8/28/2014 10:31:33 AM


Assembly Name	Part #	Serial #	Matl Rev	Rev	OF Rev	Hw Id	Cal Date	Cal Interval
PXIe Controller	PXIe-8135			0		174		
Reference	M9300A	US52360177				0165	2013.06.28	2 years
Reference Carrier	M9300A	W1312-63281-23-1249-00095		01	1.2.1	0166		
Reference Plugin	M9300A	W1312-63286-30-1250-00043		01	1.2.0	0167		
Chassis	M9018A	TW53330145		0		171		
Wideband Digitizer	M943063005	00000000028	001	0	0	138		
DownConverter	M943063006	00000000028	001	0	A	139		
RFIO	M943063009	00000000028	001	0	A	140		
Power Supply	M943063002	00000000028	001	0	A	141		
(M9430 TRX)	9142601882	SN000000000	0	0	0	142	1970.01.01	1 year
Baseband Generator	M943063007	00000000028	002	0	0	136		

The same hardware information is also returned by the SCPI command:
`SYST:CONF:HARD?`

Certificates of Calibration

Customers receive 2 to 5 Certificates of Calibration (CoCs) with the E6640A EXM test set. The EXM is a scalable test set and can be purchased with 1 to 4 TRX modules. The TRX modules, along with the Reference module (supplied with each E6640A), are the only items that require calibration, and therefore they are the only items supplied with a CoC and calibration sticker at shipment.

Each E6640A test set is shipped with a yellow envelope which contains your calibration certificates; this envelope is labeled (lower right corner) with a sticker identifying the unit model number (E6640A) and serial number. Inside each envelope will be from two to five CoCs: one for the reference module, and one for each of the TRX modules installed in the test set. This package constitutes your test set's calibration data.




You may choose to physically attach calibration stickers to your E6640A test set (recommended), or, as many customers do, you may choose to use the modules' electronic calibration information to identify the calibration status of any module at any time. The use of electronic calibration information is of benefit to those customers that may want to move modules from one instrument to another: the calibration date always follows the module.

To summarize: calibration of the E6640A is not done at the instrument level, but an E6640A is fully calibrated when all of the modules (Reference module and TRX modules) are calibrated. At initial shipment, this information is provided in the yellow envelope supplied with each unit.

Desktop Icons

Three desktop icons provide access to utilities which support the E6640A's measurement firmware.



Configure Applications

TRX Configurator

Launch XSA

TRX Configurator

The TRX Configurator tool shows which of the four possible TRXs are installed in the PXIe rack, and identifies them by SCPI address. At a minimum, TRX1 must be installed, and cannot be disabled using this tool. However, the other TRXs, if installed, can be enabled or disabled using the checkbox in the right column.



The **UpdateFPGA** button is used when installing a new TRX into an E6640A which was purchased without it, to ensure software compatibility.


Launch XSA

On bootup, the E6640A opens an XSA application window for each TRX which is (1) installed and (2) enabled in the TRX Configurator described above.

If you have closed any of the XSA application windows, you can reopen them without rebooting, by clicking the Launch XSA icon. This will only open a window for a given TRX if it is enabled in the TRX Configurator.

Configure Applications

Click the Configure Applications icon to launch this utility, which allows you to specify which of the E6640A's measurement applications will be preloaded into memory at startup. Preloading the most frequently-used applications is a convenience (however, selecting fewer of them for preload shortens the startup time and uses less memory).



A selector at the lower left of the window allows you specify which application will be running when the E6640A starts up.

Port Configuration


Any of the TRXs four RF ports can be configured as an input port (accepting a signal to be analyzed) or as an output port (providing an RF stimulus). Two of the ports (RFIO 1 and RFIO 2) can also be used in full-duplex mode (that is, with a single port used for both input and output simultaneously). The various possible configurations are described here.

NOTE

If the test set was purchased with Option 4FD, all four RF ports have full duplex capability, and port configuration is handled differently. For details, see **“Option 4FD: Full Duplex Port Setup”** on page 59 and **“Option 4FD: Separate Input and Output Port Setup”** on page 60.

Full Duplex Port Setup (RFIO 1 and RFIO 2)

Ports RFIO 1 and RFIO 2 connect to a splitter/combiner rather than a switch, so they can be used in full duplex mode.



To use the RFIO 1 port as the input and output, set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RFIO 1

Command:


```
:FEED:RF:PORT RFIO1
```

Then set up the RF Output using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > RF Output Port > RFIO 1

Command:

```
:FEED:RF:PORT:OUTP RFIO1
```



To use the RFIO 2 port as the input and output, set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RFIO 2

Command:

`:FEED:RF:PORT RFIO2`

Then set up the RF Output using this menu setting or its equivalent command:


Input/Output > More > RF Output & Test Set Config > RF Output Port > RFIO 2

Command:

`:FEED:RF:PORT:OUTP RFIO2`

Separate Input and Output Port Setup (RF3 I/O and RF4 I/O)

Ports RF3 I/O and RF4 I/O connect to a switch rather than a splitter/combiner, which means that one of these ports must be used as either an input or an output, not both, at any given time. (However, because there is less RF path loss when a switch is used, these ports are preferred in situations where it is important to maximize RF power to the analyzer or from the source.)



To define RF3 I/O as an input and RF4 I/O as an output, use the following steps.

Define RF3 I/O as an input, using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > Half Duplex Config > RF Input Port > RF3 I/O

Command:

```
:HDUPlex:PORT:INPut RFIO3
```

Define RF4 I/O as an output, using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > Half Duplex Config > RF Output Port > RF4 I/O

Command:

```
[ :SENSe ] :HDUPlex:PORT:OUTPut RFIO4
```

NOTE

This does not complete the port setup; two more settings are required!

Set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RF Input

Command:


```
:FEED:RF:PORT RFIN
```

Set up the RF Output using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > RF Output Port > RF Output

Command:

`:FEED:RF:PORT:OUTP RFO`



To define RF4 I|O as an input and RF3 I|O as an output, use the following steps.

Define RF4 I|O as an input, using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > Half Duplex Config > RF Input Port > RF4 I|O

Command:

`:HDUPlex:PORT:INPut RFIO4`

Define RF3 I|O as an output, using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > Half Duplex Config > RF Output Port > RF3 I|O

Command:

`[:SENSe] :HDUPlex:PORT:OUTPut RFIO3`

NOTE

This does not complete the port setup; two more settings are required!

Set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RF Input

Command:

`:FEED:RF:PORT RFIN`

3 Operating Tasks

Set up the RF Output using this menu setting or its equivalent command:


Input/Output > More > RF Output & Test Set Config > RF Output Port > RF Output

Command:

```
:FEED:RF:PORT:OUTP RFO
```


Separate Input and Output Port Setup (RFIO 1 and RFIO 2)

Although Ports RFIO 1 and RFIO 2 are designed to operate as full-duplex ports, they can function as separate inputs and outputs.



To use the RFIO 1 port as the input and the RFIO 2 port as the output, set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RFIO 1

Command:


```
:FEED:RF:PORT RFIO1
```

Then set up the RF Output using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > RF Output Port > RFIO 2

Command:

```
:FEED:RF:PORT:OUTP RFIO2
```



To use the RFIO 1 port as the input and the RFIO 2 port as the output, set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RFIO 2

Command:

`:FEED:RF:PORT RFIO2`

Then set up the RF Output using this menu setting or its equivalent command:


Input/Output > More > RF Output & Test Set Config > RF Output Port > RFIO 1

Command:

`:FEED:RF:PORT:OUTP RFIO1`

Option 4FD: Full Duplex Port Setup

If the test set was purchased with Option 4FD, then each of the IO ports connects to a splitter/combiner rather than a switch, and it can be used in full duplex mode, as a simultaneous input and output. In the example illustrated below, the RFIO 2 port is used in this way. (Other ports could be selected simply by substituting different port names in the instructions provided below.)



To use the RFIO 2 port as the input and output, set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RFIO 2

Command:

```
:FEED:RF:PORT RFIO2
```

Then set up the RF Output using this menu setting or its equivalent command:


Input/Output > More > RF Output & Test Set Config > RF Output Port > RFIO 2

Command:

```
:FEED:RF:PORT:OUTP RFIO2
```

Option 4FD: Separate Input and Output Port Setup

Although, with Option 4FD, all of the RFIO ports are designed to operate as full-duplex ports, they can function as separate inputs and outputs. In this example, the RFIO 2 port is used as an input, and the RFIO 3 port as an output. (Other port combinations could be selected simply by substituting different port names in the instructions provided below.)



To use the RFIO 2 port as the input and the RFIO 3 port as the output, set up the RF Input using this menu setting or its equivalent command:

Input/Output > RF Input > RF Input Port > RFIO 2

Command:

```
:FEED:RF:PORT RFIO2
```

Then set up RFIO 3 as the RF Output, using this menu setting or its equivalent command:

Input/Output > More > RF Output & Test Set Config > RF Output Port > RFIO 3

Command:

```
:FEED:RF:PORT:OUTP RFIO3
```

LAN Address Configuration

The E6640A supports both dynamic and static assignment of its IP address, using ports LAN 1 and LAN 2 respectively.

The LAN 1 port is designed for dynamic IP addressing, using the Dynamic Host Configuration Protocol. If your site network supports DHCP, the E6640A will be assigned an IP address automatically when it is connected to the LAN. Once the address is assigned, it is listed, along with the computer name, under **System > Show > System**. The address or computer name can be used to find the test set on the LAN. However, to see the **System > Show > System** screen and find this information, it is necessary either to connect a monitor to the front panel monitor port, or else to use the LAN 2 port to connect test set to a PC directly.

Revision Date	1/7/2014 2:00:46 PM
Computer System	Windows 7 , Service Pack 1
Computer Name	A-E6640A-01111
IP Address	148.5.97.84
IPv6 Address	2001:0:5ef5:79fd:380c:3239:6bfa:9eab
Link-Local IPv6 Address	fe80::380c:3239:6bfa:9eab%3
Host ID	PCSERNO,GA47677499
mDNS Enabled	Yes
mDNS Host Name	A-E6640A-0111-2
mDNS Service Name	Agilent PCSERNO Wireless Connectivity Test Set - SC29...

Using the LAN 1 port to connect the test set directly to the public LAN is potentially insecure, because the test set does not provide anti-virus protection. Connecting the test set to the public LAN by way of a PC with antivirus protection is the preferred solution.

The LAN 2 port is designed to use a private static IP address, 192.168.1.2 (changing this address is possible but not recommended). This port can be used for direct connection to a PC.

To find the test set on the LAN, you will need to know its IP address or computer name. If you cannot easily obtain a monitor view of **System > Show > System**, you can determine the computer name using the instrument serial number. The computer name is in the format **A-E6640A-nnnnn**, ending in the last five digits of the serial number.

Switched MIMO

For WLAN modulation accuracy measurements, a form of MIMO measurement can be made with a single TRX, using a time-division multiplexing process. The same source/analyzer pair within the TRX switches between the antenna ports of the Device Under Test.

In contrast to true MIMO functionality, which is available in WLAN mode but not in Sequence Analyzer mode, the switched MIMO feature is a WLAN measurement which is available *only* in Sequence Analyzer mode.


Constraints on Radio Standards

Only certain radio standards support MIMO. Under **Mode Setup > Radio Std**, select **802.11n** or **802.11ac**, but not **802.11a/b/g**.


Also, if you select **802.11ac**, do not select **(80+80 MHz)** on the menu for that choice.

MIMO Port Configuration

For 2x2 MIMO, the RFIO 1 port (a full-duplex port) captures the data stream from the first antenna port of the DUT. The other full-duplex port, RFIO 2, captures the data stream from the second antenna port of the DUT.



For 3x3 MIMO, the RFIO 1 port (a full-duplex port) captures the data stream from the first antenna port of the DUT. The other full-duplex port, RFIO 2, captures the data stream from the second antenna port of the DUT. One of the half-duplex ports, RF3 I|O or RF4 I|O, captures the data stream from the third antenna port of the DUT.



Source Sync Control

Switched MIMO is a measurement process only; it is not necessary to use Source Sync Control (required for true MIMO) in the case of switched MIMO.

RF Input Port Mode

The RF Input Port Mode must be set to List, so that the sequence can change the input port path from one acquisition to the next. In the **Meas Setup > RF Input Port Mode** menu, select **List**. The equivalent SCPI commands is:

```
[ :SENSe ] :LSEQuencer :PORT :INPut :MODE LIST
```

Half Duplex RF Input

In the case of 3x3 switched MIMO, one of the half-duplex ports (RF3 I|O or RF4 I|O) must be designated as an RF input port for the third input. In the **Input/Output > RF Output & Test Set Config > Half Duplex Config > RF Input** menu, select **RFIO3** or **RFIO4**, so that the RF3 IO port or the RF4 I|O port is defined as the input port. The equivalent SCPI command is:

```
[ :SENSe ] :HDUPlex :PORT :INPut RFIO3 | RFIO4
```

Acquisition Triggers

Each of the acquisitions which make up a sequence used for switched MIMO testing needs to have the same trigger settings as the others.

Acquisition Integration

For 2x2 switched MIMO in Sequence Analyzer mode, two separate acquisitions are used for the two separate ports, and must be integrated. Set up acquisition integration as follows: in the **Meas Setup > Acquisition Setup > Acquisition Integration** menu, select **Primary** for the first acquisition and **Switch MIMO** for the second. The equivalent SCPI commands are:

```
[ :SENSe ] :LSEQuencer:ACQuire{1:512}:SETup:INTEgration PRIMary  
[ :SENSe ] :LSEQuencer:ACQuire{1:512}:SETup:INTEgration SMIMo
```

For 3x3 switched MIMO in Sequence Analyzer mode, three separate acquisitions are used for the three separate ports, and must be integrated. Set up acquisition integration as follows: in the **Meas Setup > Acquisition Setup > Acquisition Integration** menu, select **Primary** for the first acquisition and **Switch MIMO** for the second and third. The equivalent SCPI commands are:

```
[ :SENSe ] :LSEQuencer:ACQuire{1:512}:SETup:INTEgration PRIMary  
[ :SENSe ] :LSEQuencer:ACQuire{1:512}:SETup:INTEgration SMIMo  
[ :SENSe ] :LSEQuencer:ACQuire{1:512}:SETup:INTEgration SMIMo
```

NOTE

The integration parameter can also be set within the complex command which is used to set the Acquisition RF Input Port (see **"Acquisition RF Input Port"** on page 65; the integration parameter is the one which comes immediately before the Acquisition RF Input Port parameter.

Acquisition RF Input Port

An RF input path must be selected for each acquisition. In the **Meas Setup > Acquisition Setup > Acquisition RF Input Port** menu, select **RFIO1**, **RFIO2**, or **RF Input**, as needed for each acquisition. The **RF Input** selection, which is used for the third port in 3x3 switched MIMO, means that whichever of the half-duplex ports (RF3 I/O or RF4 I/O) has been designated as the RF input is used (see **“Half Duplex RF Input”** on page 63). In SCPI, the acquisition RF input port is controlled by the last parameter in the complex command which configures the acquisition (the values for the relevant parameter are italicized here):

```
[ :SENSE ] :LSEQuencer:ACQuire [ 1 ] | 2 | . . . | 4 . . 512 :SETup
NONE | GSM | EDGE | WCDMA | CDMA2K | CDMA1XEV | LTE | LTETDD |
TDSCDMA | BTooth | WLAN,
NONE | PGSM | EGSM | RGSM | DCS1800 | PCS1900 | TGSM810 | GSM450
| GSM480 | GSM850 | GSM700 | BANDI | BANDII | BANDIII | BANDIV |
BANDV | BANDVI | BANDVII | BANDVIII | BANDIX | BANDX | BANDXI |
BANDXII | BANDXIII | BANDXIV | USCELL | USPCS | JAPAN | KOREAN |
NMT | IMT2K | UPPER | SECOND | PAMR400 | PAMR800 | IMTEXT |
PCS1DOT9G | AWS | US2DOT5G | PUBLIC | LOWER | BAND1 | BAND2 |
BAND3 | BAND4 | BAND5 | BAND6 | BAND7 | BAND8 | BAND9 | BAND10 |
BAND11 | BAND12 | BAND13 | BAND14 | BAND17 | BAND18 | BAND19 |
BAND20 | BAND21 | BAND24 | BAND25 | BAND26 | BAND27 | BAND28 |
BAND31 | BAND33 | BAND34 | BAND35 | BAND36 | BAND37 | BAND38 |
BAND39 | BAND40 | BAND41 | BAND42 | BAND43 | BAND44 | BANDA |
BANDB | BANDC | BANDD | BANDE | BANDF,
BTS | MS,
<real>,
<integer>,
<ampl>,
LOW | ZERO | HIGH,
<time>,
<time>,
IMMediate | VIDEo | Video2 | INTernal | EXT1 | EXT2,
<amptd>,
<time>,
NONE | INT/,
<relative ampl>,
OFF | RFIO0 | RFIO1 | RFIO2 | RFIO3 | RFIO4 | RFIO5 | RFIO6 |
RFIO7,
ON | OFF,
NORMal | PRIMary | LOWer | UPPer | RANGE | SMIMo,
RFIO1 / RFIO2 / RFIN
```

NOTE

The parameter which comes immediately before the Acquisition RF Input Port parameter (italicized above) is the Acquisition Integration parameter, and can be set to PRIMary or SMIMo as needed (see **“Acquisition Integration”** on page 64).

Switched MIMO Modulation Accuracy measurement SCPI commands

The following commands are used to set up the MIMO Modulation Accuracy (“EVM MIMO”) measurement.

Table 8-1

Command	Description	Notes
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:FERRor <real>	Freq Error	Sets a frequency error limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:CLKerror <real>	Clock Error	Sets a clock error limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:CFLeakage <rel_ampl>	Center Freq Leakage	Sets Center Freq Leakage limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:CFLeakage <rel_ampl>	BPSK-1/2 RMS EVM	Sets BPSK coding rate 1/2 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:BPSK:R1B2 <rel_ampl>	QPSK-1/2 RMS EVM	Sets QPSK coding rate 1/2 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QPSK:R1B2 <rel_ampl>	QPSK-3/4 RMS EVM	Sets QPSK coding rate 3/4 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QPSK:R3B4 <rel_ampl>	16QAM-1/2 RMS EVM	Sets 16QAM coding rate 1/2 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QA16:R1B2 <rel_ampl>	16QAM-3/4 RMS EVM	Sets 16QAM coding rate 3/4 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QA16:R3B4 <rel_ampl>	64QAM-2/3 RMS EVM	Sets 64QAM coding rate 2/3 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QA64:R2B3 <rel_ampl>	64QAM-3/4 RMS EVM	Sets 64QAM coding rate 3/4 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QA64:R5B6 <rel_ampl>	64QAM-5/6 RMS EVM	Sets 64QAM coding rate 5/6 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QA256:R3B4 <rel_ampl>	256QAM-3/4 RMS EVM	Sets 256QAM coding rate 3/4 RMS EVM limit
:CALCulate:LSEQuencer:WLAN:EVMMimo:LIMit:RMS:QA256:R5B6 <rel_ampl>	256QAM-5/6 RMS EVM	Sets 256QAM coding rate 5/6 RMS EVM limit

Table 8-1

Command	Description	Notes
:CALCulate:LSEQuencer:WLAN:EVMMimo:EQUalizer: TMODe SEQuence SDATa	Equalizer Training	Specifies how the equalizer is initialized or trained.
:CALCulate:LSEQuencer:WLAN:EVMMimo:PILot:TRACk: AMPLitude OFF ON 0 1	Track Amplitude	Specifies whether analyzer tracks amplitude changes in the pilot subcarriers.
CALCulate:LSEQuencer:WLAN:EVMMimo:PILot:TRACk: PHASe OFF ON 0 1	Track Phase	Specifies whether analyzer tracks phase changes in the pilot subcarriers.
:CALCulate:LSEQuencer:WLAN:EVMMimo:PILot:TRACk: TIMing OFF ON 0 1	Track Timing	Specifies whether analyzer tracks timing changes in the pilot subcarriers.
[:SENSe]:LSEQuencer:WLAN:EVMMimo: MTYPE DMAP FOURier	Matrix Type	Specifies MIMO matrix type.

Switched MIMO Results

The results available for a WLAN MIMO Modulation Accuracy measurement are as follows (the results for Stream 3 are returned as -1.0 (not tested) in the case of 2x2 MIMO):

Index	Result Parameter
0	Overall Pass/Fail Result 0:Pass, 1:Fail, -1:Not tested
1	Stream 1 RMS EVM pass/fail result (1.0 = fail, 0.0 = pass, -1.0 = Not tested)
2	Stream 1 RMS EVM (dB)
3	Stream 1 Peak EVM (dB)
4	Stream 1 Pilot EVM (dB)
5	Stream 1 Data EVM (dB)
6	Stream 1 Chips Peak EVM (%)
7-18	7-12: Stream 2 (as in Index 1-6 above). 13-18: Stream 3 (as in Index 1-6 above)
19-24	Reserved
25	Channel 1 RMS EVM pass/fail result (1.0 = fail, 0.0 = pass, -1.0 = Not tested)
26	Channel 1 Frequency Error pass/fail result (1.0 = fail, 0.0 = pass, -1.0 = Not tested)
27	Channel 1 Symbol Clock Error pass/fail result (1.0 = fail, 0.0 = pass, -1.0 = Not tested)
28	Channel 1 IQ Offset pass/fail result (1.0 = fail, 0.0 = pass, -1.0 = Not tested)
29	Channel 1 RMS EVM (dB)
30	Channel 1 Peak EVM (dB)
31	Channel 1 Pilot EVM (dB)
32	Channel 1 Data EVM (dB)
33	Channel 1 Frequency Error (Hz)
34	Channel 1 Symbol Clock Error (ppm)
35	Channel 1 IQ Gain Imbalance (dB)
36	Channel 1 IQ Offset (dB)
37	Channel 1 IQ Quadrature Error (deg)
38	Channel 1 IQ Timing Skew (pps)
39	Channel 1 Cross Power (dB)
40	Channel 1 Sync Corr

Index	Result Parameter
41-72	41-56: Channel 2 (as in Index 25-40 above). 57-72: Channel 3 (as in Index 25-40 above)
73-88	Reserved


True MIMO

In WLAN mode, True MIMO (simultaneous rather than switched MIMO) is supported for the MIMO Modulation Analysis measurement. (This functionality is not available in Sequence Analyzer mode.) Setting up this type of measurement requires configuration of two separate TRXs.

For MIMO Modulation Analysis, two or more measurement data streams are captured simultaneously, by two or more separate TRXs in the EXM test set. In the EXM, the TRXs share a common frequency reference. They also share a common IP address, but have separate socket numbers. In MIMO testing, a single TRX is designated the primary, and it controls the other TRX or TRXs used.


TRX Configurations

Here is an example of 2x2 MIMO:




In the example setup shown here, TRX 1 runs the MIMO Modulation Analysis measurement, and is also configured to control TRX 2, which collects the second data stream. The primary TRX controls the measurement settings for the secondary TRX, and also processes the measurement data collected by the secondary. (The display window for the secondary will display a mostly-blank screen, with a message indicating that it is in use by the primary.)

Here is an example of 3x3 MIMO:



In the example setup shown here, TRX 1 runs the MIMO Modulation Analysis measurement, and is also configured to control TRX 2 and 3, which collect the second and third data streams.

Here is an example of 4x4 MIMO:



In the example setup shown here, TRX 1 runs the MIMO Modulation Analysis measurement, and is also configured to control TRX 2 to 4, which collect the second, third, and fourth data streams.

Source Settings for MIMO

Unlike measurement settings, which are made only to the primary TRX and applied to the others, source settings are made separately for each TRX, and cannot be made from the primary TRX alone. However, when the measurement is run, the primary TRX communicates with the other TRXs to ensure that they begin playing ARB files simultaneously.

To run the sources synchronously:

1. Make all source settings (such as frequency, amplitude, and modulation) individually on each TRX which will be used in the MIMO test.
2. Select the source synchronization type on each of the TRXs which will be used in the MIMO test:
 - For 2x2 MIMO, set up source synchronization as follows: in the **Source > Source Sync Control** menu, select **2x2**. Do this for both TRXs. The equivalent SCPI command is:
:SOURCE:SYNC:CONFIG TWO
 - For 3x3 MIMO, set up source synchronization as follows: in the **Source > Source Sync Control** menu, select **3x3**. Do this for all three TRXs. The equivalent SCPI command is:
:SOURCE:SYNC:CONFIG THREE
 - For 4x4 MIMO, set up source synchronization as follows: in the **Source > Source Sync Control** menu, select **4x4**. Do this for all four TRXs. The equivalent SCPI command is:
:SOURCE:SYNC:CONFIG FOUR
3. In the window for the primary TRX, activate synchronization in the **Source > Source Sync Control** menu by pressing **Sync Start**. At this point, the other TRX sources are synchronized to the primary and will play their arb files simultaneously with it. The equivalent SCPI command is:
:SOURCE:SYNC:START
You can stop synchronization in the **Source > Source Sync Control** menu by pressing **Sync Stop**. The equivalent SCPI command is:
:SOURCE:SYNC:STOP

NOTE

If it is necessary to change source settings after synchronization is started, the synchronization has to be stopped. The changes can then be made, after which synchronization can be started again (as described above). While the primary is controlling the other TRXs, the front panel interface for those TRXs is not available; either use SCPI commands to change the settings of those sources, or temporarily release the other TRXs from control by the primary, as described under **"Remote Analyzer Configuration"** on page 73.

Constraints on TRX Selection

Not every TRX can be designated the primary for purposes of MIMO testing. The permitted configurations are shown in the table below:

Table 8-2

MIMO Configuration	Primary TRX	Secondary TRXs
2x2	TRX 1	TRX 2, Socket 5125
2x2 (alternative setup)	TRX 3	TRX 4, Socket 5325
3x3	TRX 1	TRX 2, Socket 5125 TRX 3, Socket 5225
4x4	TRX 1	TRX 2, Socket 5125 TRX 3, Socket 5225 TRX 4, Socket 5325

Constraints on Radio Standards

Only certain radio standards support MIMO. Under **Mode Setup > Radio Std**, select **802.11n** or **802.11ac**, but not **802.11a/b/g**.

Also, if you select **802.11ac**, do not select **(80+80 MHz)** on the menu for that choice.

MIMO Type

For the MIMO Modulation Analysis measurement, go to the **Meas Setup > MIMO Type** menu and select **2x2**, **3x3**, or **4x4**. (This setting determines whether the display shows two, three, or four constellation diagrams for the measurement.)

The equivalent SCPI command is:

```
[ :SENSE ] :EVMMimo:TYPE M2M2 | M3M3 | M4M4
```

Remote Analyzer Configuration


This feature is used to coordinate TRXs within the test set; it is named this way because its original purpose was to coordinate the test set with external equipment.

NOTE

The 2x2 or 3x3 MIMO configuration that is set up here needs to match the **MIMO Type** selection, as described above. Error messages will be displayed until the two are in agreement.

1. From the primary TRX window (TRX 1 in this example), select **Mode Setup > Global Settings > Remote Analyzer Config**.
2. A list is displayed which includes any TRX which you have previously used (unless you have deleted it) and any added while in the **Add Analyzer to List** menu. However, the list may be blank, or may not include the desired TRX.
3. To add a TRX to the list, press **Add Analyzer to List**.

4. Click **IP Address**, use the numeric keys to enter an IP address for the test set, and press **Enter**. In this example, the test set's localhost address is used (for the EXM test set, the localhost address has a fixed value of 127.0.0.1).
5. Click **SCPI Socket Port**, use the numeric keys to enter the socket number of the secondary TRX, and press **Enter**. In this example socket number 5125 (TRX 2) is used.
6. Press **Add**. The secondary TRX is now displayed in the list.
7. Select the desired TRX from the displayed list (use the up and down arrow keys to move the highlighted selection to the desired TRX and then press **Select Highlighted Analyzer**).
8. The secondary TRX is now being run from the present TRX. The window for the secondary TRX displays a "This instrument is in use..." message. If you need to release it from control, go to the primary TRX and select **Mode Setup > Global Settings > Remote Analyzer Config > Release Current Analyzer**, or send to the primary the equivalent SCPI command:
:SYST:COMM:ANAL:REL
9. The equivalent SCPI command for the steps described above is as follows (modify the address string "127.0.0.1:5125" as needed):
:SYST:COMM:ANAL:SEL "127.0.0.1:5125"



The procedure given above is for 2x2 MIMO; the process is the same for 3x3 or 4x4 MIMO, except that more TRXs are added.

The command equivalent for 3x3 MIMO would be:

```
:SYST:COMM:ANAL:SEL "127.0.0.1:5125"  
:SYST:COMM:ANAL:SEL "127.0.0.1:5225"
```

The command equivalent for 4x4 MIMO would be:

```
:SYST:COMM:ANAL:SEL "127.0.0.1:5125"  
:SYST:COMM:ANAL:SEL "127.0.0.1:5225"  
:SYST:COMM:ANAL:SEL "127.0.0.1:5325"
```

MIMO Modulation Analysis measurement SCPI commands

The following commands are used to set up the MIMO Modulation Analysis measurement.

Table 8-3

Command	Description	Notes
<code>[:SENSe] :EVMMimo:MTYPE DMAP FOURier</code>	Matrix Type	Sets the type as DMAP or Fourier
<code>[:SENSe] :EVMMimo:TYPe M2M2 M3M3 M4M4</code>	MIMO Type	Sets the type as 2x2, 3x3, or 4x4
<code>[:SENSe] :EVMMimo:RESult:STReam <integer></code>	MIMO Result Stream Index	Specifies the result stream by number.
<code>[:SENSe] :EVMMimo:RESult:CHANnel <integer></code>	MIMO Result Channel Index	Specifies the result channel by number.

MIMO Modulation Analysis Results

The results available for a WLAN MIMO Modulation Analysis measurement are as follows (the results for Stream 3 are returned as -1.0 (not tested) in the case of 2x2 MIMO):

Index	Result Parameter
0	<p>Returns unprocessed I/Q trace data of Capture Interval, as a series of trace point values for Channel m. Channel index m can be set and queried through EVMM:RESult:CHAN command.</p> <p>The I values are listed first in each pair, using the 0 through even-indexed values. The Q values are the odd-indexed values.</p>
1	<p>Returns the comma-separated EVM results by stream. The EVM scalar results for each stream are in the following order:</p> <ol style="list-style-type: none"> 1. RMS EVM Max (dB) 2. RMS EVM Avg (dB) 3. Peak EVM Max (dB) 4. Peak EVM Avg (dB) 5. Pilot EVM Max (dB) 6. Pilot EVM Avg (dB) 7. Data EVM Max (dB) 8. Data EVM Avg (dB) 9. CPE EVM Max (dB) 10. CPE EVM Avg (dB) <p>When MIMO Type is 2x2, this result gives stream 1 EVM scalar results first, then stream 2 EVM scalar results. When MIMO Type is 3x3, stream 1, 2 and 3 EVM scalar results are provided in order. When MIMO Type is 4x4, returns 40 comma-separated EVM scalar results are provided in stream order.</p>
2	<p>Returns a value of the pass/fail (0.0 = passed, or 1.0 = failed) results determined by testing the result of RMS EVM. Only when all the stream RMS EVM Average results are passed, will this value be passed.</p>
3	<p>Returns series of floating point numbers that alternately represent I and Q pairs of the corrected measured trace for STRM m. Stream index m can be set and queried through EVMM:RESult:STReam command.</p> <p>The magnitude of each I and Q pair are normalized to 1.0, when the IQ Normalize is turned on. The first number is the in-phase (I) sample and the second is the quadrature-phase (Q) sample.</p>
4	<p>Returns series of floating point numbers (in dB) that represent each sample in the Stream EVM vs. Carrier trace of Capture Interval for STRM m. Stream index m can be set and queried through EVMM:RESult:STReam command.</p>
5	<p>Returns series of floating point numbers (in dB) that represent each sample in the Stream EVM vs. Symbol trace of Capture Interval for STRM m. The first number is the symbol 0 decision point. Stream index m can be set and queried through EVMM:RESult:STReam command.</p>

Index	Result Parameter
6	Returns series of floating point numbers that represent each sample in the Channel Response for STRM m CH k. Stream index m can be set and queried through EVMM:RESult:STReam command. Channel index k can be set and queried through EVMM:RESult:CHAN command.
7	<p>Returns comma-separated EVM scalar results by channel. The EVM scalar results for each channel are in the following order:</p> <ol style="list-style-type: none"> 1. RMS EVM Max (dB) 2. RMS EVM Avg (dB) 3. Peak EVM Max (dB) 4. Peak EVM Avg (dB) 5. Pilot EVM Max (dB) 6. Pilot EVM Avg (dB) 7. Data EVM Max (dB) 8. Data EVM Avg (dB) 9. Freq Error Max (Hz) 10. Freq Error Avg (Hz) 11. Freq Error Ppm Max (ppm) 12. Freq Error Ppm Avg (ppm) 13. Sym Clock Error Max (ppm) 14. Sym Click Error Avg (ppm) 15. IQ Gain Imb Max (dB) 16. IQ Gain Imb Avg (dB) 17. IQ Offset Max (dB) 18. IQ Offset Avg (dB) 19. IQ Quad Error Max (degree) 20. IQ Quad Error Avg (degree) 21. IQ Timing Skew Max (s) 22. IQ Timing Skew Avg (s) 23. Cross Pwr Max (dB) 24. Cross Pwr Avg (db) 25. Sync Corr Max 26. Sync Corr Avg 27. Avg Burst Power Max (dBm) 28. Avg Burst Power Avg (dBm) 29. Peak Burst Power Max (dBm) 30. Peak Burst Power Avg (dBm) <p>When MIMO Type is 2x2, 60 comma-separated scalar results are returned; When MIMO Type is 3x3, 90 comma-separated scalar results are returned; When MIMO Type is 4x4, 120 comma-separated scalar results are returned.</p>
8	Return the Channel Matrix results.
9	Return the decode bits of HT-SIG Signal.
10	Returns series of floating point numbers (in dB) that represent each sample in the Equalizer Impulse Response for STRM m CH k. Stream index m can be set and queried through EVMM:RESult:STReam command. Channel index k can be set and queried through EVMM:RESult:CHAN command.

Tx Beam Forming

If license V9077B-KFP is installed, Tx Beam-Forming results are available from the Modulation Analysis measurement in WLAN mode.

Tx Beam Forming Commands

Two commands are added for Tx Beam Forming, as described in the table below. The first command sets up the measurement's EVM demodulation algorithm to provide EVM results or Tx Beam Forming Results. The second command is used to specify the channel count of the measured MIMO signal.

Table 8-4

Command	Description
[:SENSe]:EVM:BFORMing:MTYPE EVM TXBFcal	<p>This command is only available when license V9077B-KFP is installed; otherwise, “-113, Undefined Header” error will be reported.</p> <p>This parameter takes effect when the Radio Std is 802.11n(20MHz), 802.11n (40MHz), 802.11ac (20MHz), 802.11ac (40MHz), 802.11ac (80MHz) and 802.11ac (160MHz)</p> <p>When EVM is selected, the result index 1~20 are calculated with meaningful values for EVM measurement, and result 21 and 22 results are -999.0;</p> <p>When TXBFcal is selected, the result index 1~20 values will not be calculated; Instead, result 21 and 22 are calculated and returned</p>
[:SENSe]:EVM:ACCHannel <integer> <real>	<p>For Radio Std 802.11n (20MHz) and 802.11n (40MHz), Tx Beam Forming with channel count <=4 is supported;</p> <p>For Radio Std 802.11ac (20MHz), 802.11ac (40MHz), 802.11ac (80MHz), and 802.11ac (160MHz), Tx Beam Forming with channel count <=8 is supported</p>

Tx Beam Forming Results

The Tx Beam Forming results are added to the Modulation Analysis measurement results as Index 21 and Index 22, as described in the table below.

Index	Result Parameter
21	<p>Tx Beam Forming Calibrated Angle Error returns a series of floating point numbers (in rad) that represent the calibrated angle error vectors on channels, when</p> <ul style="list-style-type: none"> - Radio Std is 802.11n (20 MHz), or 802.11n (40 MHz), and MIMO Tx Channel Count <= 4 - Radio Std is 802.11ac(20MHz), 802.11ac(40MHz), 802.11ac(80MHz), or 802.11ac(160MHz), and MIMO Tx Channel Count <= 8 - MIMO Usage Type is TXBForming (available when license V9077B-KFP is installed) <p>Otherwise, returns -999.0.</p> <p>The first channel angle vector is used as reference; the angle error vectors are listed in order of channels. For the 20MHz BW signals, each channel's angle error vector contains 64 values; For the 40MHz BW signals, each channels angle error vector contains 128 values; For the 80MHz BW signals, each channels angle error vector contains 256 values, For 160MHz BW signals, each channels angle error vector contains 512 values.</p>
22	<p>Tx Beam Forming Gain returns gain after the calibrated angle error vector are compensated for the MIMO input signal, when</p> <ul style="list-style-type: none"> - Radio Std is 802.11n (20 MHz), or 802.11n (40 MHz), and MIMO Tx Channel Count <= 4 - Radio Std is 802.11ac(20MHz), 802.11ac(40MHz), 802.11ac(80MHz), or 802.11ac(160MHz), and MIMO Tx Channel Count <= 8 - MIMO Usage Type is TXBForming (available when license V9077B-KFP is installed) <p>Otherwise, returns -999.0.</p>

4 Test Set Operating System

This chapter describes the Microsoft Windows configuration and the settings used with the Keysight test set software. It includes information about changing some of the system settings. And it describes the Windows operating system configuration and the software installations that are present on the hard disk drive when the test set leaves the factory.

"Keysight Software Installed" on page 82

"Customer Installation of Software" on page 83

"User Accounts" on page 84

"Keysight Licensing Options" on page 86

"Licensing New Measurement Application Software - After Initial Purchase" on page 88

"Windows Configuration" on page 93

"Configuring Printers" on page 96

"Configuring LAN" on page 96

"Windows Security" on page 97

"System Maintenance" on page 101

"USB Connections" on page 102

"Hard Drive Partitioning and Use" on page 103

"Hard Drive Recovery Process" on page 104

"Updating the software" on page 106

Keysight Software Installed

Measurement Software Applications

A set of Keysight measurement applications for different modes (GSM/EDTE, cdma2000, Sequence Analyzer mode, etc.) is installed by default in the test set, but these must be licensed in order to be used. Whichever licenses were purchased with the instrument are installed at the factory (see **"Options and Licenses"** on page 21. Other licenses can be purchased and installed later (including transportable licenses). For more information, see: <http://www.keysight.com/find/e6640a>

Customer Installation of Software

3rd party software verified by Keysight

Keysight has verified that the following programs are compatible with the test set applications:

- Symantec AntiVirus™ Corporate Edition version 10
- MathWorks MATLAB

Installation of other 3rd party software

The E6640A platform is an open Windows environment, so you can install non-approved software on the test set. However, installation of non-approved software may affect test set performance. Keysight does not warrant the performance of the test set with non-approved software installed.

NOTE

Before installing any additional programs on the test set, you should exit the test set application.

Also, you must not remove any applications or programs that are installed on the test set when it is shipped from the factory.

If you install programs other than those that Keysight has tested, it could cause problems with the test set's applications. If this happens, you should try uninstalling the program that has caused the problem, or try changing the program's configuration. If this does not correct the problem, you may have to use the Keysight Recovery system to reinstall the test set's system software.

User Accounts

Administrator login

The Administrator account ships from the factory with the password “Keysight4u!”. Using the Administrator account, you can perform the following operations:

- Install software
- Configure network and printer access
- Access all files on the test set
- Add or change user accounts and passwords
- Change Firewall settings
- Change Windows settings (for example, using Device Manager)
- Change the time and date
- Run any application

NOTE

The E6640A was formerly a product of Agilent Technologies. If the front panel of the test set shows Agilent as the manufacturer, then “agilent4u” may have been set as the password for the Administrator account.

User login

The default user account that ships from the factory is “Instrument” with the password “measure4u”. This user is a member of the Power Users group. Using the Instrument account you can perform the following operations:

- Install software
- Configure network and printer access
- Access files on the test set that are accessible to the Power Users group
- Run applications that are accessible to the Power Users group

KeysightOnly user account

The test set contains a user account called “KeysightOnly” that can be used by Keysight’s customer support in the event that the Administrator password was changed and has since been lost or forgotten. You must not remove or modify the KeysightOnly account. (If the front panel of the test set shows Agilent Technologies as the manufacturer, the account may be called “AgilentOnly”.)

Keysight service user accounts

User accounts are defined for Keysight’s use if it is necessary to service the test set.

Customer creation of accounts

You can create additional user accounts and decide on the level of security granted to any new user accounts created. For example, the level of security can be assigned as administrator, power user, user, backup operators. User names are not case sensitive but passwords are case sensitive.

NOTE

For the test set software to operate, the user account executing the software must be assigned Administrator or Power User privileges. Otherwise, the test set software will not operate correctly.

It is Keysight's expectation that each user's My Documents folder is mapped to the D: drive. This is to avoid overwriting the user's data in the event the Keysight Recovery must be performed. Also, this facilitates convenient backup by copying the contents of the D: drive to external media. All users accounts created by the factory already have My Documents mapped to the D: drive. Please map all new users My Documents folders to the D: drive.

Keysight Licensing Options

Keysight test sets use three licensing types: Fixed Perpetual, Transportable Perpetual, and Trial. Fixed Perpetual licenses are also required to enable hardware options.

Fixed Perpetual

Fixed Perpetual licenses are the traditional license type (Fixed) with the same duration (Perpetual) that have been available for all features since the Keysight X-Series introductions. Fixed Perpetual licenses are identified by the “F” in the second character and “P” in the third character of the option designator:

Example: V9079B-1FP

A license key is instrument model and serial number dependent. You can only install the license key on the specific instrument for which it was created.

Transportable Perpetual

Transportable Perpetual licenses are an optional license type offering deployment duration which is not fixed to a specific instrument model and serial number. Transportable Perpetual licenses are identified in the product structure by a “T” in the second character and “P” in the third character of the option designator:

Example: V9079B-1TP

Transportable Perpetual licenses require a connection to the Keysight server only for managing the check-in/out of the license. The Keysight licensing server also provides for storage of unused licenses which have been transported off instruments but are awaiting assignment to new instruments. The server will limit the number of transports per 30 day period per application license to 10.

Unlike Fixed Perpetual licenses which are pre-installed at the factory with new instrument purchases, Transportable Perpetual licenses require redemption and installation of the license before the first use. This allows the user to determine on which instrument to initially install the application license.

Keysight recommends that instruments be at the same instrument software release to ensure the latest code is available on each instrument so that the user experience is identical between instruments. This is particularly important when transporting the license for a newly-released application which may only be available in the latest software release.

Time-Limited Licenses (Trial Licenses)

Trial licenses are available so that you may try applications before you buy the full applications. These licenses are time limited for a 14 day period, and are restricted to one trial for an individual application per instrument. The restriction is enforced through the license redemption process of the Keysight Software Manager (KSM) system.

Trial licenses are only available for measurement applications and not available for enabling hardware capability.

Trial licenses are not installed in the factory and no entitlement certificates are created for distribution with new instrument shipments. Trial licenses are not available for order, but are available from the Keysight Web site after completing a brief registration:

http://www.keysight.com/find/e6640a_trial

Trial Licenses are identified by the license designator “-TRL”.

Example: V9079B-TRL

If a product has multiple tiers (levels of functionality), the -TRL license enables all tiers for the duration of the license. If you have a base tier and want a trial license for a higher tier, you can install the -TRL license. When the duration expires the entitlement is reverted back to the base tier.

Licensing New Measurement Application Software – After Initial Purchase

Additional measurement application software can be ordered after your initial purchase of the E6640A test set. Software upgrades are provided in a kit that includes an option based Entitlement Certificate, a license agreement. The licenses are downloaded from the license Web site onto a storage device so they can be loaded into the instrument.

If you do not want to wait for your upgrade kit to arrive, licenses can be installed using any available USB storage device. If you choose to do this, we recommend that the latest version of the instrument software be installed. This ensures that the measurement application being licensed and activated is installed and is the most current version.

The latest revision of the software may be downloaded from:

<http://www.keysight.com/find/e6640a>

A license key is usually for one instrument model and serial number combination. The license key will only install itself on that instrument.

NOTE

No calibration is required after a test set application installation.

Installation procedure over USB

Step	Action	Notes
1 Redeem the Option Upgrade Entitlement Certificate	<ul style="list-style-type: none"> Follow the instructions on the Certificate 	After redeeming your Option Upgrade Entitlement Certificate you will receive an e-mail with an attached License File.
2 Save the license file	<ul style="list-style-type: none"> Save the .lic file to the root directory of a USB storage device 	
3 Load the license file	<ul style="list-style-type: none"> Connect the USB storage device to one of the test set USB ports. 	Windows will detect the new hardware and may display the configuration menu. The test set automatically loads the license file. (This may take a few minutes) Upon completion, the Keysight License Manager displays a "Successful License Installation" message.

NOTE

Alternatively the license file can be manually installed over USB or LAN by placing the license file in the following folder on the test set.

C:\Program Files\Keysight\licensing

Step	Action	Notes
4 Verify installation	Cycle the power on the signal test set. On the Virtual Front Panel press System , Show , System . Verify that the new application appears in the list.	The application will not be available for use until after the power has been cycled This displays the list of installed applications. If you require further assistance, please contact the Keysight support team. Online assistance: http://www.keysight.com/find/assist If you do not have access to the Internet, contact your local Keysight Technologies Sales and Service Office, or if in the United States, call 1-800-829-4444.

Transporting a License Between Test Sets

Transportable licenses can be identified by the letters "TP" in their option designator. For example, V9079B-2TP indicates the license is transportable and perpetual. To transport this license from one test set to another, Keysight recommends that both test sets be at the same instrument software release. This ensures that the user experience is identical between instruments.

As a minimum, the instrument software release in the test set that will receive the transportable license (the "target instrument") must at least be able to support the desired application.

The E6640A test sets support several ways of transporting licenses. The procedure below will focus on the most common procedure, where neither of the test sets has access to an internet connection. For this procedure to work, a PC with an internet connection is required

You will need the following:

- USB flash drive
- USB keyboard
- USB mouse

We will refer to the test set from which the transportable license will be removed as the "source instrument". We will refer to the test set which will receive the transportable license as the "target instrument".

Procedure for Transporting a License, Neither Test Set Connected to the Internet

Step	Action	Notes
1 Connect the USB devices to the source instrument	<ul style="list-style-type: none"> • Connect the USB flash drive, USB keyboard, and USB mouse to the USB ports on the source instrument. 	It will be necessary to use one or more of the rear panel ports to connect the USB devices.
2 Verify software version in each test set	<ul style="list-style-type: none"> • On the Virtual Front Panel for each test set, press System, Show, System and make note of the Instrument S/W Revision on each. 	Ideally, the revisions will be the same. But at least, both versions should be able to support the application whose license is being transported.
3 Obtain the Host ID from the Target Instrument	<ul style="list-style-type: none"> • On the target instrument, press System, Show, System, and make note of the Host ID. 	This information will be needed to issue the license for the target instrument. The Host ID is the model number, followed by a comma, followed by the serial number.
4 Start the License Manager on the Source Instrument	<ul style="list-style-type: none"> • On the source instrument, press System, More, Licensing... It may take a minute for the Keysight License Manager screen to be fully populated with all the installed licenses. 	

Step	Action	Notes
5	<p>Transport the Transportable License from Source Instrument</p> <p>a On the source instrument, locate the desired license to be transported and highlight it. Its option designator should include the letters "TP".</p> <p>b .Right-click on the mouse and select Delete.</p> <p>c Click Yes in the License Deletion Confirmation dialog box.</p> <p>d After a few seconds, a Transport License dialog box will appear. Click OK and save the *.url files on the USB flash drive, with a name such as "PhaseNoise_License.url"</p>	<p>Be sure you are selecting a transportable license and not a fixed perpetual license.</p> <p>Although you are "deleting" the transportable license, the ultimate action will be to transport it.</p> <p>The information in this dialog box will be used to issue a new license for the target instrument.</p>
6	<p>Get New License from ASM Transportation Web page</p> <p>a Insert USB flash drive into PC that is connected to the internet</p> <p>b On the PC, locate the *.url file on the USB flash drive that was saved in the previous step and click on it.</p> <p>c After a few seconds, the ASM Transportation Web page will appear. One field, the New Host ID will be empty.</p> <p>d Enter the Host ID of the target instrument (step 3 above) into the field marked New Host ID.</p> <p>e Click Submit.</p>	<p>The PC must have an internet connection and an available USB port.</p> <p>Most of the fields have been populated from the information in the *.url file.</p> <p>Be very careful when entering the New Host ID. A mistake made in entering the New Host ID will result in an invalid license being issued. Be sure there is a comma between the model number and serial number. There should be no spaces in the Host ID.</p>
7	<p>Save License File to USB Flash Drive</p> <p>a The ASM Transportation Web page will indicate that a license file is available and display two links to the license file. Right-click on the license file and select Save Target As...</p> <p>b Save the license file to the root level of the USB flash drive, keeping the ".lic" file extension.</p> <p>c Close the Save dialog and exit the ASM Transportation Web page.</p>	<p>Either link can be used to either display or save the license file.</p> <p>The license file must be saved to the root level of the USB flash drive for it to be recognized by the target instrument.</p>

Step	Action	Notes
8 Install License File in Target Instrument	<p data-bbox="493 338 948 495">a With the target instrument running, insert the USB flash drive into one of the front panel USB ports. After a few minutes, you should see a message saying "Successful License Installation"</p> <p data-bbox="493 596 948 625">b Cycle power on the target instrument.</p> <p data-bbox="493 726 948 814">c Once the test set has re-booted, the application should be ready to use on the target instrument.</p>	<p data-bbox="992 338 1425 527">The Keysight License Services running on the target instrument looks for *.lic files whenever it detects a USB device has been inserted. If the contents of the *.lic file are appropriate for the instrument, the license will be installed automatically.</p> <p data-bbox="992 548 1425 638">Newly-installed licenses are only recognized by the instrument software on power-up.</p>

Windows Configuration

The Windows settings have been optimized for the best measurement performance. Any modifications to these settings may degrade test set performance and measurement speed. In general, most Windows System settings (typically set through the Windows Control Panel) should not be modified. Those that can be safely modified are listed below.

CAUTION







To recover from problems caused by changing Windows systems settings, you may have to reinstall the Windows system and test set application using the Keysight Recovery process.


Settings that can be changed

You may change the following Windows settings or administrative tasks (available from the Windows Control Panel) to select your personal preferences.

NOTE







Before changing any Windows System settings, exit the test set application.

You may use this feature	To do this...
 Windows Update	Configure Microsoft Windows Automatic Updates.
 Action Center	Install and configure an Anti Virus program.
 User Accounts	Set up new test set user accounts. CAUTION Do not delete or modify the “KeysightOnly” or “AgilentOnly” user account.
 Network and Sharing Center	Add the test set to a network.
 Devices and Printers	Install and configure a printer.
 Date and Time	Set the time and date.

You may use this feature	To do this...
 System	Under Advanced System Settings, System Properties, Advanced, Performance: choose either “Let Windows choose what’s best for my computer” or “Adjust for best performance”. Leave all other settings unchanged.

Settings that must not be changed

Avoid changing any settings in this section. Changes to the following settings may degrade test set performance, screen displays, and measurement speed.

Do NOT use this feature	To do this...
 Power Options	Do not change Power Options.
 System	Do not modify System Properties, Hardware Tab settings. Do not modify System Properties, Advanced Tab settings (except as recommended under “Settings that can be changed” above).
 Fonts	Do not remove installed Fonts.
 Display	Do not change the following Display Settings: <ul style="list-style-type: none"> • Screen Saver settings (under “Personalization”) • Screen resolution, 1024 x 768 (under “Adjust Resolution”) • DPI setting (under “Set custom text size”)
 Region and Language	Do not change any settings under “Region and Language” or the instrument keyboard and display may not operate properly
 User Accounts	Do not delete or modify the “KeysightOnly” or “AgilentOnly” user account.

In addition, DO NOT:

- Add, delete, or modify hard-disk drive partitions.
- Delete or modify Agilent/Keysight registry entries.
- Change the contents of any directories containing the name “Agilent” or “Keysight”.
- Stop these services:

- The MSSQL\$CDF service or uninstall the “Microsoft SQL Server Desktop Engine”
- The IIS server or tamper with any virtual directories (or their contents) that came configured with the test set.

- Uninstall these libraries, interfaces, or programs:
 - The Agilent/Keysight I/O Libraries
 - The .NET Framework or any Hotfixes or Service Packs for the .NET Framework
 - The “Microsoft Visual J# .NET Redistributable Package 1.1”
 - Programs that begin with “Agilent” or “Keysight”
 - The Adobe Acrobat Reader

- Modify
 - The Agilent/Keysight I/O Library “GPIB27”, “GPIB28” interfaces shown as configured Instrument I/O in the Agilent/Keysight Connection Expert or I/O Config.

Configuring Printers

Printers are configured using the Microsoft Windows Control Panel. It is easily accessed from the Windows Start menu or from under the front panel **System** key. This setup process is most easily done using a USB mouse and an external keyboard.

When setting up a new printer, you may need to load the printer driver (unless you are using a network printer that your IT department has set up to include the driver). The manufacturer of the printer supplies the driver software and process. That may require that you attach an external USB disk drive. An alternative is to connect the test set to the LAN and download the driver from the printer manufacturer's internet site.

Configuring LAN

Hostname

The Computer Name, or hostname, is pre-configured from the factory. It must be a unique name so that it does not conflict with other equipment on your LAN. The pre-configured Computer Name is A-E6640A-xxxxx, where xxxxx represents the last 5 digits of the test set serial number.

To change the Computer Name consult the Microsoft Windows Help and Support Center.

IP Address & Gateway


The test set is pre-configured to obtain an IP Address using DHCP. The IP Address and Gateway can be changed. Consult the Microsoft Windows Help and Support Center to configure the LAN.

Windows Security

Microsoft recommends the following to ensure the test set Windows operating system is protected:

- Use an internet firewall.
- Get the latest critical Windows updates.
- Use up-to-date antivirus software.

To check the status or make changes in the security settings for your test set, click **Start, Control Panel**, and then **System and Security**. The following window appears:



NOTE

The window may look slightly different on your test set.

Windows Firewall

Click **Windows Firewall** in the System and Security window to check the firewall status of your test set. The test set is shipped with the Windows Firewall enabled.



NOTE


The window may look slightly different on your test set.

Windows Firewall exceptions for programs and ports have been added to allow proper operation of the test set over a network. Modifying these settings may prevent the test set from operating properly.

Automatic updates

The default test set setting is to automatically check for critical Windows Updates and notify you, if the test set has internet access.

You can change the configuration of the Microsoft Automatic Updates. Under **Windows Update** in the **System and Security** window, click **Turn automatic updating on or off** to change settings. If you choose not to have automatic updates, you can manually update windows by clicking **Check for updates** on the **System and Security** window.



NOTE


Be aware that downloading and installing Windows Updates can be network and CPU usage intensive (impacting the test set performance), and some Windows Updates automatically reboot the test set. It is recommended that Windows Updates be performed when the test set is not in normal use.

Virus protection

Under **Action Center** in the **System and Security** window, click **Review your computer's status and resolve issues** to check the status of virus protection on your test set. There is no antivirus software included with your test set. Antivirus application software has been tested to be compatible with the test set. See the section on **"3rd party software verified by Keysight"** on page 83 for anti-virus software that has been tested by Keysight.

NOTE

Having antivirus software installed may have a slight impact on the test set performance.



Spyware protection

The **Review your computer's status and resolve issues** window described above also shows the status of spyware protection. There is no anti-spyware software installed on the test set. This should not be a problem if you do not use the test set for a lot of internet browsing. Having spyware in the test set could have an impact on the test set performance.

System Maintenance

Backup

It is recommended that you have a regular backup strategy. Your IT department may already have a backup strategy in place which is suitable for the test set and its data. Using the Keysight Recovery system in conjunction with a regular backup strategy should allow full recovery of the test set data.

The Windows operating system has a Backup utility that you can use to archive files and folders in case of a hard disk drive failure. See the Microsoft Windows Help and Support Center for more information on this utility. You can also use third party backup utilities. However, you must ensure that this third party software is compatible with the test set's system software. See **“Customer Installation of Software”** on page 83 for more information.

When performing backups, we recommend that you backup the data to an external storage device connected to the network or one of the test set's USB connectors. Also, you should perform backups at times when the test set is not being used for normal operations as it may impact the test set's overall performance.

System restore

Windows contains the capability to restore the system to a previous point in time. System Restore is enabled with default settings as provided by Microsoft. However, System Restore is not 100% successful. Therefore, it is not the recommended method to backup the test set. System Restore has not been tested to verify successful restoring.

Disk defragmenting

Over time the hard disk on the test set becomes fragmented. Windows has a Disk Defragmenter utility that you can use to defragment the hard disk. See the Microsoft Windows Help and Support Center for more information on this utility.


NOTE

Running Disk Defragmenter should be performed when the measurement application is not running. Measurement throughput is significantly impacted while disk defragmentation is in process.

USB Connections

The controller front panel provides two USB 3.0 ports and four USB 2.0 ports to which you can connect USB mass storage devices and accessories such as printers and keyboards. The test set USB Host support includes the standard Microsoft Windows USB class drivers for human interface, mass storage, printing, scanning, and imaging devices. A complete up-to-date list of the Windows USB class driver support is available on the Microsoft website.

http://www.microsoft.com/whdc/system/bus/usb/USBFAQ_intro.msp



The test set USB device driver included in the test set software supports the test and measurement industry standard USBTMC-USB488 device class.

In addition, the Agilent/Keysight IO Libraries CD that was included with your test set contains USB Host drivers that allow control of other test sets connected to the USB bus.

Keysight Technologies does not support or warrant correct test set operation if additional USB drivers from third parties are installed in the test set. It is possible that additional drivers could break the normal USB operation. If USB operation is broken, recovery would require reinstalling the test set application using the hard drive recovery process.

Hard Drive Partitioning and Use

The drive is partitioned into 3 sections: C:, D: and E:

- The C: partition contains the Windows operating system and software installed by Keysight. This is an Open System which means you can install additional software, and these should be installed on the C: drive. However, only a limited set of software applications are tested for use with the Keysight measurement software. The installation and/or use of other software is not warranted, and could interfere with the operation of the measurement software. If test set repair is ever needed, the Keysight version of the C: drive is the only part of the test set software that is restored by the Keysight Recovery process. You must reload any other software that you have added in the test set.
- The D: partition is reserved for data storage. The User Accounts that are configured by Keysight have their My Documents folder mapped to the D: drive. This is for the convenience of backing-up the measurement data. You should always back-up the data on the D: drive to an external device. This allows you to restore the data if you ever need to replace the hard drive.
- The E: partition is reserved for Keysight's use. The primary use of the E: drive is for housing the Calibration and Alignment data. Do not change or overwrite the files on this drive. This could cause your test set to not meet specifications, or even to stop functioning correctly. Do not use this drive for data storage. It is also recommended that you back up the contents of this drive by using the factory calibration data backup utility.

Hard Drive Recovery Process

The Keysight Recovery System can be used to repair errors on the test set's C: drive partition, or to restore the original factory configuration of the system software. The Keysight Recovery System is stored in a separate hidden hard disk drive partition.

Repairing errors on the hard disk drive may result in loss of data or files. If you need more information about the Windows "chkdsk" error repair process, see the chkdsk documentation in the Microsoft Windows Help and Support Center.

Restoring the original factory system software does not restore any of the following items:

- Windows system configurations that were made after the test set was shipped from the factory. For example, Windows and Service Pack updates, user accounts, and Windows configuration settings. After an Keysight Recovery, these configurations need to be redone.
- Additional software that was installed after the test set was shipped from the factory. After an Keysight Recovery, that software needs to be re-installed.
- Any data or programs saved on the D: or E: drives.
- Any upgrades that were made to the Keysight measurement application software.



Restoring the original factory system software will not synchronize the code in the Field Programmable Gate Arrays (FPGAs) on the various hardware assemblies. As a result, you may see an error dialog box appear during the final boot-up at the end of the recovery process. This typically occurs when there are significant differences between the software version installed prior to performing the recovery and the version restored by the recovery. In these situations, upgrade the software to the latest version.

NOTE

It is recommended that you use a regular back up strategy. Your IT department may already have a back up strategy in place which is suitable for the test set and its data. See "**System Maintenance**" on page 101. Using the Keysight Recovery System in conjunction with a regular back up strategy should allow you to fully recover the test set software and data.

It is recommended that routine backups of the test set information be performed to keep current archives of the test set information. This allows a full recovery of the test set information after the test set recovery system operations are performed. See "**Backup**" on page 101 for more details.

Table 3-1 Using the test set recovery system

Step	Notes
1 Make sure the test set is turned off.	
2 Turn on the test set. Press the down arrow key to move the highlight to Keysight Recovery System , then press Enter .	<p data-bbox="933 447 1347 510">After the Keysight Technologies screen is displayed,</p> <div data-bbox="963 520 1295 674" style="text-align: center;">  </div> <p data-bbox="933 709 1347 741">This screen is displayed for five seconds.</p> <div data-bbox="937 747 1448 1014" style="text-align: center;">  </div>
3 When the Keysight Recovery System has booted, follow the on-screen instructions to recover the image of the C: drive. <ul style="list-style-type: none"> <li data-bbox="410 1150 893 1213">• Press 2, then press Enter to select the recovery. <li data-bbox="410 1224 893 1266">• Press 1, then press Enter to continue. <li data-bbox="410 1276 893 1308">• Press 1, then Enter to confirm. 	<p data-bbox="933 1213 1412 1276">It may take up to 25 minutes for this process to complete</p>
<p data-bbox="410 1329 1448 1432">After exiting the Keysight Recovery System, the test set reboots. If the original factory test set system has been restored, the test set re-executes the following process: Turning on the test set the first time" on page 17</p>	

NOTE

Additional recovery steps may be required to fully recover the system to a more current working state. This could involve restoring your own backups of the test set configuration, including re-installing applications, data, and performing system customizations.

Updating the software

The purpose of this update is to ensure that the E6640A's software, its measurement mode applications, and the FPGA content of its PC boards are all current and up to date.

NOTE An update to the instrument software revision does not require a new license key for the measurement applications, so long as the applications were licensed prior to the update.

NOTE This procedure will only update the instrument software that currently resides on the C: drive of the instrument. It will not update the version that exists on the recovery partition. Because of this, if the recovery process is performed, the instrument will revert back to the instrument software revision that was originally shipped from the factory.

Determining the current software revision

To determine what revision of instrument software is currently installed, press **System, Show, System** using the instrument's virtual front panel. You can access the virtual front panel using a mouse and right-clicking in the application window, and selecting **Utility, Virtual Front Panel**. Once in the show system screen, look for the "Instrument S/W Revision" entry (for example: "A.13.85"). If the installed revision is the same as the latest revision on the web, there is no need to update the instrument software.

Requirements

The following requirements must be met to use this procedure:

- The E6640A must be functioning properly before the update is attempted.
- You must have a USB keyboard and mouse.
- You must be able to log in to the instrument as the administrator.
- You must be able to transfer a large file (~1 GByte) to the E6640A being updated.

File Download

1. Go to <http://www.keysight.com/find/e6640a> and go to the Technical Support section; select the Driver, Firmware & Software tab.
2. Download the executable file by clicking on the file link. Save the file to a location of your choice. If you do not have the ability to download files directly to the C: drive of the E6640A being updated, store the file on a large capacity USB storage device for transfer to the instrument. (The file size is ~1 GByte.)

User account and log-in

The default user account is Instrument, which does not have the required permissions to install the instrument software updates. For the process outlined below, the user must be logged in as Administrator.

The automated instrument software upgrade process has an install wizard which removes the old software version and installs new software version without manual steps. The removal of the old software version and the installation of the new software version completes in approximately 15 minutes.

1. Connect a USB keyboard and mouse to the E6640A.
2. If the E6640A is not already running, power it up and allow it to boot up completely.
3. Close the E6640A measurement application if it is running. Right-click in the application window, and select **Utility > File** from the right-click menu; select the **Exit** softkey and click **OK** in the confirmation window.
4. Log out as the default user (Instrument). Click the Start icon at the lower left corner of the screen, and select **Log Off**.
5. Select the Administrator icon on the Windows desktop, to log in as follows:
User Name: `administrator`
Password: `Keysight4u!`
6. If a “Found New Hardware” wizard window pops up, select “No, not at this time” to bypass the activity.

Instrument software installation

If you downloaded the instrument software update file to the instrument via LAN, access that file from the location that it was stored and proceed to step 1.

If you downloaded the instrument software update file to a USB storage device, plug that into one of the USB ports on the front of the instrument. Access the file for the removable drive, then proceed to step 1.

Table 3-2 **Preparing the installation**

Step	Notes
1 Double click on the installer file, and click Yes in the confirmation window that appears.	The file name is in the format: <code>XSA_Installer_A.XX.XX.exe</code>

Table 3-2 Preparing the installation

Step	Notes
<p>2 A window appears showing the extraction of the software installer. Click Install in the confirmation window to begin the installation process. When the extraction process is complete, a software updater display appears, showing progress of the software removal and installation.</p>	<p>The update process causes various messages and windows to appear and disappear on the screen, but the process doesn't require further interaction. Examples of what will be displayed are shown below, as a guide to what the process looks like while it is running.</p>

Extraction window:




Table 3-2 Preparing the installation

Step	Notes
------	-------

Software updater window, showing uninstallation progress:




While the old instrument software is being uninstalled, the text “Uninstall in Progress” text rolls across the screen, to be replaced finally by “Completed Successfully”. No interaction is required at this point.

Software updater window, showing installation progress:



While the new instrument software is being installed, the text “Installation in Progress” text rolls across the screen, to be replaced finally by “Completed Successfully”. No interaction is required at this point. Please be patient: the total installation process can take 30 to 40 minutes.


Table 3-2 Preparing the installation

Step	Notes
FPGA update warning window (may be displayed, if any FPGA updating occurs):	
CAUTION	During FPGA programming, DO NOT power off the E6640A for ANY reason! Interrupting the FPGA update process can place the instrument in an unusable state, requiring it to be returned to Keysight for repair.
3 When the software updater window shows “The software upgrade was successful”, you will be prompted to shut down and restart the system.	Reboot the E6640A by clicking Finish in the window. Windows will display a “System Shutdown” warning window; this is normal.


Software updater window, showing completion of the process:



Table 3-2 Preparing the installation

Step	Notes
4 Click the “Finish” button. This results in a Windows shutdown.	
	
5 After you power the E6650A back on, it will boot up in the default “Instrument” account, perform alignments, and finally display the measurement screen.	If the instrument does not reboot, remove the USB storage device and cycle power. The boot order for the instrument may have been changed in the instrument BIOS previously, and the instrument may be instructed to boot from USB first, and the internal drive second. If this is the case, failure to remove the USB device will result in an error, as the instrument attempts to boot from it.
6 Check for any errors, and verify that the new software version is installed.	In the right-click menu, Press System . In the System menu, press Show, System . Look for the line “Instrument S/W Revision”.

Instrument S/W Revision:



It is possible a newer version of the instrument software is available on the web at <http://www.keysight.com/find/e6640a>

NOTE

Additional recovery steps may be required to fully recover the system to a more current working state. This could involve restoring your own backups of the instrument configuration, including re-installing applications, data, and performing system customizations.

Configuring recovery prompt timing

You can configure the time at which the test set power-up process waits for the selection of the recovery process by performing the following steps:

Step	Notes
1 From the Windows Start menu, right-click Computer and click Properties .	This opens the System window.
2 Click Advanced System Settings in the left pane.	This opens the System Properties window.
3 In the Startup and Recovery section, click Settings .	This opens the Startup and Recovery window.
4 Under the System Startup section: <ul style="list-style-type: none"> • you can either clear the Time to display a list of operating systems check box, or, • select the Time to display recovery option when needed check box and change the seconds to delay for it 	

NOTE

You must be logged in as an administrator to change these settings. See **"User Accounts"** on page 84 for more information.

5

Using Microsoft Windows Operating System

NOTE

The capabilities described in this section are Microsoft Windows features. The following information provides some guidelines for using the capabilities with the test set. You need to refer to the Windows help documentation for more information. Your version of Windows may not match these instructions exactly.

You need an external keyboard and mouse to fully use these features.

[“Navigating Windows Without a Mouse”](#) on page 114

[“Remote Desktop: Using the Test Set Remotely”](#) on page 115

[“Programmatic Control of the Test Set”](#) on page 129

[“Capturing/Printing Displays and Windows”](#) on page 130

[“Windows Shortcuts and Miscellaneous Tasks”](#) on page 131

Navigating Windows Without a Mouse

Key Presses	Actions
Esc	Exits/closes a Windows dialog box (does not exit an Application window)
Enter	Does the current "default action". If a menu item or a button is currently highlighted, then the Enter key activates that menu item or button.
Alt	Moves focus/control to the pull down menus bar in the active Window
Right Arrow	In pull-down menu: opens the next menu to the right, or opens a submenu In a dialog box: selects an option button
Left Arrow	In pull-down menu: opens the next menu to the left, or opens a submenu In dialog box: selects an option button
Up Arrow	In pull-down menu: Moves to next selection up in the menu In dialog box: selects an option button
Down Arrow	In pull-down menu: Moves to next selection down in the menu In dialog box: selects an option button
Tab	In dialog box: moves to the next/previous field
Del	Deletes the currently selected item
Alt + Tab	Switches between the next/previous Application
Alt + Enter	Shows the Properties of the currently selected item
Alt + Esc	Cycles through items in the order that they had been opened
Backspace	In My Computer or Windows Explorer: move up one level In Internet Explorer: works like the BACK arrow key
Ctrl + Left arrow	Moves to the left one word at a time
Ctrl + Right arrow	Moves to the right one word at a time
Ctrl + Tab	In dialog box: moves to the next/previous Tab location
Alt + Space	Opens the window control menu for the currently active window, allowing you to minimize, maximize, move and restore (size) the window
Ctrl + Esc	Opens the Windows Start Menu
Ctrl + Alt + Delete	Opens the Windows Task Manager

Remote Desktop: Using the Test Set Remotely

Windows Remote Desktop is recommended for remote control of the test set. The Remote Desktop functionality is a Microsoft Windows capability. The following discussion provides some guidelines for using this capability with the test set. You need to refer to the Windows help documentation for more information. As Windows evolves, these instructions may no longer be exact.

NOTE

Overview of Remote Desktop operation

Using the Remote Desktop functionality of the test set allows you to control and interact with the test set from a remote computer, as though you were sitting in front of the test set.

When you have configured the test set for remote connectivity, and configured a separate computer to act as a Remote Desktop Host, you can send commands to the test set from the remote computer, and you can see the test set display on the screen of the remote computer.

This section provides full details of how to set up the test set for remote connectivity, and also how to set up a computer running any 32-bit version of Microsoft Windows as a Remote Desktop Host.

Setting up Remote Desktop operation

Setting up the test set

Before the test set can be controlled through a Remote Desktop Connection, it must be set up to allow connection from a remote computer.


Table 3-3 **Setting up a remote desktop connection**

Step	Notes
1 To perform this operation successfully, you must have Administrator level access to the test set.	The default Administrator password is: Keysight4u!
2 On the test set, open the Windows Control Panel . Using the Virtual Front Panel within the test set application, press System, Control Panel , <i>or</i> , From the Windows Desktop, click Start, Control Panel .	
3 Within the Control Panel, select System, System and Security .	

Table 3-3 **Setting up a remote desktop connection**

Step	Notes
4 Click Remote Settings .	A System Properties window appears, with the Remote tab displayed.
5 Check the box labeled Allow Remote Assistance connections to this computer, if it is not already checked.	
6 To add users, click Select Users, Add .	
<ul style="list-style-type: none"> • Follow the on screen instructions. 	

Figure 3-1 Basic setup for remote desktop operation



Setting up the remote computer

The procedure depends on whether the Remote Computer to be set up is running Windows XP, Windows 7, or another version of Microsoft Windows.

Remote computer running Windows XP or Windows 7 Windows XP and Windows 7 include the Remote Desktop Connectivity Client software, so no additional setup is required.

Remote computer running another version of Windows You can use any 32-bit version of Windows (Windows 95, 98, ME, NT4, or 2000) to install and run the Client software for Remote Desktop Connectivity. However, you need to have available a Windows XP or Windows 7 installation CD-ROM, because that contains the Client software.

NOTE The following instructions relate to software provided by Microsoft Corporation. Keysight offers no warranty regarding the operation of such software. The procedure described here may be changed by Microsoft at some future time.

Table 3-4 Installing the Client software

Step	Notes
1 When the Welcome Screen appears, click Perform additional tasks	
2 From the What do you want to do? screen, click Set up Remote Desktop Connection .	The Remote Desktop Connection InstallShield Wizard appears.
3 Click Next .	Follow the on screen instructions provided by the Wizard.
4 To access the installed software, click Start > All Programs > Accessories > Communications > Remote Desktop Connection .	

How to locate the computer name of the test set

To connect a remote computer to the test set, you need to know its Computer Name. The Computer Name can be displayed as follows:

Table 3-5 Locating the name from the Keysight application

Step	Notes
• On the test set front panel, press System, Show, System .	A page listing various parameters appears. The test set's computer name is shown in the list next to the title Computer Name.

Table 3-6 Locating the name from the Windows desktop (with a mouse):

Step	Notes
1 Click Start, Control Panel .	In the Control Panel window, select System and Security, System .
2 The computer name is displayed in the System window, under the heading "Computer name, domain, and workgroup settings".	The Computer Name is listed as Full computer name .

Table 3-7 Locating the name from the Windows desktop (without a mouse):

Step	Notes
1 Press Ctrl+Esc to display the Windows Start menu.	
2 Use the Up Arrow or Down Arrow keys to select the Control Panel item.	
3 Press Enter to open the Control Panel dialog.	
4 Use the Tab key to highlight System and Security , then press Enter .	The System and Security window is displayed.
5 Use the Tab key to highlight System , then press Enter .	The System window is displayed.
6 The computer name is displayed in the System window, under the heading Computer name, domain, and workgroup settings .	The Computer Name is listed as Full computer name .
7 To close the Control Panel dialog: <ul style="list-style-type: none"> • Press Alt to open the drop-down File menu • Press the Down Arrow until the Close menu item is selected • Press Enter to close the Control Panel dialog. 	

Running a Remote Desktop session


Initializing a Remote Desktop session

NOTE

To initialize a Remote Desktop Session, you need to know the Computer Name of the test set. This information can be shown on the test set display by following the procedure in the section **“How to locate the computer name of the test set”** on page 117.

After setting up both the test set and the remote computer for Remote Desktop Connectivity, as described in **“Setting up Remote Desktop operation”** on page 115, you are ready to start a Remote Desktop session.

Table 3-8 Starting a session

Step	Notes
<p>1 Click Start > All Programs > Accessories > Remote Desktop Connection.</p>	<p>A Remote Desktop Connection dialog appears:</p> 
<p>2 Enter the computer name or IP address of the test set.</p>	
<p>3 Click Connect.</p>	<p>A login dialog box appears.</p>
<p>4 Enter the login account name and password.</p>	<p>The default account name is <i>Instrument</i> and the default password is <i>measure4u</i>, but these parameters may be changed by instrument users.</p>

NOTE

Only the current User or an Administrator can remotely log into the test set. To see who the current user of the test set is, press **Ctrl+Esc** on the test set until you can view the current user name on the Start menu. If no one is currently logged into the test set, any valid instrument user can remotely log in.


The test set display appears on the screen of the remote computer. Because the test set front-panel keys are not available when using the test set remotely, three alternative methods are available to perform the functions of the front-panel keys, as detailed in the following sections:

- **“Accessing menus for Remote Desktop operation”** on page 123
- **“Keycode commands for Remote Desktop operation”** on page 124
- **“The Virtual Front Panel”** on page 128

Setting Remote Desktop options

Step	Notes
------	-------

- 1 On the Remote Desktop Connection window, click **Options**. The window expands vertically to show several configuration tabs.




The Options dialog has several tabs. Generally, the default settings are correct.

- 2 Under the **General** tab, ensure that the Computer name, User name and Domain name are set correctly.


You may choose to enter the password and save it for future sessions, by checking the **Save my password** box.

Step**Notes****3** Click the **Display** tab.


- Under **Remote desktop size**, you may select the size of the window in which the test set display appears. Do *not* select any size smaller than 1024 x 768 pixels. Selecting a remote desktop size smaller than 1024 x 768 results in the test set display not being fully visible. In such circumstances, scroll bars do not appear, so portions of the display are not accessible.
- Under **Colors**, you may select any setting that uses 15 bits or more. Selecting a color setting that uses less than 15 bits results in dithering and incorrect color rendition of the remote desktop window.

**4** Click the **Local Resources** tab.

- Click the **More** button to see a selection of local resources; selecting the **Drives** checkbox enables you to transfer data between the remote desktop and the local PC. To transfer data, click **Start** on the task bar of the remote computer, then click **My Computer**. Explorer opens on the remote computer and displays the drives of both the remote computer and the local computer. You can now copy and paste between the two disk drives.




Step	Notes
5 Click the Experience tab.	To Optimize the performance of the Remote Desktop session, choose the appropriate connection format from the drop-down list.



Accessing menus for Remote Desktop operation

Using the mouse, right-click the application display window to access the following remote desktop menus:



The function of each selection in the menu is generally identical to that of the corresponding Virtual Front Panel key. The following lists additional functionality:

Exit – Selecting this item closes the Application software.

Utility > Page Setup – Selecting this item opens a printer setup dialog.

Utility > Virtual Front Panel – Selecting this item opens the Virtual Front Panel window, as described in the Section **“The Virtual Front Panel”** on page 128.

NOTE

The menus are also available when using the test set directly, provided that a mouse or other pointing device is attached to it.

Keycode commands for Remote Desktop operation

When using the test set in remote desktop mode, the following combinations of remote keyboard keys can be used to perform the operation of the virtual front-panel keys.




Table 3-9 Keycode commands

To initiate the following virtual key:	Press these keys on the remote computer keyboard:
AMPTD Y Scale	Ctrl+Shift+A
Auto Couple	Ctrl+Shift+C
Bk Sp	Backspace
BW	Ctrl+Alt+B
Cancel (Esc)	Esc

Table 3-9 Keycode commands

To initiate the following virtual key:	Press these keys on the remote computer keyboard:
Cont	Ctrl+Alt+C
Ctrl	Ctrl
Decrease Audio Volume	the Volume Control slider
Del	Delete
Down Arrow	Down Arrow
Enter	Enter (Return)
File	Ctrl+F
FREQ Channel	Ctrl+Shift+F
Full Screen	Ctrl+Shift+B
Help	F1
Increase Audio Volume	the Volume Control slider
Input/Output	Ctrl+Shift+O
Left Arrow	Left Arrow
Marker	Ctrl+Alt+K
Marker ->	Ctrl+Alt+N
Marker Function	Ctrl+Alt+F
Meas	Ctrl+Alt+M
Meas Setup	Ctrl+Alt+E
Menu (Alt)	Alt
Mode	Ctrl+Shift+M
Mode Preset	Ctrl+M
Mode Setup	Ctrl+Shift+E
Mute	the Mute checkbox under the Volume Control
Next Window	Ctrl+Shift+N
Peak Search	Ctrl+Alt+P
Print	Ctrl+P
Quick Save	Ctrl+Q
Recall	Ctrl+R
Restart	Ctrl+Alt+R
Return	Ctrl+Shift+R

Table 3-9 Keycode commands

To initiate the following virtual key:	Press these keys on the remote computer keyboard:
Right Arrow	Right Arrow
Save	Ctrl+S
Select	Space Bar
Single	Ctrl+Alt+S
Softkey 1	Ctrl+Shift+F1
Softkey 2	Ctrl+Shift+F2
Softkey 3	Ctrl+Shift+F3
Softkey 4	Ctrl+Shift+F4
Softkey 5	Ctrl+Shift+F5
Softkey 6	Ctrl+Shift+F6
Softkey 7	Ctrl+Shift+F7
Source	Ctrl+Alt+U
SPAN X Scale	Ctrl+Shift+S
Split Screen	Ctrl+Shift+L
Sweep/Control	Ctrl+Shift+W
System	Ctrl+Shift+Y
Tab	Tab
Trace/Detector	Ctrl+Alt+T
Trigger	Ctrl+Shift+T
Up Arrow	Up Arrow
User Preset	Ctrl+U
View/Display	Ctrl+Shift+V
Zoom	Ctrl+Shift+Z
1	1 ^a
2	2 ^a
3	3 ^a
4	4 ^a
5	5 ^a
6	6 ^a

Table 3-9 Keycode commands

To initiate the following virtual key:	Press these keys on the remote computer keyboard:
7	7 ^a
8	8 ^a
9	9 ^a
-	Use the – key to enter a negative value, as appropriate ^a
. (Decimal Point)	. (Period) ^a
0 (Zero)	0 (Zero) ^a

- a. For remote keyboards that feature a numeric keypad, use either the appropriate numeric keypad key or the main keypad key.

NOTE

The effect of pressing **Ctrl+Alt+Delete** on the keyboard of the remote computer *always* applies to the remote computer, and not to the test set. Generally, the effect of pressing this key combination is to display the Windows Task Manager. Therefore, pressing this key combination on the remote computer does *not* allow you to reboot the test set.

Ending a Remote Desktop session

There are two ways to disconnect the remote computer from the test set to end the session:


Step	Notes
1 Click the X , then click OK .	For full-screen, the X appears at the top center of the window. For non full-screen, the X appears in a red box at the right of the window's title bar.
or	
2 When the remote desktop is full screen, move the cursor to the bottom left of the window: <ul style="list-style-type: none"> Click Start, Disconnect. Click Disconnect. 	You are asked to confirm that you want to disconnect.

The Virtual Front Panel

The Virtual Front Panel is a software equivalent of the front-panel key set, which provides another alternate method for test set control. It may be displayed in a separate window on the test set (or remote desktop) display.

NOTE The Virtual Front Panel is also available when using the test set directly, provided that a mouse or other pointing device is attached to it.

Using the mouse, right-click the application display window, then select **Utility > Virtual Front Panel**. The Virtual Front Panel appears, as shown below:



When you click a button in the Virtual Front Panel, it performs the operation of the corresponding test set front-panel key. The buttons at the left of the panel, named Key 1 through Key 7, perform the operation of the respective softkeys. The Rpg Up and Rpg Down keys perform the operation of the knob.

Programmatic Control of the Test Set

For information on how to send commands to the test set, using a VISA connection, see the X-Series Signal Analyzer Programmer's Guide, which is available online at:

http://www.keysight.com/find/mxa_manuals

NOTE

The individual TRX instruments within the test set chassis are controlled using separate VISA connections; see **"TRX Selection"** on page 45.

Capturing/Printing Displays and Windows

You need an external keyboard and mouse to use this feature.

Save the desktop:

Step	Notes
1 Press Print Screen on the external keyboard	This captures the desktop and saves it on the Windows clipboard.
2 Open a graphics software program like Microsoft Paint	
3 Paste the clipboard contents into the program	The keyboard shortcut Ctrl + v will paste the contents of the clipboard.
4 Save the image in a file	

Save the current active window:

Step	Notes
1 Click on the window you want to capture	This activates the window.
2 Press Alt + Print Screen on the external keyboard	This captures the window and saves it on the Windows clipboard.
3 Open a graphics software program like Microsoft Paint	
4 Paste the clipboard contents into the program.	The keyboard shortcut Ctrl + v will paste the contents of the clipboard.
5 Save the image in a file.	

Windows Shortcuts and Miscellaneous Tasks

This section provides a list of Windows shortcuts (key combinations) that are useful when you operate the test set without an attached mouse and keyboard. (See also “[Navigating Windows Without a Mouse](#)” on page 114.) Although these shortcuts are available in any Windows system, they are not commonly used when a mouse and keyboard are attached.

Windows shortcuts (key combinations)

You can use the following combinations of front panel keys to perform basic Windows tasks when using the test set without an attached mouse and keyboard.


Table 3-10 Windows shortcut key combinations

To do the following:	Press:
Display the Windows Start Menu	Ctrl+Esc
Cycle through all open applications	Alt+Tab
Select the first menu of a menu bar	Alt
Move through menu headings	Left Arrow, Right Arrow
Open (drop down) a menu	Down Arrow
Move through items in an expanded menu	Up Arrow, Down Arrow
Close the current menu selection	Esc
Cancel the current menu bar selection	Alt
Open an application’s control menu (usually the left-most menu on the menu bar, starting with F ile)	Alt+Select
In a dialog: move between tabs	Ctrl+Tab
In a dialog: move forward through dialog box items	Tab
In a dialog: move backward through dialog box items	Shift+Tab
In a dialog: open a list box	Alt+Down Arrow
In a dialog list box or check box: select or deselect items	Select
In a dialog list box or check box: select or deselect one item at a time	Shift+Up Arrow, Shift+Down Arrow
In My Computer, expand a selected folder	Enter
In My Computer, open a folder one level up from the current folder	Bk Sp

Navigating an Application Menu without a Mouse or Keyboard

This example uses Microsoft Excel, but you can use a similar sequence of operations to select and execute any menu item of any application.

To select and execute the item **Sheet > Rename** option from the **Format** menu in the menu bar shown below,



perform the following operations:

Step	Notes
1 Press Alt to select the File menu in the menu bar.	The focus must be in the appropriate window.
2 Use the Right Arrow and Left Arrow keys to move horizontally to the Format menu.	
3 Press the Down Arrow to expand the Format menu.	
4 Use the Down Arrow and Up Arrow keys to move vertically to the Sheet menu item.	
5 Press the Right Arrow key to expand the Sheet sub-menu.	The Rename sub-menu item appears already selected. If another item in the same sub-menu is required, use the Down Arrow and Up Arrow keys to move vertically to that item.
6 Press Enter to execute the selected action.	

Windows taskbar auto-hide feature

The Windows taskbar should always be in the auto-hide mode when using the test set application. If the taskbar is not set to auto-hide, the lower part of the test set display is obscured by the taskbar.


If a mouse is attached to the test set, and you move the mouse cursor to the bottom of the display (either deliberately or accidentally), the taskbar automatically appears. Provided that the taskbar is in auto-hide mode, you can make it disappear again by moving the mouse cursor away from the bottom of the screen.

If at any time the Windows taskbar is inadvertently set to the non-auto-hide mode, you can restore the auto-hide behavior by doing the following:

Table 3-11 Restoring taskbar auto-hide mode

Step	Notes
1 Click Start > Control Panel	If not using a mouse, press Ctrl+Esc .
2 Click Appearance and Personalization, Taskbar and Start Menu	If you are not using a mouse, use the shortcut key combinations specified in the Section "Windows shortcuts (key combinations)" on page 131 to make these selections.
3 Click the Taskbar tab	The Taskbar and Start Menu Properties dialog appears.

Table 3-11 Restoring taskbar auto-hide mode

Step	Notes
	
4 Select the Auto-hide the taskbar check box	If you are not using a mouse, press Tab repeatedly until the auto-hide option is selected, then press Select to toggle the check box state.
5 Click OK .	This applies the change and closes the dialog box.

Windows startup folder

All Windows systems include a special folder, called the Startup folder. If a program, or a shortcut to a program, is placed in the Startup folder (either by Windows itself, or by a third-party application, or by any user), that program automatically runs every time Windows is restarted.

When your test set is first configured by Keysight, shortcuts to the appropriate application software and supporting programs are placed in the Startup folder. The exact contents of the Startup folder depends on the options you purchased with your test set.

Important You should *never* delete items that appear in the Startup folder, or move any item from the Startup folder to another folder. Deleting or moving any item in the Startup folder may cause applications not to start automatically when Windows is restarted, or may cause certain options to be unavailable when using the test set.

If you wish to view the contents of the Startup folder, do one of the following:

- Either click **Start, All Programs**, then select **Startup**,
or,
- Click **Start, My Computer**, then navigate to the folder:
C:\Documents and Settings\All Users\Start Menu\Programs\Startup.

Additional startup items may also appear in the corresponding startup folder for the user that is currently logged on. For example, the startup folder for the **administrator** user is **C:\Documents and Settings\administrator\Start Menu\Programs\Startup**.

5 Using Microsoft Windows Operating System

6

Troubleshooting

“Check the Basics” on page 138

“Problems with Microsoft Windows Operating System” on page 140

“Returning a test set for Service” on page 141

WARNING

No operator serviceable parts inside. Refer servicing to qualified personnel. To prevent electrical shock do not remove covers.

Check the Basics

- o Is there power at the receptacle?
- o Is the test set turned on? Check to see if the blue power light and the green fan LEDs beside the power switch are on and steady. Also, listen for internal fan noise to determine if the test set cooling fans are running.
- o If other equipment, cables, and connectors are being used with your test set, make sure they are connected properly and operating correctly.
- o If the test set cannot completely load or run the operating system, or the instrument application is not successfully launched, the problem could be a corrupt hard drive. If the test set gets far enough along in the boot process to run the "Keysight Recovery System", perform the **"Hard Drive Recovery Process" on page 104**. (The bootup process will be visible only if you have connected an external monitor.)
- o Is the Measurement Application running? If not, there is a software launch shortcut/icon on the desktop.
- o Does the test set application have the focus? (That is, is the blue window banner highlighted?) If not, move focus to the application with Alt-Tab.
- o Review the measurement procedures being performed when the problem first appeared. Are all of the settings correct?
- o If the test set is not functioning as expected, return the test set to a known state by pressing **Mode Preset**.

NOTE

Some test set settings are not affected by a Preset. If you wish to reset the test set settings, press **System, Power On, Restore Power On Defaults**.

-
- o Is the measurement being performed, and the results that are expected, within the specifications and capabilities of the test set? Refer to the specifications in the data sheet for your test set, which is available on the Keysight website: <http://www.keysight.com/find/e6640a>
 - o If the test set is not communicating via the LAN connection, check the LAN cable connection at the front panel.
 - o To meet specifications, the test set must be aligned. Either the Auto Align (On) feature must be selected (press **System, Alignments, Auto Align, Normal**), or the test set must be manually aligned.
 - o Perform an Alignment. Press **System, Alignments, Align Now, All**.

- o If the previously performed alignments did not resolve the problem, press **System, Alignments, Restore Align Defaults**. Then press **System, Alignments, Align Now, All**.

NOTE

The test set must be powered on for 45 minutes prior to performing alignments.

- o Is the test set displaying an error message? If so, refer to the Instrument Messages Guide.
- o Check if the external frequency reference is selected but not available. Verify that it is selected by pressing **Input/Output, Freq Ref In**. If **External** is selected, changing the setting to **Sense** allows the test set to sense the presence of an external reference and use it only if it is available. The frequency of the reference should be set correctly.
- o If you are using a Windows program, other than the test set application, you may notice it running slow. Place the test set application in single sweep/measurement.

Problems with Microsoft Windows Operating System

The Microsoft Windows operating system settings have been optimized for the best performance. Modification of these settings may degrade test set performance and measurement speed. Those that can be safely modified are described in **“Settings that can be changed”** on page 93.

The E6640A EXM Wireless Test Set operates in an open Windows environment, so you can install software on the test set. However, installation of non-approved software may affect test set performance. Keysight does not warrant the performance with non-approved software installed.

Returning a test set for Service

Calling Keysight Technologies

Keysight Technologies has offices around the world to provide you with complete support for your test set. To obtain servicing information or to order replacement parts, contact the nearest Keysight Technologies office listed in the following table. In any correspondence or telephone conversations, refer to your test set by its product number, full serial number, and software revision.

Press **System, Show, System**, and the product number, serial number, and software revision information is displayed on your test set screen. A serial number label is also attached to the rear panel of the test set.

Locations for Keysight Technologies

Americas

Canada
(877) 894 4414

Brazil
55 11 3351 7010

Mexico
001 800 254 2440

United States
(800) 829 4444

Asia & Pacific

Australia
1 800 629 485

China
800 810 0189

Hong Kong
800 938 693

India
1 800 112 929

Japan
0120 (421) 345

Korea
080 769 0800

Malaysia
1 800 888 848

Singapore
1 800 375 8100

Taiwan
0800 047 866

Other Asia-Pacific countries:
(65) 6375 8100

Europe & Middle East

Austria
0800 001122

Belgium
0800 58580

Finland
0800 523252

France
0805 980333

Germany
0800 6270999

Ireland
1800 832700

Israel
1 809 343051

Italy
800 599100

Luxembourg
+32 800 58580

Netherlands
0800 0233200

Russia
8800 5009286

Spain
0800 000154

Sweden
0200 882255

Switzerland
0800 805353
Opt. 1 (DE), Opt. 2 (FR), Opt. 3 (IT)

United Kingdom
0800 0260637

Online assistance: <http://www.keysight.com/find/assist>

Contact us: <http://www.keysight.com/find/contactus>

Read the Warranty

The warranty for your test set is in the front of your Specifications Guide. Please read it and become familiar with its terms.

If your test set is covered by a separate maintenance agreement, please be familiar with its terms.

Service Options

Keysight Technologies offers several optional maintenance plans to service your test set after the warranty has expired. Call your Keysight Technologies office for full details.

Packaging the Test Set

Use original packaging or comparable. It is best to pack the unit in the original factory packaging materials if they are available.

CAUTION

Test set damage can result from using packaging materials other than those specified. Never use styrene pellets in any shape as packaging materials. They do not adequately cushion the test set or prevent it from shifting in the carton. They cause test set damage by generating static electricity and by lodging in the test set louvers, blocking airflow.

You can repackage the test set with commercially available materials, as follows:

Step	Notes
1 Wrap the test set in antistatic plastic to reduce the possibility of damage caused by electrostatic discharge.	
2 Use a strong shipping container.	The carton must be both large enough and strong enough to accommodate the test set. A double-walled, corrugated cardboard carton with 159 kg (350 lb) bursting strength is adequate. Allow at least 3 to 4 inches on all sides of the test set for packing material.
3 Surround the equipment with three to four inches of packing material and prevent the equipment from moving in the carton.	If packing foam is not available, the best alternative is plastic bubble-pak. This material looks like a plastic sheet filled with 1-1/4 inch air bubbles. Use the pink-colored bubble which reduces static electricity. Wrapping the equipment several times in this material should both protect the equipment and prevent it from moving in the carton.
4 Seal the shipping container securely with strong nylon adhesive tape.	
5 Mark the shipping container "FRAGILE, HANDLE WITH CARE" to assure careful handling.	
6 Retain copies of all shipping papers.	

Index

Symbols

.NET, 95

A

acquisition integration, 64
add
 a network connection, 93
 a printer, 93
Administrator logon, 84
Adobe Acrobat Reader, 95
Agilent directories and files, 94
Agilent I/O Libraries, 95
Agilent Recovery System, 104, 105
Agilent Technologies
 locations, 142
Agilent Technologies, calling, 141
Agilent-Only user accounts, 94
alignments, 44
anti virus programs, 93
anti-virus software, 83, 100
applications, 51
Auto-hide Taskbar, 133
automatic updates, 97, 99
Automatic Updates, Microsoft
 Windows, 93

B

backing up your instrument, 101
battery replacement, 27
beam forming, 79
boot wait time, 112

C

C drive usage, 103
calibration certificates, 48
calibration status, 47
changing
 Windows system settings, 93
classic Windows display, 94
commands, 129
Computer Name, 120
computer name
 how to locate it, 117
configurator, 50
configure applications, 51

connectors, front panel, 30
contents of shipment, 14
Control Panel
 taskbar, 133
Control Panel access, 96
Control Panel settings that can be
 changed, 93
controlling Windows without a
 mouse, 114
cooling the instrument, 16

D

D drive data storage, 103
data storage, 103
date and time setting, 93
defragmenting your disk, 101
desktop icons, 50
desktop saving, 130
Device Manager settings, Microsoft
 Windows, 94
Device USB port, 102
directories, Agilent, 94
disk defragmentation, 101
disk drive partitioning, 94
display annotations, 38
display print, 130
display save, 130
Display settings, Microsoft
 Windows, 94
Domain Name, 120
drive C use, 103
drive D use, 103
drive partitioning, 94
drive recovery, 104
Drivers settings, Microsoft
 Windows, 94
duplex, 52

E

E6640A
 Computer Name, 120
 computer name, 117
 Domain Name, 120
 popup menus, 123
 virtual front panel, 128
electrostatic discharge (ESD)
 protecting against, 28
Environment Variables settings,
 Microsoft Windows, 94

environmental conditions, 25
Error Reporting settings, Microsoft
 Windows, 94

F

file backup, 101
firewalls, 97, 98
fixing the hard disk, 105
font size, changing, 94
Fonts removal, Microsoft Windows, 94
front panel
 connectors and keys, 30
 keycode alternates, 124
 popup menu alternative, 123
 shortcut keys, 124
 symbols, 41
 virtual, 128
fuses, 24

G

gateway setting, 96
GPIB27 or GPIB28, 95

H

half-duplex, 54, 57, 63
hard disk recovery process, 105
hard drive recovery, 104
Hibernate, Microsoft Windows, 94
HiSLIP, 45
Host USB ports, 102
hostname setting, 96

I

icons, 50
IIS server, 94
initialization with mouse, 17
installing other software, 83
installing programs to drive C, 103
instrument location, 16
Instrument logon, 84, 85
instrument recovery, 105
instrument ventilation, 16
integration, 64
IP address setting, 96

Index

K

Keycode Commands
Windows, 131
keycode commands, 124
keys, 30
shortcut, 124

L

LAN address, 61
LAN configuration, 96
launch XSA, 51
licenses for software, 82
line voltage, 24
loading a printer driver, 96
loading other software, 83
locating the instrument, 16
logon
accounts, 85
administrator, 84
instrument, 84

M

message URL http
//www.keysight.com/find/e6640a, 4
//www.keysight.com/find/tips, 4
Microsoft SQL Server Desktop
Engine, 94
Microsoft Visual J# .NET, 95
Microsoft Windows
Automatic Updates, 93
Display settings, 94
Fonts removal, 94
network connection Windows
network connection, 93
Power Options, 94
printer install and configure, 93
set screen saver, 94
set time and date, 93
System Properties, 94
Tab settings, 94
user accounts, 93
virus protection, 93
Microsoft Windows system
settings, 93
MIMO, 70
MIMO (switched), 62
MIMO source settings, 72
MIMO type, 73
modulation accuracy, 66
modulation analysis, 76, 77
mouseless Windows navigation, 114
MSSQL\$CDF service, 94
My Documents location, 103

N

navigating without a mouse, 114
network connection, Microsoft
Windows, 93

O

operating remotely, 115

P

packaging, 143
partitioning the hard-disk drive, 94
passwords, 84, 85
Performance settings, Microsoft
Windows, 94
port configuration, 52, 62
power cords, 25
Power Options, Microsoft Windows, 94
power source, 24
power-on time, 23
print screen, 130
printer driver loading, 96
printer install and configure, Microsoft
Windows, 93
printer setup, 96
printing a window, 130
printing the desktop, 130
problems with shipment, 15
program storage, 103
programmatic control, 129

R

rack mounting, 16
radio standards, 62, 73
rear panel
symbols, 41
rear panel features, 40
recovering the hard drive, 104
recovery process, 105
recovery timing, 112
reduce power-on time, 23
registry entries, 94
remote analyzer configuration, 73
Remote Desktop
keycodes, 124
running, 118
Remote Operation
Connection Speed, 120
remote operation, 115
replacing battery, 27
restoring your instrument, 101
RF input port, 65
RF input port mode, 63
running Windows without a
mouse, 114

S

saving a window image, 130
saving data to internal drives, 103
saving programs on C drive, 103
saving the desktop image, 130
screen capture, 130
screen resolution setting, Microsoft
Windows, 94
screen saver settings, Microsoft
Windows, 94
security, 97
service options, 142
services running, 94
setting
gateway, 96
hostname, 96
ip address, 96
LAN, 96
shipment contents, 14
Shortcut keys
Windows, 131
shortcut keys, 124
SICL, 46
socket numbers, 129
socket ports, 46
software installed by customers, 83
software licenses, 82
source settings, 72
source sync control, 63
spyware, 100
Startup and Recovery settings,
Microsoft Windows, 94
Startup Folder
Windows, 135
storing data on D drive, 103
sub-instruments, 129
symbols, on front and rear panels, 41
system backup, 101
System Properties, Microsoft
Windows, 94
system recovery, 104
system restore, 101
system settings that can be
changed, 93

T

Tab settings, Microsoft Windows, 94
Taskbar
Windows, 133
telnet ports, 46
test set software, 82
third party software, 83
time and date setting, Microsoft
Windows, 93
triggers, 64
TRX configuration, 70
TRX configurator, 50
TRX selection, 45, 73

turn on
 with a mouse, 17
 Tx beam forming, 79

U

updates, automatic, 93
 updating Windows, 99
 URL
 contact us, 142
 sales and service, 142
 USB host ports, 102
 USB Series A or B ports, 102
 User Accounts location, 103
 user accounts, Microsoft Windows, 93
 user logon, 84
 user logon accounts, 85
 User Profiles settings, Microsoft
 Windows, 94
 user startup folder, 135
 using the test set remotely, 115
 using Windows without a mouse, 114

V

V9060A software, 82
 V9077A software, 82
 verifying contents, 14
 virtual front panel, 128
 virus protection, 83, 100
 virus protection programs, 93
 VXI-11, 46

W

warranty, 142
 window saving, 130
 Windows, 114
 Automatic Updates, 93
 Control Panel, 133
 Display settings, 94
 Fonts removal, 94
 Power Options, 94
 print screen, 130
 printer install and configure, 93
 Remote Desktop
 Installing, 116
 Setting up, 116
 screen capture, 130
 set screen saver, 94
 set time and date, 93
 Shortcut keys, 131
 Startup Folder, 135
 System Properties, 94
 Tab settings, 94
 taskbar
 auto-hide, 133
 user accounts, 93

 virus protection, 93
 Windows 7
 operating system, 113
 Windows Classic Style, 94
 Windows firewall, 98
 Windows Remote Desktop, 115
 keycodes, 124
 options, 120
 Options Dialog
 Experience Tab, 120
 General Tab, 120
 popup menus, 123
 running, 118
 Setting up Remote Computer, 116
 setting up the E6640A, 115
 Setting up Windows Systems, 116
 32-bit, 116
 virtual front panel, 128
 windows security, 97
 Windows services, 94
 Windows settings, 93
 Windows Style, 94
 windows system, 113
 Windows updates, 99

X

XSA, 51

