

Certificate Of Calibration

Certificate No: 33621AMY53700926

Manufacturer: Keysight Technologies
Model No: 33621A
Options Installed With Specifications: N/A

Description: Function/Arbitrary Waveform Generator
Serial No: MY53700926

Date of Calibration: 08-JUN-2018
Temperature: (23 ± 5) °C
Procedure: ATM-09-A2600

Humidity: (20 to 80)% RH

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures in compliance with a quality management system registered to ISO 9001:2015.

As Received Conditions: Factory tested - No incoming data available.

As Shipped Conditions: At the completion of the calibration, measured values were IN-SPECIFICATION at the points tested.

Remarks or special requirements:

Notes:

1) This calibration report shall not be reproduced, except in full.

Traceability Information: Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

Calibration Equipment Used:

Model Number	Model Description	<i>Date Used: Date equipment used in this Calibration.</i>		
-----	-----	Trace Number	Date Used	Cal Due Date
-----	-----	-----	-----	-----
1395D-1-09-10	THERMAL VOLTAGE CONVERTER	PD2013	08-JUN-2018	12-APR-2019
33250A	FUNC/ARB WAVEFORM GENERATOR	P3814	08-JUN-2018	21-DEC-2018
3458A	8-1/2 DIGIT MULTIMETER	P2865	08-JUN-2018	11-SEP-2018
53230A	UNIVERSAL COUNTER/TIMER,350MHZ	P2403	08-JUN-2018	16-MAR-2019
BAL_1395D-6-09-10	THERMAL VOLTAGE CONVERTER	PD2001	08-JUN-2018	18-JAN-2019
E9304A	E-SERIES AVERAGE POWER SENSOR	P2472	08-JUN-2018	29-MAR-2019
E9304A	E-SERIES AVERAGE POWER SENSOR	P2835	08-JUN-2018	19-MAR-2019
FS725	RUBIDIUM FREQUENCY STANDARD	P2871	08-JUN-2018	09-DEC-2018

Print Date: 08-JUN-2018



 Tay Eng Su
 Quality Manager

Keysight Technologies				
DD	MM	YY	BY:	
CAL	08	06	18	SY
DUE				