

Certificate Of Calibration

Certificate No: E36313AMY57040854

Manufacturer: Keysight Technologies

Model No: E36313A

Options Installed With Specifications: 0E3 STD PLT

Description: TRIPLE OUTPUT DC POWER SUPPLY

Serial No: MY57040854

Date of Calibration: 28-JAN-2018

Temperature: (23 ± 2)°C

Procedure: 07-TE40-KYSHLACUMULONIMBUS-001

Humidity: (40 to 75)% RH

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures in compliance with a quality management system registered to ISO 9001:2008.

As Received Conditions: Factory tested - No incoming data available.

As Shipped Conditions: At the completion of the calibration, measured values were IN-SPECIFICATION at the points tested.

Remarks or special requirements: For inquiries regarding this calibration please contact Keysight Technologies at the above address.

Notes:

1) This calibration report shall not be reproduced, except in full.

Traceability Information: Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

Calibration Equipment Used:

Model Number	Model Description	<i>Date Used: Date equipment used in this Calibration.</i>		
-----	-----	Trace Number	Date Used	Cal Due Date
-----	-----	-----	-----	-----
3458A	8-1/2 DIGIT MULTIMETER	P3219	28-JAN-2018	21-MAR-2018
3458A	8-1/2 DIGIT MULTIMETER	P3453	28-JAN-2018	11-JUN-2018
3458A	8-1/2 DIGIT MULTIMETER	P3591	28-JAN-2018	21-MAR-2018
3458A	8-1/2 DIGIT MULTIMETER	P3807	28-JAN-2018	22-DEC-2018
DA1855A	DIFFERENTIAL AMPLIFIER	P3922	28-JAN-2018	14-AUG-2018
MSO8104A	4GSA/S, 1GHZ 4+16CH MSO	PA3984	28-JAN-2018	03-MAY-2018
ROH URE2	RMS VOLTMETER	P3372	28-JAN-2018	28-AUG-2019
SHUNT	SHUNT RESISTOR	P3402	28-JAN-2018	29-NOV-2018

Print Date: 28-JAN-2018



Arumugam Varathaiah
 Site Quality Manager
 Keysight Calibration Provider # 25865

Keysight Provider #25865				
	DD	MM	YY	BY:
CAL	28	01	18	SY
DUE				